

CROSS-APPLICATION CONTROL AND TRIGGERING

Specifications

R&S®FSW3-KM135

Introduction

The specifications of the R&S®FSW3-KM135 cross-application control and triggering group measurements are based on the specifications document of the FSWX signal and spectrum analyzer.

General remarks

The specifications in this document cover the R&S®FSW3-KM135 option. R&S®FSW3-KM135 is not a separate application, but rather a general feature of the FSWX instrument. It requires the R&S®FSW3-KM135 license. The abbreviation "CrossACT" is used for cross-application control and triggering. CrossACT enables multiple applications to measure jointly.

Key features

- Synchronous data capture between multiple application instances, e.g. to analyze LTE and 5G carriers simultaneously
- Synchronous data capture between multiple application instances at different RF inputs, e.g. from different hardware paths
- Synchronous evaluation of time-aligned data in different applications, e.g. multiple carriers being demodulated in different applications
- Evaluation of time-aligned data using different measurement settings (frequency, sample rate etc.), e.g. to measure interference between WLAN and radar signals
- Common display of all data captures for multiple inputs in the CrossACT group (CrossACT visualization)
- Easy configuration of CrossACT groups via drag and drop

Additional software licenses

The cross-application control and triggering option can only be used in combination with other applications whose software licenses are required.

The following applications support the cross-application control and triggering option:

- R&S®FSW3 I/Q capture (no license needed)
- R&S®FSW3-KM100 I/Q spectrum analyzer
- R&S®FSW3-KM320 5G NR Release 15/16 downlink measurements
- R&S®FSW3-KM321 5G NR Release 15/16 uplink measurements
- R&S®FSW3-KM326 5G NR Release 17/18/19 measurements
- R&S®FSW3-KM327 O-RAN measurements
- R&S®FSW3-KM335 Beyond 5G measurements

Specifications | Version 01.00

ROHDE & SCHWARZ
Make ideas real



Restrictions and limitations

The following limitations apply to CrossACT:

- A CrossACT group must contain two application instances. The instances can be of the same or different application types
- Measurements that require multiple data captures cannot be performed in a CrossACT group

The following restrictions and limitations apply to CrossACT with the firmware version 1.10:

- Only a single CrossACT group can be defined
- A CrossACT group contains a maximum of two application instances
- Each application instance in a CrossACT group can only use a single input source
- Each input source can only be used by a single application instance within a CrossACT group
- Measurements based on file input cannot be part of a CrossACT group
- Trace averaging (sweep count > 1) is not supported
- Cross-correlation functionality is not supported
- Auto level is not supported

Ordering information

Designation	Type	Order No.
Cross-application control and triggering	R&S®FSW3-KM135	1349.2129.02

Rohde & Schwarz GmbH & Co. KG

www.rohde-schwarz.com

Rohde & Schwarz training

www.training.rohde-schwarz.com

Rohde & Schwarz customer support

www.rohde-schwarz.com/support

Certified Quality Management

ISO 9001

Certified Environmental Management

ISO 14001

R&S® is a registered trademark of Rohde & Schwarz GmbH & Co. KG

Trade names are trademarks of the owners

PD 3610.0695.22 | Version 01.00 | November 2025 (jr)

Cross-Application Control and Triggering

Data without tolerance limits is not binding | Subject to change

© 2025 Rohde & Schwarz GmbH & Co. KG | 81671 Munich, Germany