Quantitative Analysis Results Summary Rep

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin				
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

Di(2-ethylhexyl)adipate

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
DEHA ST-01.d	Cal	6.928	5313	118480	0.5897
DEHA ST-02.d	Cal	6.928	11524	116940	1.0378
DEHA ST-03.d	Cal	6.928	32512	118167	2.5111
DEHA ST-04.d	Cal	6.928	63329	116656	4.7449
DEHA ST-05.d	Cal	6.928	138198	116455	10.1165
BK.d	Blank			116405	ND
FBK(RE 26-03-04).d	Blank			116641	ND
FBK(KD 26-03-04).d	Blank			111076	ND
RE-01.d	Sample			110274	ND
RE-02.d	Sample			106125	ND
RE-03.d	Sample			106557	ND
RE-04.d	Sample			106610	ND
RE-09.d	Sample			106771	ND
RE-10.d	Sample			106656	ND
RE-11.d	Sample			106003	ND
RE-12.d	Sample			105704	ND
KD-01.d	Sample			105063	ND
KD-02.d	Sample			103391	ND
KD-03.d	Sample			102740	ND
MBK.d	Blank			103327	ND
CCV.d	QC	6.928	76385	112817	5.8645

Di(2-ethylhexyl)adipate



Cal. Path	Type	Level	Enable	Conc.	Response
V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\DEHA ST-01.d	Calibration	1	x	0.500	5313
V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\DEHA ST-02.d	Calibration	2	x	1.000	11524
V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\DEHA ST-03.d	Calibration	3	x	2.500	32512
V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\DEHA ST-04.d	Calibration	4	x	5.000	63329
V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\CCV.d	QC	4	x	5.000	76385
V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\DEHA ST-05.d	Calibration	5	x	10.000	138198

Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 2:54:15	**Data File**	DEHA ST-01.d
Type	Cal	**Name**	DEHA ST-01
Dil.	1	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate	6.928	5313	118480	0.5897 mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:04:51	**Data File**	DEHA ST-02.d
Type	Cal	**Name**	DEHA ST-02
Dil.	1	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate	6.928	11524	116940	1.0378 mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:15:26	**Data File**	DEHA ST-03.d
Type	Cal	**Name**	DEHA ST-03
Dil.	1	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate	6.928	32512	118167	2.5111 mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:26:01	**Data File**	DEHA ST-04.d
Type	Cal	**Name**	DEHA ST-04
Dil.	1	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate	6.928	63329	116656	4.7449 mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:36:38	**Data File**	DEHA ST-05.d
Type	Cal	**Name**	DEHA ST-05
Dil.	1	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate	6.928	138198	116455	10.1165 mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:47:13	**Data File**	BK.d
Type	Blank	**Name**	BK
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			116405	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:57:51	**Data File**	FBK(RE 26-03-04).d
Type	Blank	**Name**	FBK(RE 26-03-04)
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			116641	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:08:28	**Data File**	FBK(KD 26-03-04).d
Type	Blank	**Name**	FBK(KD 26-03-04)
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			111076	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:19:04	**Data File**	RE-01.d
Type	Sample	**Name**	RE-01
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			110274	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:29:43	**Data File**	RE-02.d
Type	Sample	**Name**	RE-02
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			106125	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:40:19	**Data File**	RE-03.d
Type	Sample	**Name**	RE-03
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			106557	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:50:58	**Data File**	RE-04.d
Type	Sample	**Name**	RE-04
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			106610	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:01:34	**Data File**	RE-09.d
Type	Sample	**Name**	RE-09
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			106771	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:12:12	**Data File**	RE-10.d
Type	Sample	**Name**	RE-10
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			106656	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:22:50	**Data File**	RE-11.d
Type	Sample	**Name**	RE-11
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			106003	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:33:27	**Data File**	RE-12.d
Type	Sample	**Name**	RE-12
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			105704	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:44:03	**Data File**	KD-01.d
Type	Sample	**Name**	KD-01
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			105063	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:54:42	**Data File**	KD-02.d
Type	Sample	**Name**	KD-02
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			103391	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 6:05:20	**Data File**	KD-03.d
Type	Sample	**Name**	KD-03
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			102740	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 6:16:01	**Data File**	MBK.d
Type	Blank	**Name**	MBK
Dil.	0.002	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate			103327	ND mg/L

Di(2-ethylhexyl)adipate







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-03\DEHA(RE 3.4, KD 3.4)\QuantResults\DEHA.batch.bin		
Analysis Time	2026-03-10 오후 4:27:41	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:06:47	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:27:41	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 6:26:41	**Data File**	CCV.d
Type	QC	**Name**	CCV
Dil.	1	**Acq. Method File**	ORG(DEHA) METHOD

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Di(2-ethylhexyl)adipate	6.928	76385	112817	5.8645 mg/L

Di(2-ethylhexyl)adipate