Quantitative Analysis Results Summary Rep

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin				
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

1,1-DCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	4.991	8628	1737	0.016
VOC10 ST-02.D	Cal	4.996	12978	1607	0.022
VOC10 ST-03.D	Cal	4.991	34841	1277	0.055
VOC10 ST-04.D	Cal	4.996	55140	1265	0.083
VOC10 ST-05.D	Cal	4.991	115765	1258	0.168
VOC10 ST-06.D	Cal	4.996	264526	1162	0.403
BK-.D	Blank			1373	ND
FBK(PL 25-12-30).D	Blank			1902	ND
PL-01.D	Sample			1638	ND
PL-02.D	Sample			1683	ND
MBK.D	Blank			1552	ND
CCV.D	QC	4.990	119812	1073	0.202

DCM

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.148	18202	1737	0.011
VOC10 ST-02.D	Cal	5.153	17595	1607	0.014
VOC10 ST-03.D	Cal	5.153	20429	1277	0.050
VOC10 ST-04.D	Cal	5.153	30751	1265	0.108
VOC10 ST-05.D	Cal	5.153	44281	1258	0.185
VOC10 ST-06.D	Cal	5.153	77247	1162	0.405
BK-.D	Blank	5.153	8440	1373	0.000
FBK(PL 25-12-30).D	Blank	5.153	11761	1902	0.000
PL-01.D	Sample	5.147	11579	1638	0.000
PL-02.D	Sample	5.153	47474	1683	0.136
MBK.D	Blank	5.148	9090	1552	0.000
CCV.D	QC	5.147	38888	1073	0.192

Chloroform

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.720	5002	1737	0.015
VOC10 ST-02.D	Cal	5.720	8786	1607	0.023
VOC10 ST-03.D	Cal	5.720	19163	1277	0.054
VOC10 ST-04.D	Cal	5.720	34951	1265	0.095
VOC10 ST-05.D	Cal	5.720	68759	1258	0.183
VOC10 ST-06.D	Cal	5.720	144024	1162	0.409
BK-.D	Blank			1373	ND
FBK(PL 25-12-30).D	Blank			1902	ND
PL-01.D	Sample			1638	ND
PL-02.D	Sample			1683	ND
MBK.D	Blank			1552	ND
CCV.D	QC	5.720	62838	1073	0.196

Carbon Chloride

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.922	5257	1737	0.017
VOC10 ST-02.D	Cal	5.922	7416	1607	0.021
VOC10 ST-03.D	Cal	5.922	20674	1277	0.056
VOC10 ST-04.D	Cal	5.922	32212	1265	0.083
VOC10 ST-05.D	Cal	5.922	67473	1258	0.166
VOC10 ST-06.D	Cal	5.922	155500	1162	0.403
BK-.D	Blank			1373	ND
FBK(PL 25-12-30).D	Blank			1902	ND
PL-01.D	Sample			1638	ND
PL-02.D	Sample			1683	ND
MBK.D	Blank			1552	ND

Carbon Chloride

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
CCV.D	QC	5.922	70556	1073	0.202

1,2-DCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.986	2222	1737	0.011
VOC10 ST-02.D	Cal	5.986	4494	1607	0.022
VOC10 ST-03.D	Cal	5.987	9322	1277	0.052
VOC10 ST-04.D	Cal	5.986	18319	1265	0.100
VOC10 ST-05.D	Cal	5.986	35711	1258	0.192
VOC10 ST-06.D	Cal	5.986	69767	1162	0.404
BK-.D	Blank			1373	ND
FBK(PL 25-12-30).D	Blank			1902	ND
PL-01.D	Sample			1638	ND
PL-02.D	Sample			1683	ND
MBK.D	Blank	5.967	0	1552	ND
CCV.D	QC	5.986	33955	1073	0.214

Benzene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.993	13857	1737	0.015
VOC10 ST-02.D	Cal	5.993	23310	1607	0.023
VOC10 ST-03.D	Cal	5.993	53116	1277	0.056
VOC10 ST-04.D	Cal	5.993	92682	1265	0.095
VOC10 ST-05.D	Cal	5.993	182991	1258	0.183
VOC10 ST-06.D	Cal	5.993	382018	1162	0.409
BK-.D	Blank			1373	ND
FBK(PL 25-12-30).D	Blank			1902	ND
PL-01.D	Sample			1638	ND
PL-02.D	Sample			1683	ND
MBK.D	Blank			1552	ND
CCV.D	QC	5.993	163656	1073	0.192

TCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	6.232	3830	1737	0.017
VOC10 ST-02.D	Cal	6.232	5901	1607	0.024
VOC10 ST-03.D	Cal	6.232	14549	1277	0.058
VOC10 ST-04.D	Cal	6.232	24337	1265	0.092
VOC10 ST-05.D	Cal	6.232	48732	1258	0.178
VOC10 ST-06.D	Cal	6.232	107034	1162	0.412
BK-.D	Blank			1373	ND
FBK(PL 25-12-30).D	Blank			1902	ND
PL-01.D	Sample			1638	ND
PL-02.D	Sample			1683	ND
MBK.D	Blank			1552	ND
CCV.D	QC	6.232	45787	1073	0.195

Toluene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	6.851	14225	1737	0.016
VOC10 ST-02.D	Cal	6.851	22776	1607	0.023
VOC10 ST-03.D	Cal	6.851	52769	1277	0.056
VOC10 ST-04.D	Cal	6.851	90883	1265	0.094
VOC10 ST-05.D	Cal	6.851	179678	1258	0.181
VOC10 ST-06.D	Cal	6.851	381691	1162	0.410
BK-.D	Blank			1373	ND
FBK(PL 25-12-30).D	Blank			1902	ND
PL-01.D	Sample			1638	ND
PL-02.D	Sample			1683	ND
MBK.D	Blank			1552	ND
CCV.D	QC	6.851	162245	1073	0.192

PCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	7.145	3754	1737	0.019
VOC10 ST-02.D	Cal	7.145	5373	1607	0.025
VOC10 ST-03.D	Cal	7.145	14073	1277	0.059
VOC10 ST-04.D	Cal	7.145	22442	1265	0.089
VOC10 ST-05.D	Cal	7.145	45918	1258	0.174
VOC10 ST-06.D	Cal	7.145	104548	1162	0.414
BK-.D	Blank			1373	ND
FBK(PL 25-12-30).D	Blank			1902	ND
PL-01.D	Sample			1638	ND
PL-02.D	Sample			1683	ND
MBK.D	Blank			1552	ND
CCV.D	QC	7.150	44517	1073	0.196

Xylene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	7.931	16586	1737	0.015
VOC10 ST-02.D	Cal	7.931	26048	1607	0.021
VOC10 ST-03.D	Cal	7.931	61222	1277	0.053
VOC10 ST-04.D	Cal	7.931	106191	1265	0.089
VOC10 ST-05.D	Cal	7.931	211930	1258	0.173
VOC10 ST-06.D	Cal	7.931	462735	1162	0.402
BK-.D	Blank	0.000	0	1373	ND
FBK(PL 25-12-30).D	Blank	0.000	0	1902	ND
PL-01.D	Sample	0.000	0	1638	ND
PL-02.D	Sample	0.000	0	1683	ND
MBK.D	Blank	0.000	0	1552	ND
CCV.D	QC	7.934	186375	1073	0.178

1,1-DCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	8628
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	12978
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	34841
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	55140
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5		0.200	115765
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	119812
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	264526

DCM



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	18202
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2		0.020	17595
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	20429
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	30751
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5	x	0.200	44281
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	38888
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	77247

Chloroform



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	5002
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	8786
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	19163
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	34951
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5	x	0.200	68759
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	62838
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	144024

Carbon Chloride



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	5257
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	7416
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	20674
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	32212
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5		0.200	67473
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	70556
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	155500

1,2-DCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	2222
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	4494
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	9322
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	18319
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5	x	0.200	35711
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	33955
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	69767

Benzene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	13857
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	23310
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	53116
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	92682
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5	x	0.200	182991
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	163656
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	382018

TCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	3830
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	5901
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	14549
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	24337
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5	x	0.200	48732
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	45787
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	107034

Toluene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	14225
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	22776
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	52769
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	90883
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5	x	0.200	179678
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	162245
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	381691

PCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	3754
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	5373
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	14073
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	22442
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5	x	0.200	45918
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	44517
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	104548

Xylene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-01.D	Calibration	1	x	0.010	16586
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-02.D	Calibration	2	x	0.020	26048
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-03.D	Calibration	3	x	0.050	61222
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-04.D	Calibration	4	x	0.100	106191
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-05.D	Calibration	5		0.200	211930
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\CCV.D	QC	5	x	0.200	186375
D:\DATA\2026년\1. 2026년 1월 \VOC10(PL 12.30)\VOC10 ST-06.D	Calibration	6	x	0.400	462735

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오전 10:18:33	**Data File**	VOC10 ST-01.D
Type	Cal	**Name**	VOC10 ST-01
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.991	8628	1737	0.016 mg/L
DCM	5.148	18202	1737	0.011 mg/L
Chloroform	5.720	5002	1737	0.015 mg/L
Carbon Chloride	5.922	5257	1737	0.017 mg/L
1,2-DCE	5.986	2222	1737	0.011 mg/L
Benzene	5.993	13857	1737	0.015 mg/L
TCE	6.232	3830	1737	0.017 mg/L
Toluene	6.851	14225	1737	0.016 mg/L
PCE	7.145	3754	1737	0.019 mg/L
Xylene	7.931	16586	1737	0.015 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-01.D
Counts
x10^3
Area=3754
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 76.0 (83.5 %)
Ratio = 47.7 (101.5 %)
Acquisition Time (min)


+ SIM (7.103-7.192 min, 16 scans) (**) VOC
Counts
x10^3
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-01.D
Counts
x10^3
Area=11718
Area=4868
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 214.8 (85.9 %)
Ratio = 31.4 (78.5 %)
Ratio = 226.9 (87.3 %)
Ratio = 62.8 (83.8 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^3
91.0
106.0
77.0
129.0
166.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오전 10:31:10	**Data File**	VOC10 ST-02.D
Type	Cal	**Name**	VOC10 ST-02
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	12978	1607	0.022 mg/L
DCM	5.153	17595	1607	0.014 mg/L
Chloroform	5.720	8786	1607	0.023 mg/L
Carbon Chloride	5.922	7416	1607	0.021 mg/L
1,2-DCE	5.986	4494	1607	0.022 mg/L
Benzene	5.993	23310	1607	0.023 mg/L
TCE	6.232	5901	1607	0.024 mg/L
Toluene	6.851	22776	1607	0.023 mg/L
PCE	7.145	5373	1607	0.025 mg/L
Xylene	7.931	26048	1607	0.021 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-02.D
Counts
x10^3
Area=5373
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 76.2 (83.7 %)
Ratio = 47.5 (101.1 %)
Acquisition Time (min)


+ SIM (7.097-7.196 min, 19 scans) (**) VOC
Counts
x10^3
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-02.D
Counts
x10^4
Area=18141
Area=7907
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 216.6 (86.6 %)
Ratio = 31.6 (79.1 %)
Ratio = 229.5 (88.3 %)
Ratio = 66.2 (88.2 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^3
91.0
106.0
77.0
129.0
166.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오전 10:43:57	**Data File**	VOC10 ST-03.D
Type	Cal	**Name**	VOC10 ST-03
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.991	34841	1277	0.055 mg/L
DCM	5.153	20429	1277	0.050 mg/L
Chloroform	5.720	19163	1277	0.054 mg/L
Carbon Chloride	5.922	20674	1277	0.056 mg/L
1,2-DCE	5.987	9322	1277	0.052 mg/L
Benzene	5.993	53116	1277	0.056 mg/L
TCE	6.232	14549	1277	0.058 mg/L
Toluene	6.851	52769	1277	0.056 mg/L
PCE	7.145	14073	1277	0.059 mg/L
Xylene	7.931	61222	1277	0.053 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-03.D
Counts
x10^3
Area=14073
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 76.6 (84.2 %)
Ratio = 47.7 (101.5 %)
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.098-7.199 min, 19 scans) (**) VOC
Counts
x10^3
166.0
129.0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-03.D
Counts
x10^4
Area=43192
Area=18030
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 218.3 (87.3 %)
Ratio = 32.2 (80.4 %)
Ratio = 230.0 (88.5 %)
Ratio = 64.7 (86.2 %)
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^3
91.0
106.0
77.0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오전 10:56:44	**Data File**	VOC10 ST-04.D
Type	Cal	**Name**	VOC10 ST-04
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	55140	1265	0.083 mg/L
DCM	5.153	30751	1265	0.108 mg/L
Chloroform	5.720	34951	1265	0.095 mg/L
Carbon Chloride	5.922	32212	1265	0.083 mg/L
1,2-DCE	5.986	18319	1265	0.100 mg/L
Benzene	5.993	92682	1265	0.095 mg/L
TCE	6.232	24337	1265	0.092 mg/L
Toluene	6.851	90883	1265	0.094 mg/L
PCE	7.145	22442	1265	0.089 mg/L
Xylene	7.931	106191	1265	0.089 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-04.D
Counts
x10^4
Area=22442
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 76.6 (84.2 %)
Ratio = 47.7 (101.5 %)
Acquisition Time (min)


+ SIM (7.097-7.203 min, 20 scans) (**) VOC
Counts
x10^3
166.0
129.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-04.D
Counts
x10^4
Area=73917
Area=32274
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 211.8 (84.7 %)
Ratio = 32.5 (81.2 %)
Ratio = 231.7 (89.1 %)
Ratio = 67.5 (90.0 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^3
91.0
106.0
77.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오전 11:09:21	**Data File**	VOC10 ST-05.D
Type	Cal	**Name**	VOC10 ST-05
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.991	115765	1258	0.168 mg/L
DCM	5.153	44281	1258	0.185 mg/L
Chloroform	5.720	68759	1258	0.183 mg/L
Carbon Chloride	5.922	67473	1258	0.166 mg/L
1,2-DCE	5.986	35711	1258	0.192 mg/L
Benzene	5.993	182991	1258	0.183 mg/L
TCE	6.232	48732	1258	0.178 mg/L
Toluene	6.851	179678	1258	0.181 mg/L
PCE	7.145	45918	1258	0.174 mg/L
Xylene	7.931	211930	1258	0.173 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-05.D
Counts
x10^4
Area=45918
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 76.7 (84.3 %)
Ratio = 47.6 (101.3 %)
Acquisition Time (min)


+ SIM (7.097-7.234 min, 26 scans) (**) VOC
Counts
x10^3
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-05.D
Counts
x10^4
Area=147539
Area=64390
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 219.8 (87.9 %)
Ratio = 32.7 (81.8 %)
Ratio = 232.0 (89.2 %)
Ratio = 67.9 (90.5 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^4
91.0
106.0
77.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오전 11:22:00	**Data File**	VOC10 ST-06.D
Type	Cal	**Name**	VOC10 ST-06
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	264526	1162	0.403 mg/L
DCM	5.153	77247	1162	0.405 mg/L
Chloroform	5.720	144024	1162	0.409 mg/L
Carbon Chloride	5.922	155500	1162	0.403 mg/L
1,2-DCE	5.986	69767	1162	0.404 mg/L
Benzene	5.993	382018	1162	0.409 mg/L
TCE	6.232	107034	1162	0.412 mg/L
Toluene	6.851	381691	1162	0.410 mg/L
PCE	7.145	104548	1162	0.414 mg/L
Xylene	7.931	462735	1162	0.402 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-06.D
Counts
x10^4
Area=104548
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 76.9 (84.5 %)
Ratio = 47.6 (101.3 %)
Acquisition Time (min)


+ SIM (7.097-7.241 min, 27 scans) (**) VOC
Counts
x10^4
166.0
129.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-06.D
Counts
x10^5
Area=324026
Area=138709
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 219.9 (88.0 %)
Ratio = 32.8 (82.1 %)
Ratio = 232.6 (89.5 %)
Ratio = 67.7 (90.3 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^4
91.0
106.0
77.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오전 11:47:16	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			1373	ND mg/L
DCM	5.153	8440	1373	0.000 mg/L
Chloroform			1373	ND mg/L
Carbon Chloride			1373	ND mg/L
1,2-DCE			1373	ND mg/L
Benzene			1373	ND mg/L
TCE			1373	ND mg/L
Toluene			1373	ND mg/L
PCE			1373	ND mg/L
Xylene	0.000	0	1373	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) BK-.D
Counts
x10^1
6
5.8
5.6
5.4
5.2
5
4.8
4.6
4.4
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.047-7.247 min, 38 scans) (**) BK-.D
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene BK-.D
Counts
x10^2
0.8
0.7
0.6
0.5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
1.6
1.4
1.2
1
0.8
0.6
0.4
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) BK-.
Counts
x10^1
77.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오전 11:59:54	**Data File**	FBK(PL 25-12-30).D
Type	Blank	**Name**	FBK(PL 25-12-30)
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			1902	ND mg/L
DCM	5.153	11761	1902	0.000 mg/L
Chloroform			1902	ND mg/L
Carbon Chloride			1902	ND mg/L
1,2-DCE			1902	ND mg/L
Benzene			1902	ND mg/L
TCE			1902	ND mg/L
Toluene			1902	ND mg/L
PCE			1902	ND mg/L
Xylene	0.000	0	1902	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) FBK(PL 25-12-30).D
Counts
x10^1
5.4
5.2
5
4.8
4.6
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.047-7.247 min, 38 scans) (**) FBK(
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene FBK(PL 25-12-30).D
Counts
x10^1
7.5
7
6.5
6
5.5
5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
1.2
1.1
1
0.9
0.8
0.7
0.6
0.5
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) FBK
Counts
x10^1
77.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오후 12:12:33	**Data File**	PL-01.D
Type	Sample	**Name**	PL-01
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			1638	ND mg/L
DCM	5.147	11579	1638	0.000 mg/L
Chloroform			1638	ND mg/L
Carbon Chloride			1638	ND mg/L
1,2-DCE			1638	ND mg/L
Benzene			1638	ND mg/L
TCE			1638	ND mg/L
Toluene			1638	ND mg/L
PCE			1638	ND mg/L
Xylene	0.000	0	1638	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) PL-01.D
Counts
x10^1
5.1
5
4.9
4.8
4.7
4.6
4.5
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.047-7.247 min, 38 scans) (**) PL-0
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene PL-01.D
Counts
x10^1
7
6.5
6
5.5
5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) PL-(
Counts
x10^1
77.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오후 12:25:10	**Data File**	PL-02.D
Type	Sample	**Name**	PL-02
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			1683	ND mg/L
DCM	5.153	47474	1683	0.136 mg/L
Chloroform			1683	ND mg/L
Carbon Chloride			1683	ND mg/L
1,2-DCE			1683	ND mg/L
Benzene			1683	ND mg/L
TCE			1683	ND mg/L
Toluene			1683	ND mg/L
PCE			1683	ND mg/L
Xylene	0.000	0	1683	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) PL-02.D
Counts
x10^1
5
4.9
4.8
4.7
4.6
4.5
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.047-7.247 min, 38 scans) (**) PL-02
Counts
x10^1
5
4
3
2
1
0
129.0
166.0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene PL-02.D
Counts
x10^1
6.5
6.25
6
5.75
5.5
5.25
5
4.75
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) PL-0
Counts
x10^1
5
4
3
2
1
0
77.0
106.0
129.0
166.0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오후 12:37:48	**Data File**	MBK.D
Type	Blank	**Name**	MBK
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			1552	ND mg/L
DCM	5.148	9090	1552	0.000 mg/L
Chloroform			1552	ND mg/L
Carbon Chloride			1552	ND mg/L
1,2-DCE	5.967	0	1552	ND mg/L
Benzene			1552	ND mg/L
TCE			1552	ND mg/L
Toluene			1552	ND mg/L
PCE			1552	ND mg/L
Xylene	0.000	0	1552	ND mg/L

Agilent Trusted Answers

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) MBK.D
Counts
x10^1
6
5.75
5.5
5.25
5
4.75
4.5
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.047-7.247 min, 37 scans) (**) MBK.
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene MBK.D
Counts
x10^2
0.9
0.8
0.7
0.6
0.5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
1.6
1.4
1.2
1
0.8
0.6
0.4
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) MBK
Counts
x10^1
91.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(PL 12.30)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-02 오후 1:22:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:58:12	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-02 오후 1:22:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-01 오후 12:50:27	**Data File**	CCV.D
Type	QC	**Name**	CCV
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.990	119812	1073	0.202 mg/L
DCM	5.147	38888	1073	0.192 mg/L
Chloroform	5.720	62838	1073	0.196 mg/L
Carbon Chloride	5.922	70556	1073	0.202 mg/L
1,2-DCE	5.986	33955	1073	0.214 mg/L
Benzene	5.993	163656	1073	0.192 mg/L
TCE	6.232	45787	1073	0.195 mg/L
Toluene	6.851	162245	1073	0.192 mg/L
PCE	7.150	44517	1073	0.196 mg/L
Xylene	7.934	186375	1073	0.178 mg/L

1,1-DCE



DCM



Chloroform



Carbon Chloride



1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) CCV.D
Counts
x10^4
Area=44517
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 77.3 (85.0 %)
Ratio = 47.7 (101.4 %)
Acquisition Time (min)


+ SIM (7.102-7.208 min, 20 scans) (**) CCV.
Counts
x10^3
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene CCV.D
Counts
x10^4
Area=130967
Area=55407
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 229.8 (91.9 %)
Ratio = 34.7 (86.8 %)
Ratio = 244.6 (94.1 %)
Ratio = 71.6 (95.4 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) CCV
Counts
x10^4
91.0
106.0
77.0
Mass-to-Charge (m/z)