Quantitative Analysis Results Summary Rep

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin				
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

1,1-DCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	4.996	13189	2202	0.017
VOC10 ST-02.D	Cal	4.991	31990	2346	0.026
VOC10 ST-03.D	Cal	4.996	81813	2380	0.053
VOC10 ST-04.D	Cal	4.996	168688	2568	0.093
VOC10 ST-05.D	Cal	4.996	380014	2825	0.181
VOC10 ST-06.D	Cal	4.996	819003	2602	0.411
BK-.D	Blank			3075	ND
FBK(KD 26-01-15).D	Blank			3257	ND
KD-01.D	Sample			6558	ND
KD-02.D	Sample			3338	ND
JR 26-01-19.D	Sample			2819	ND
MBK.D	Blank			3618	ND
-CCV-.D	QC	4.996	372515	2975	0.169

DCM

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.153	11728	2202	0.007
VOC10 ST-02.D	Cal	5.153	23820	2346	0.019
VOC10 ST-03.D	Cal	5.153	56550	2380	0.054
VOC10 ST-04.D	Cal	5.153	110903	2568	0.103
VOC10 ST-05.D	Cal	5.153	223092	2825	0.194
VOC10 ST-06.D	Cal	5.153	417489	2602	0.402
BK-.D	Blank	5.153	2763	3075	0.000
FBK(KD 26-01-15).D	Blank	5.153	2881	3257	0.000
KD-01.D	Sample	5.153	6081	6558	0.000
KD-02.D	Sample	5.153	3083	3338	0.000
JR 26-01-19.D	Sample	5.153	85582	2819	0.071
MBK.D	Blank	5.147	3154	3618	0.000
-CCV-.D	QC	5.153	220281	2975	0.182

Chloroform

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.720	19634	2202	0.008
VOC10 ST-02.D	Cal	5.720	32828	2346	0.020
VOC10 ST-03.D	Cal	5.720	66722	2380	0.054
VOC10 ST-04.D	Cal	5.720	124185	2568	0.102
VOC10 ST-05.D	Cal	5.720	243884	2825	0.193
VOC10 ST-06.D	Cal	5.720	451863	2602	0.402
BK-.D	Blank	5.720	11409	3075	0.000
FBK(KD 26-01-15).D	Blank	5.720	12328	3257	0.000
KD-01.D	Sample			6558	ND
KD-02.D	Sample			3338	ND
JR 26-01-19.D	Sample			2819	ND
MBK.D	Blank	5.720	14222	3618	0.000
-CCV-.D	QC	5.720	245445	2975	0.184

Carbon Chloride

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.922	14317	2202	0.013
VOC10 ST-02.D	Cal	5.922	34255	2346	0.025
VOC10 ST-03.D	Cal	5.922	84147	2380	0.054
VOC10 ST-04.D	Cal	5.922	166771	2568	0.096
VOC10 ST-05.D	Cal	5.922	357253	2825	0.183
VOC10 ST-06.D	Cal	5.922	741876	2602	0.408
BK-.D	Blank			3075	ND
FBK(KD 26-01-15).D	Blank			3257	ND

Carbon Chloride

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
KD-01.D	Sample			6558	ND
KD-02.D	Sample			3338	ND
JR 26-01-19.D	Sample			2819	ND
MBK.D	Blank			3618	ND
-CCV-.D	QC	5.922	349988	2975	0.171

1,2-DCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.986	4525	2202	0.010
VOC10 ST-02.D	Cal	5.986	9886	2346	0.021
VOC10 ST-03.D	Cal	5.987	25456	2380	0.052
VOC10 ST-04.D	Cal	5.986	52796	2568	0.101
VOC10 ST-05.D	Cal	5.986	111338	2825	0.193
VOC10 ST-06.D	Cal	5.986	214042	2602	0.403
BK-.D	Blank			3075	ND
FBK(KD 26-01-15).D	Blank			3257	ND
KD-01.D	Sample			6558	ND
KD-02.D	Sample			3338	ND
JR 26-01-19.D	Sample			2819	ND
MBK.D	Blank			3618	ND
-CCV-.D	QC	5.986	108228	2975	0.178

Benzene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.999	15540	2202	0.017
VOC10 ST-02.D	Cal	5.999	36633	2346	0.025
VOC10 ST-03.D	Cal	5.999	97880	2380	0.051
VOC10 ST-04.D	Cal	5.999	216663	2568	0.093
VOC10 ST-05.D	Cal	5.999	503740	2825	0.186
VOC10 ST-06.D	Cal	5.999	1052280	2602	0.408
BK-.D	Blank			3075	ND
FBK(KD 26-01-15).D	Blank			3257	ND
KD-01.D	Sample			6558	ND
KD-02.D	Sample			3338	ND
JR 26-01-19.D	Sample			2819	ND
MBK.D	Blank			3618	ND
-CCV-.D	QC	5.999	502102	2975	0.176

TCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	6.238	5321	2202	0.009
VOC10 ST-02.D	Cal	6.238	12583	2346	0.020
VOC10 ST-03.D	Cal	6.238	32505	2380	0.050
VOC10 ST-04.D	Cal	6.238	69184	2568	0.099
VOC10 ST-05.D	Cal	6.238	156334	2825	0.204
VOC10 ST-06.D	Cal	6.238	281101	2602	0.398
BK-.D	Blank			3075	ND
FBK(KD 26-01-15).D	Blank			3257	ND
KD-01.D	Sample			6558	ND
KD-02.D	Sample			3338	ND
JR 26-01-19.D	Sample			2819	ND
MBK.D	Blank			3618	ND
-CCV-.D	QC	6.238	156165	2975	0.193

Toluene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	6.854	11051	2202	0.023
VOC10 ST-02.D	Cal	6.854	25632	2346	0.030
VOC10 ST-03.D	Cal	6.855	68400	2380	0.050
VOC10 ST-04.D	Cal	6.854	155058	2568	0.086
VOC10 ST-05.D	Cal	6.858	389668	2825	0.174
VOC10 ST-06.D	Cal	6.854	911024	2602	0.415
BK-.D	Blank			3075	ND

Toluene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
FBK(KD 26-01-15).D	Blank			3257	ND
KD-01.D	Sample			6558	ND
KD-02.D	Sample	6.855	20136	3338	0.025
JR 26-01-19.D	Sample	6.854	20036	2819	0.026
MBK.D	Blank			3618	ND
-CCV-.D	QC	6.858	396939	2975	0.169

PCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	7.150	6141	2202	0.015
VOC10 ST-02.D	Cal	7.150	14638	2346	0.025
VOC10 ST-03.D	Cal	7.150	37250	2380	0.052
VOC10 ST-04.D	Cal	7.150	78224	2568	0.095
VOC10 ST-05.D	Cal	7.150	174323	2825	0.185
VOC10 ST-06.D	Cal	7.150	362630	2602	0.408
BK-.D	Blank			3075	ND
FBK(KD 26-01-15).D	Blank			3257	ND
KD-01.D	Sample			6558	ND
KD-02.D	Sample			3338	ND
JR 26-01-19.D	Sample			2819	ND
MBK.D	Blank			3618	ND
-CCV-.D	QC	7.150	172734	2975	0.174

Xylene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	7.937	10134	2202	0.025
VOC10 ST-02.D	Cal	7.937	24717	2346	0.029
VOC10 ST-03.D	Cal	7.939	75240	2380	0.045
VOC10 ST-04.D	Cal	7.942	219081	2568	0.084
VOC10 ST-05.D	Cal	7.939	636680	2825	0.187
VOC10 ST-06.D	Cal	7.939	1377811	2602	0.410
BK-.D	Blank	0.000	0	3075	ND
FBK(KD 26-01-15).D	Blank	0.000	0	3257	ND
KD-01.D	Sample	0.000	0	6558	ND
KD-02.D	Sample	0.000	0	3338	ND
JR 26-01-19.D	Sample	7.725	151024	2819	0.061
MBK.D	Blank	0.000	0	3618	ND
-CCV-.D	QC	7.939	644259	2975	0.181

1,1-DCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	13189
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	31990
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	81813
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	168688
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	380014
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	372515
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	819003

DCM



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	11728
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	23820
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	56550
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	110903
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	223092
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	220281
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	417489

Chloroform



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	19634
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	32828
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	66722
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	124185
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	243884
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	245445
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	451863

Carbon Chloride



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	14317
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	34255
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	84147
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	166771
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	357253
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	349988
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	741876

1,2-DCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	4525
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	9886
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	25456
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	52796
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	111338
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	108228
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	214042

Benzene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	15540
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	36633
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	97880
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	216663
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	503740
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	502102
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	1052280

TCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	5321
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	12583
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	32505
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	69184
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	156334
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	156165
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	281101

Toluene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	11051
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	25632
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	68400
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	155058
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	389668
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	396939
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	911024

PCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	6141
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	14638
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	37250
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	78224
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	174323
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	172734
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	362630

Xylene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-01.D	Calibration	1	x	0.010	10134
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-02.D	Calibration	2	x	0.020	24717
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-03.D	Calibration	3	x	0.050	75240
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-04.D	Calibration	4	x	0.100	219081
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-05.D	Calibration	5	x	0.200	636680
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\-CCV-.D	QC	5	x	0.200	644259
D:\DATA\2026년\1. 2026년 1월 \VOC10(KD 1.15)\VOC10 ST-06.D	Calibration	6	x	0.400	1377811

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 8:55:55	**Data File**	VOC10 ST-01.D
Type	Cal	**Name**	VOC10 ST-01
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	13189	2202	0.017 mg/L
DCM	5.153	11728	2202	0.007 mg/L
Chloroform	5.720	19634	2202	0.008 mg/L
Carbon Chloride	5.922	14317	2202	0.013 mg/L
1,2-DCE	5.986	4525	2202	0.010 mg/L
Benzene	5.999	15540	2202	0.017 mg/L
TCE	6.238	5321	2202	0.009 mg/L
Toluene	6.854	11051	2202	0.023 mg/L
PCE	7.150	6141	2202	0.015 mg/L
Xylene	7.937	10134	2202	0.025 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-01.D
Counts
x10^3
Area=6141
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 79.8 (87.7 %)
Ratio = 47.6 (101.2 %)
Acquisition Time (min)


+ SIM (7.102-7.202 min, 19 scans) (**) VOC
Counts
x10^3
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-01.D
Counts
x10^3
Area=7137
Area=2997
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 214.3 (85.7 %)
Ratio = 33.7 (84.2 %)
Ratio = 225.7 (86.8 %)
Ratio = 62.2 (83.0 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^2
91.0
106.0
77.0
129.0
166.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 9:08:45	**Data File**	VOC10 ST-02.D
Type	Cal	**Name**	VOC10 ST-02
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.991	31990	2346	0.026 mg/L
DCM	5.153	23820	2346	0.019 mg/L
Chloroform	5.720	32828	2346	0.020 mg/L
Carbon Chloride	5.922	34255	2346	0.025 mg/L
1,2-DCE	5.986	9886	2346	0.021 mg/L
Benzene	5.999	36633	2346	0.025 mg/L
TCE	6.238	12583	2346	0.020 mg/L
Toluene	6.854	25632	2346	0.030 mg/L
PCE	7.150	14638	2346	0.025 mg/L
Xylene	7.937	24717	2346	0.029 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-02.D
Counts
x10^3
Area=14638
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 80.0 (87.9 %)
Ratio = 47.6 (101.3 %)
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.102-7.204 min, 20 scans) (**) VOC
Counts
x10^3
166.0
129.0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-02.D
Counts
x10^4
Area=17422
Area=7295
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 214.8 (85.9 %)
Ratio = 33.6 (83.9 %)
Ratio = 226.8 (87.2 %)
Ratio = 62.4 (83.2 %)
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^3
91.0
106.0
77.0
129.0
166.0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 9:21:36	**Data File**	VOC10 ST-03.D
Type	Cal	**Name**	VOC10 ST-03
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	81813	2380	0.053 mg/L
DCM	5.153	56550	2380	0.054 mg/L
Chloroform	5.720	66722	2380	0.054 mg/L
Carbon Chloride	5.922	84147	2380	0.054 mg/L
1,2-DCE	5.987	25456	2380	0.052 mg/L
Benzene	5.999	97880	2380	0.051 mg/L
TCE	6.238	32505	2380	0.050 mg/L
Toluene	6.855	68400	2380	0.050 mg/L
PCE	7.150	37250	2380	0.052 mg/L
Xylene	7.939	75240	2380	0.045 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-03.D
Counts
x10^4
Area=37250
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 80.0 (87.9 %)
Ratio = 47.7 (101.5 %)
Acquisition Time (min)


+ SIM (7.102-7.220 min, 23 scans) (**) VOC
Counts
x10^3
166.0
129.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-03.D
Counts
x10^4
Area=53078
Area=22162
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 216.9 (86.8 %)
Ratio = 33.7 (84.1 %)
Ratio = 230.5 (88.7 %)
Ratio = 63.4 (84.5 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^3
91.0
106.0
77.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 9:34:27	**Data File**	VOC10 ST-04.D
Type	Cal	**Name**	VOC10 ST-04
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	168688	2568	0.093 mg/L
DCM	5.153	110903	2568	0.103 mg/L
Chloroform	5.720	124185	2568	0.102 mg/L
Carbon Chloride	5.922	166771	2568	0.096 mg/L
1,2-DCE	5.986	52796	2568	0.101 mg/L
Benzene	5.999	216663	2568	0.093 mg/L
TCE	6.238	69184	2568	0.099 mg/L
Toluene	6.854	155058	2568	0.086 mg/L
PCE	7.150	78224	2568	0.095 mg/L
Xylene	7.942	219081	2568	0.084 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-04.D
Counts
x10^4
Area=78224
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 80.2 (88.1 %)
Ratio = 47.7 (101.5 %)
Acquisition Time (min)


+ SIM (7.102-7.241 min, 27 scans) (**) VOC
Counts
x10^4
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-04.D
Counts
x10^4
Area=158808
Area=60272
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 218.9 (87.6 %)
Ratio = 33.2 (83.0 %)
Ratio = 231.5 (89.0 %)
Ratio = 64.0 (85.4 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^4
91.0
106.0
77.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 9:47:19	**Data File**	VOC10 ST-05.D
Type	Cal	**Name**	VOC10 ST-05
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	380014	2825	0.181 mg/L
DCM	5.153	223092	2825	0.194 mg/L
Chloroform	5.720	243884	2825	0.193 mg/L
Carbon Chloride	5.922	357253	2825	0.183 mg/L
1,2-DCE	5.986	111338	2825	0.193 mg/L
Benzene	5.999	503740	2825	0.186 mg/L
TCE	6.238	156334	2825	0.204 mg/L
Toluene	6.858	389668	2825	0.174 mg/L
PCE	7.150	174323	2825	0.185 mg/L
Xylene	7.939	636680	2825	0.187 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-05.D
Counts
x10^5
Area=174323
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 80.2 (88.2 %)
Ratio = 47.7 (101.5 %)
Acquisition Time (min)


+ SIM (7.102-7.247 min, 28 scans) (**) VOC
Counts
x10^4
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-05.D
Counts
x10^5
Area=468614
Area=168066
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 217.2 (86.9 %)
Ratio = 32.4 (80.9 %)
Ratio = 231.2 (88.9 %)
Ratio = 64.5 (86.0 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^4
91.0
106.0
77.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 10:04:31	**Data File**	VOC10 ST-06.D
Type	Cal	**Name**	VOC10 ST-06
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	819003	2602	0.411 mg/L
DCM	5.153	417489	2602	0.402 mg/L
Chloroform	5.720	451863	2602	0.402 mg/L
Carbon Chloride	5.922	741876	2602	0.408 mg/L
1,2-DCE	5.986	214042	2602	0.403 mg/L
Benzene	5.999	1052280	2602	0.408 mg/L
TCE	6.238	281101	2602	0.398 mg/L
Toluene	6.854	911024	2602	0.415 mg/L
PCE	7.150	362630	2602	0.408 mg/L
Xylene	7.939	1377811	2602	0.410 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) VOC10 ST-06.D
Counts
x10^5
Area=362630
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 80.2 (88.1 %)
Ratio = 47.7 (101.5 %)
Acquisition Time (min)


+ SIM (7.102-7.247 min, 28 scans) (**) VOC
Counts
x10^4
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene VOC10 ST-06.D
Counts
x10^5
Area=1004754
Area=373057
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 215.7 (86.3 %)
Ratio = 32.1 (80.2 %)
Ratio = 230.2 (88.5 %)
Ratio = 64.3 (85.7 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) VOC
Counts
x10^4
91.0
106.0
77.0
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 10:29:50	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3075	ND mg/L
DCM	5.153	2763	3075	0.000 mg/L
Chloroform	5.720	11409	3075	0.000 mg/L
Carbon Chloride			3075	ND mg/L
1,2-DCE			3075	ND mg/L
Benzene			3075	ND mg/L
TCE			3075	ND mg/L
Toluene			3075	ND mg/L
PCE			3075	ND mg/L
Xylene	0.000	0	3075	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) BK-.D
Counts
x10^1
8
7.5
7
6.5
6
5.5
5
4.5
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.050-7.250 min, 37 scans) (**) BK-.D
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene BK-.D
Counts
x10^2
0.9
0.8
0.7
0.6
0.5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
2.5
2
1.5
1
0.5
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) BK-.
Counts
x10^1
77.0
91.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 10:42:30	**Data File**	FBK(KD 26-01-15).D
Type	Blank	**Name**	FBK(KD 26-01-15)
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3257	ND mg/L
DCM	5.153	2881	3257	0.000 mg/L
Chloroform	5.720	12328	3257	0.000 mg/L
Carbon Chloride			3257	ND mg/L
1,2-DCE			3257	ND mg/L
Benzene			3257	ND mg/L
TCE			3257	ND mg/L
Toluene			3257	ND mg/L
PCE			3257	ND mg/L
Xylene	0.000	0	3257	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) FBK(KD 26-01-15).D
Counts
x10^1
6.5
6
5.5
5
4.5
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.050-7.250 min, 37 scans) (**) FBK(
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene FBK(KD 26-01-15).D
Counts
x10^1
7.5
7
6.5
6
5.5
5
4.5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
2.25
2
1.75
1.5
1.25
1
0.75
0.5
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) FBK
Counts
x10^1
77.0
91.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 10:55:10	**Data File**	KD-01.D
Type	Sample	**Name**	KD-01
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			6558	ND mg/L
DCM	5.153	6081	6558	0.000 mg/L
Chloroform			6558	ND mg/L
Carbon Chloride			6558	ND mg/L
1,2-DCE			6558	ND mg/L
Benzene			6558	ND mg/L
TCE			6558	ND mg/L
Toluene			6558	ND mg/L
PCE			6558	ND mg/L
Xylene	0.000	0	6558	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) KD-01.D
Counts
x10^1
6.25
6
5.75
5.5
5.25
5
4.75
4.5
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.050-7.250 min, 37 scans) (**) KD-0
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene KD-01.D
Counts
x10^1
6.5
6
5.5
5
4.5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
3
2.5
2
1.5
1
0.5
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) KD-
Counts
x10^1
77.0
91.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 11:07:51	**Data File**	KD-02.D
Type	Sample	**Name**	KD-02
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3338	ND mg/L
DCM	5.153	3083	3338	0.000 mg/L
Chloroform			3338	ND mg/L
Carbon Chloride			3338	ND mg/L
1,2-DCE			3338	ND mg/L
Benzene			3338	ND mg/L
TCE			3338	ND mg/L
Toluene	6.855	20136	3338	0.025 mg/L
PCE			3338	ND mg/L
Xylene	0.000	0	3338	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) KD-02.D
Counts
x10^1
6
5.8
5.6
5.4
5.2
5
4.8
4.6
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.050-7.250 min, 37 scans) (**) KD-0
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene KD-02.D
Counts
x10^2
0.9
0.8
0.7
0.6
0.5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
3
2.5
2
1.5
1
0.5
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) KD-
Counts
x10^1
77.0
91.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 11:20:32	**Data File**	JR 26-01-19.D
Type	Sample	**Name**	JR 26-01-19
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			2819	ND mg/L
DCM	5.153	85582	2819	0.071 mg/L
Chloroform			2819	ND mg/L
Carbon Chloride			2819	ND mg/L
1,2-DCE			2819	ND mg/L
Benzene			2819	ND mg/L
TCE			2819	ND mg/L
Toluene	6.854	20036	2819	0.026 mg/L
PCE			2819	ND mg/L
Xylene	7.725	151024	2819	0.061 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) JR 26-01-19.D
Counts
x10^1
5.6
5.4
5.2
5
4.8
4.6
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.050-7.250 min, 37 scans) (**) JR 26
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene JR 26-01-19.D
Counts
x10^5
Area=151024
0.8
0.6
0.4
0.2
0
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^3
Ratio = 341.9 (136.8 %)
Ratio = 38.7 (96.7 %)
Not Found
Not Found
3
2.5
2
1.5
1
0.5
0
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) JR 2
Counts
x10^4
91.0
106.0
77.0
0.8
0.6
0.4
0.2
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 11:33:12	**Data File**	MBK.D
Type	Blank	**Name**	MBK
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3618	ND mg/L
DCM	5.147	3154	3618	0.000 mg/L
Chloroform	5.720	14222	3618	0.000 mg/L
Carbon Chloride			3618	ND mg/L
1,2-DCE			3618	ND mg/L
Benzene			3618	ND mg/L
TCE			3618	ND mg/L
Toluene			3618	ND mg/L
PCE			3618	ND mg/L
Xylene	0.000	0	3618	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) MBK.D
Counts
x10^1
5.6
5.4
5.2
5
4.8
4.6
7.1
7.15
7.2
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Not Found
Not Found
1.1
1
0.9
0.8
0.7
0.6
0.5
0.4
7.1
7.15
7.2
Acquisition Time (min)


+ SIM (7.050-7.250 min, 37 scans) (**) MBK.
Counts
x10^1
129.0
166.0
5
4
3
2
1
0
130
140
150
160
Mass-to-Charge (m/z)

Xylene


Xylene MBK.D
Counts
x10^1
7
6.5
6
5.5
5
4.5
7.4
7.6
7.8
8
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Not Found
Not Found
Not Found
Not Found
2.5
2
1.5
1
0.5
7.6
7.8
8
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) MBK
Counts
x10^1
77.0
91.0
106.0
129.0
166.0
5
4
3
2
1
0
80
100
120
140
160
Mass-to-Charge (m/z)

Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC10(KD 1.15)\QuantResults\VOC10.batch.bin		
Analysis Time	2026-01-20 오후 1:17:20	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 10:55:09	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-20 오후 1:17:20	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-19 오후 11:58:41	**Data File**	-CCV-.D
Type	QC	**Name**	-CCV-
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	372515	2975	0.169 mg/L
DCM	5.153	220281	2975	0.182 mg/L
Chloroform	5.720	245445	2975	0.184 mg/L
Carbon Chloride	5.922	349988	2975	0.171 mg/L
1,2-DCE	5.986	108228	2975	0.178 mg/L
Benzene	5.999	502102	2975	0.176 mg/L
TCE	6.238	156165	2975	0.193 mg/L
Toluene	6.858	396939	2975	0.169 mg/L
PCE	7.150	172734	2975	0.174 mg/L
Xylene	7.939	644259	2975	0.181 mg/L

Agilent | Trusted Answers

1,1-DCE







DCM







Chloroform







Carbon Chloride







Agilent | Trusted Answers

1,2-DCE







Benzene







TCE







Toluene







PCE


+ Selected Ion (166.0) -CCV-.D
Counts
x10^5
Area=172734
Acquisition Time (min)


166.0, 129.0, 168.0
Relative Abundance (%)
x10^2
Ratio = 80.3 (88.2 %)
Ratio = 47.7 (101.5 %)
Acquisition Time (min)


+ SIM (7.102-7.252 min, 29 scans) (**) -CCV
Counts
x10^4
129.0
166.0
Mass-to-Charge (m/z)

Xylene


Xylene -CCV-.D
Counts
x10^5
Area=473469
Area=170791
Acquisition Time (min)


106.0, 91.0, 77.0, 106.0, 91.0, 77.0
Relative Abundance (%)
x10^2
Ratio = 217.5 (87.0 %)
Ratio = 32.3 (80.7 %)
Ratio = 231.2 (88.9 %)
Ratio = 64.2 (85.6 %)
Acquisition Time (min)


+ SIM (7.235-8.235 min, 187 scans) (**) -CC
Counts
x10^4
91.0
106.0
77.0
Mass-to-Charge (m/z)