Quantitative Analysis Results Summary Rep

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin				
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

DCM

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.153	26890	6246	0.011
VOC10 ST-02.D	Cal	5.153	38931	5947	0.021
VOC10 ST-03.D	Cal	5.153	79349	6549	0.047
VOC10 ST-04.D	Cal	5.153	147363	6329	0.100
VOC10 ST-05.D	Cal	5.153	282756	6309	0.201
BK-.D	Blank	5.153	19327	9531	0.000
FBK(HC 26-01-26).D	Blank	5.153	18280	8607	0.001
FBK(YS 26-01-27).D	Blank	5.153	18776	8636	0.001
FBK(YS 26-01-28).D	Blank	5.153	19288	9098	0.001
FBK(HJ 26-01-28).D	Blank	5.153	19298	9445	0.000
HC 26-01-26.D	Sample	5.153	69255	8354	0.030
HC 26-01-27.D	Sample	5.153	14421	3554	0.010
MBK-1.D	Blank	5.153	13944	7428	0.000
CCV-1.D	QC	5.153	104255	4795	0.093
YS-01.D	Sample	5.153	227718	4552	0.225
YS-02.D	Sample	5.153	140110	5324	0.114
YS-03.D	Sample	5.153	112402	5580	0.085
YS-04.D	Sample	5.153	179637	6410	0.122
YS-05.D	Sample	5.153	59497	5884	0.038
YS-06.D	Sample	5.153	95868	5104	0.079
YS-07.D	Sample	5.153	218436	5043	0.194
YS-08.D	Sample	5.153	75258	219	1.599
YS-09.D	Sample	5.153	94067	4322	0.093
YS-10.D	Sample	5.153	120358	4279	0.122
YS-11.D	Sample	5.153	100105	3788	0.114
YS-12.D	Sample	5.153	75183	3979	0.079
YS-13.D	Sample	5.153	258854	3261	0.362
YS-14.D	Sample	5.153	75060	3054	0.106
YS-15.D	Sample	5.153	167629	2957	0.256
YS-16.D	Sample	5.153	79961	2987	0.116
YS-17.D	Sample	5.153	66150	3036	0.093
YS-18.D	Sample	5.153	223807	2463	0.416
MBK-2.D	Blank	5.153	4413	2970	0.000
CCV-2.D	QC	5.153	36588	2104	0.072
HJ-01.D	Sample	5.153	48909	3247	0.061
HJ-02.D	Sample	5.153	30757	3017	0.038
HJ-03.D	Sample	5.153	22100	2795	0.028
HJ-04.D	Sample	5.153	22315	2272	0.037
HJ-05.D	Sample	5.153	14274	2237	0.021
HJ-06.D	Sample	5.153	21199	2866	0.025
HJ-07.D	Sample	5.153	24565	2952	0.030
HJ-08.D	Sample	5.153	41093	3021	0.054
HJ-09.D	Sample	5.153	9833	2374	0.010
HJ-10.D	Sample	5.153	10838	2480	0.011
HJ-11.D	Sample	5.153	18772	2289	0.029
HJ-12.D	Sample	5.153	20521	2087	0.037
HJ-13.D	Sample	5.153	10916	1949	0.017
HJ-14.D	Sample	5.153	24378	2082	0.046
HJ-15.D	Sample	5.153	9553	1878	0.015
HJ-16.D	Sample	5.153	26898	2150	0.049
HJ-17.D	Sample	5.153	24627	1606	0.063
MBK-3.D	Blank	5.148	4799	2965	0.000
CCV-3.D	QC	5.153	42576	2257	0.079

Chloroform

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.720	19685	6246	0.011

Chloroform

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-02.D	Cal	5.720	32171	5947	0.022
VOC10 ST-03.D	Cal	5.720	71097	6549	0.047
VOC10 ST-04.D	Cal	5.720	137329	6329	0.098
VOC10 ST-05.D	Cal	5.720	275234	6309	0.201
BK-.D	Blank	5.720	13438	9531	0.003
FBK(HC 26-01-26).D	Blank	5.720	11467	8607	0.003
FBK(YS 26-01-27).D	Blank	5.720	11950	8636	0.003
FBK(YS 26-01-28).D	Blank	5.720	12775	9098	0.003
FBK(HJ 26-01-28).D	Blank	5.720	13249	9445	0.003
HC 26-01-26.D	Sample			8354	ND
HC 26-01-27.D	Sample	5.720	0	3554	ND
MBK-1.D	Blank	5.720	9511	7428	0.002
CCV-1.D	QC	5.720	105670	4795	0.100
YS-01.D	Sample			4552	ND
YS-02.D	Sample			5324	ND
YS-03.D	Sample			5580	ND
YS-04.D	Sample			6410	ND
YS-05.D	Sample			5884	ND
YS-06.D	Sample			5104	ND
YS-07.D	Sample	5.720	0	5043	ND
YS-08.D	Sample			219	ND
YS-09.D	Sample			4322	ND
YS-10.D	Sample	5.720	42667	4279	0.043
YS-11.D	Sample			3788	ND
YS-12.D	Sample			3979	ND
YS-13.D	Sample			3261	ND
YS-14.D	Sample			3054	ND
YS-15.D	Sample			2957	ND
YS-16.D	Sample			2987	ND
YS-17.D	Sample			3036	ND
YS-18.D	Sample			2463	ND
MBK-2.D	Blank			2970	ND
CCV-2.D	QC	5.720	47504	2104	0.102
HJ-01.D	Sample			3247	ND
HJ-02.D	Sample			3017	ND
HJ-03.D	Sample			2795	ND
HJ-04.D	Sample			2272	ND
HJ-05.D	Sample			2237	ND
HJ-06.D	Sample			2866	ND
HJ-07.D	Sample			2952	ND
HJ-08.D	Sample			3021	ND
HJ-09.D	Sample			2374	ND
HJ-10.D	Sample	5.720	9034	2480	0.014
HJ-11.D	Sample	5.720	51938	2289	0.103
HJ-12.D	Sample	5.720	12843	2087	0.025
HJ-13.D	Sample			1949	ND
HJ-14.D	Sample			2082	ND
HJ-15.D	Sample	5.720	11632	1878	0.026
HJ-16.D	Sample	5.720	5118	2150	0.008
HJ-17.D	Sample			1606	ND
MBK-3.D	Blank	5.720	4150	2965	0.003
CCV-3.D	QC	5.720	53974	2257	0.109

Benzene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.999	20814	6246	0.011
VOC10 ST-02.D	Cal	5.999	51496	5947	0.022
VOC10 ST-03.D	Cal	5.999	145858	6549	0.048
VOC10 ST-04.D	Cal	5.999	306253	6329	0.098
VOC10 ST-05.D	Cal	5.999	644094	6309	0.201
BK-.D	Blank			9531	ND
FBK(HC 26-01-26).D	Blank			8607	ND

Benzene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
FBK(YS 26-01-27).D	Blank			8636	ND
FBK(YS 26-01-28).D	Blank			9098	ND
FBK(HJ 26-01-28).D	Blank			9445	ND
HC 26-01-26.D	Sample			8354	ND
HC 26-01-27.D	Sample			3554	ND
MBK-1.D	Blank			7428	ND
CCV-1.D	QC	5.999	222784	4795	0.094
YS-01.D	Sample	5.980	0	4552	ND
YS-02.D	Sample			5324	ND
YS-03.D	Sample			5580	ND
YS-04.D	Sample			6410	ND
YS-05.D	Sample			5884	ND
YS-06.D	Sample			5104	ND
YS-07.D	Sample			5043	ND
YS-08.D	Sample			219	ND
YS-09.D	Sample			4322	ND
YS-10.D	Sample			4279	ND
YS-11.D	Sample			3788	ND
YS-12.D	Sample			3979	ND
YS-13.D	Sample			3261	ND
YS-14.D	Sample			3054	ND
YS-15.D	Sample			2957	ND
YS-16.D	Sample			2987	ND
YS-17.D	Sample			3036	ND
YS-18.D	Sample			2463	ND
MBK-2.D	Blank			2970	ND
CCV-2.D	QC	5.999	79212	2104	0.077
HJ-01.D	Sample	5.999	0	3247	ND
HJ-02.D	Sample			3017	ND
HJ-03.D	Sample			2795	ND
HJ-04.D	Sample	5.999	9645	2272	0.013
HJ-05.D	Sample			2237	ND
HJ-06.D	Sample			2866	ND
HJ-07.D	Sample	5.999	43421	2952	0.033
HJ-08.D	Sample			3021	ND
HJ-09.D	Sample			2374	ND
HJ-10.D	Sample			2480	ND
HJ-11.D	Sample			2289	ND
HJ-12.D	Sample			2087	ND
HJ-13.D	Sample			1949	ND
HJ-14.D	Sample			2082	ND
HJ-15.D	Sample			1878	ND
HJ-16.D	Sample			2150	ND
HJ-17.D	Sample			1606	ND
MBK-3.D	Blank			2965	ND
CCV-3.D	QC	5.999	89311	2257	0.081

DCM



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-01.D	Calibration	1	x	0.010	26890
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-02.D	Calibration	2	x	0.020	38931
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-03.D	Calibration	3	x	0.050	79349
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-04.D	Calibration	4	x	0.100	147363
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-1.D	QC	4	x	0.100	104255
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-2.D	QC	4	x	0.100	36588
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-3.D	QC	4	x	0.100	42576
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-05.D	Calibration	5	x	0.200	282756

Chloroform



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-01.D	Calibration	1	x	0.010	19685
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-02.D	Calibration	2	x	0.020	32171

Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-03.D	Calibration	3	x	0.050	71097
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-04.D	Calibration	4	x	0.100	137329
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-1.D	QC	4	x	0.100	105670
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-2.D	QC	4	x	0.100	47504
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-3.D	QC	4	x	0.100	53974
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-05.D	Calibration	5	x	0.200	275234

Benzene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-01.D	Calibration	1	x	0.010	20814
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-02.D	Calibration	2	x	0.020	51496
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-03.D	Calibration	3	x	0.050	145858
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-04.D	Calibration	4	x	0.100	306253
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-1.D	QC	4	x	0.100	222784
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-2.D	QC	4	x	0.100	79212
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\CCV-3.D	QC	4	x	0.100	89311
D:\DATA\2026년\2. 2026년 2월 \VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\VOC10 ST-05.D	Calibration	5	x	0.200	644094

Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오전 10:34:29	**Data File**	VOC10 ST-01.D
Type	Cal	**Name**	VOC10 ST-01
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	26890	6246	0.011 mg/L
Chloroform	5.720	19685	6246	0.011 mg/L
Benzene	5.999	20814	6246	0.011 mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오전 10:47:09	**Data File**	VOC10 ST-02.D
Type	Cal	**Name**	VOC10 ST-02
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	38931	5947	0.021 mg/L
Chloroform	5.720	32171	5947	0.022 mg/L
Benzene	5.999	51496	5947	0.022 mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오전 10:59:50	**Data File**	VOC10 ST-03.D
Type	Cal	**Name**	VOC10 ST-03
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	79349	6549	0.047 mg/L
Chloroform	5.720	71097	6549	0.047 mg/L
Benzene	5.999	145858	6549	0.048 mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오전 11:12:31	**Data File**	VOC10 ST-04.D
Type	Cal	**Name**	VOC10 ST-04
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	147363	6329	0.100 mg/L
Chloroform	5.720	137329	6329	0.098 mg/L
Benzene	5.999	306253	6329	0.098 mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오전 11:25:14	**Data File**	VOC10 ST-05.D
Type	Cal	**Name**	VOC10 ST-05
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	282756	6309	0.201 mg/L
Chloroform	5.720	275234	6309	0.201 mg/L
Benzene	5.999	644094	6309	0.201 mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 12:03:22	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	19327	9531	0.000 mg/L
Chloroform	5.720	13438	9531	0.003 mg/L
Benzene			9531	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 12:16:04	**Data File**	FBK(HC 26-01-26).D
Type	Blank	**Name**	FBK(HC 26-01-26)
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	18280	8607	0.001 mg/L
Chloroform	5.720	11467	8607	0.003 mg/L
Benzene			8607	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 12:28:48	**Data File**	FBK(YS 26-01-27).D
Type	Blank	**Name**	FBK(YS 26-01-27)
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	18776	8636	0.001 mg/L
Chloroform	5.720	11950	8636	0.003 mg/L
Benzene			8636	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 12:41:33	**Data File**	FBK(YS 26-01-28).D
Type	Blank	**Name**	FBK(YS 26-01-28)
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	19288	9098	0.001 mg/L
Chloroform	5.720	12775	9098	0.003 mg/L
Benzene			9098	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 12:54:16	**Data File**	FBK(HJ 26-01-28).D
Type	Blank	**Name**	FBK(HJ 26-01-28)
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	19298	9445	0.000 mg/L
Chloroform	5.720	13249	9445	0.003 mg/L
Benzene			9445	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 1:06:58	**Data File**	HC 26-01-26.D
Type	Sample	**Name**	HC 26-01-26
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	69255	8354	0.030 mg/L
Chloroform			8354	ND mg/L
Benzene			8354	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 1:19:43	**Data File**	HC 26-01-27.D
Type	Sample	**Name**	HC 26-01-27
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	14421	3554	0.010 mg/L
Chloroform	5.720	0	3554	ND mg/L
Benzene			3554	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 1:32:28	**Data File**	MBK-1.D
Type	Blank	**Name**	MBK-1
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	13944	7428	0.000 mg/L
Chloroform	5.720	9511	7428	0.002 mg/L
Benzene			7428	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 1:45:10	**Data File**	CCV-1.D
Type	QC	**Name**	CCV-1
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	104255	4795	0.093 mg/L
Chloroform	5.720	105670	4795	0.100 mg/L
Benzene	5.999	222784	4795	0.094 mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 2:10:38	**Data File**	YS-01.D
Type	Sample	**Name**	YS-01
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	227718	4552	0.225 mg/L
Chloroform			4552	ND mg/L
Benzene	5.980	0	4552	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 2:23:22	**Data File**	YS-02.D
Type	Sample	**Name**	YS-02
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	140110	5324	0.114 mg/L
Chloroform			5324	ND mg/L
Benzene			5324	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 2:36:06	**Data File**	YS-03.D
Type	Sample	**Name**	YS-03
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	112402	5580	0.085 mg/L
Chloroform			5580	ND mg/L
Benzene			5580	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 2:48:51	**Data File**	YS-04.D
Type	Sample	**Name**	YS-04
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	179637	6410	0.122 mg/L
Chloroform			6410	ND mg/L
Benzene			6410	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 3:01:34	**Data File**	YS-05.D
Type	Sample	**Name**	YS-05
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	59497	5884	0.038 mg/L
Chloroform			5884	ND mg/L
Benzene			5884	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 3:14:21	**Data File**	YS-06.D
Type	Sample	**Name**	YS-06
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	95868	5104	0.079 mg/L
Chloroform			5104	ND mg/L
Benzene			5104	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 3:27:05	**Data File**	YS-07.D
Type	Sample	**Name**	YS-07
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	218436	5043	0.194 mg/L
Chloroform	5.720	0	5043	ND mg/L
Benzene			5043	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 3:39:52	**Data File**	YS-08.D
Type	Sample	**Name**	YS-08
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	75258	219	1.599 mg/L
Chloroform			219	ND mg/L
Benzene			219	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 3:52:38	**Data File**	YS-09.D
Type	Sample	**Name**	YS-09
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	94067	4322	0.093 mg/L
Chloroform			4322	ND mg/L
Benzene			4322	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 4:05:37	**Data File**	YS-10.D
Type	Sample	**Name**	YS-10
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	120358	4279	0.122 mg/L
Chloroform	5.720	42667	4279	0.043 mg/L
Benzene			4279	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 4:18:23	**Data File**	YS-11.D
Type	Sample	**Name**	YS-11
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	100105	3788	0.114 mg/L
Chloroform			3788	ND mg/L
Benzene			3788	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 4:31:09	**Data File**	YS-12.D
Type	Sample	**Name**	YS-12
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	75183	3979	0.079 mg/L
Chloroform			3979	ND mg/L
Benzene			3979	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 4:49:39	**Data File**	YS-13.D
Type	Sample	**Name**	YS-13
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	258854	3261	0.362 mg/L
Chloroform			3261	ND mg/L
Benzene			3261	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 5:02:36	**Data File**	YS-14.D
Type	Sample	**Name**	YS-14
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	75060	3054	0.106 mg/L
Chloroform			3054	ND mg/L
Benzene			3054	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 5:15:32	**Data File**	YS-15.D
Type	Sample	**Name**	YS-15
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	167629	2957	0.256 mg/L
Chloroform			2957	ND mg/L
Benzene			2957	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 5:28:27	**Data File**	YS-16.D
Type	Sample	**Name**	YS-16
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	79961	2987	0.116 mg/L
Chloroform			2987	ND mg/L
Benzene			2987	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 5:41:20	**Data File**	YS-17.D
Type	Sample	**Name**	YS-17
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	66150	3036	0.093 mg/L
Chloroform			3036	ND mg/L
Benzene			3036	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 5:54:12	**Data File**	YS-18.D
Type	Sample	**Name**	YS-18
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	223807	2463	0.416 mg/L
Chloroform			2463	ND mg/L
Benzene			2463	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 6:07:08	**Data File**	MBK-2.D
Type	Blank	**Name**	MBK-2
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	4413	2970	0.000 mg/L
Chloroform			2970	ND mg/L
Benzene			2970	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 6:20:01	**Data File**	CCV-2.D
Type	QC	**Name**	CCV-2
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	36588	2104	0.072 mg/L
Chloroform	5.720	47504	2104	0.102 mg/L
Benzene	5.999	79212	2104	0.077 mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 6:45:44	**Data File**	HJ-01.D
Type	Sample	**Name**	HJ-01
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	48909	3247	0.061 mg/L
Chloroform			3247	ND mg/L
Benzene	5.999	0	3247	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 6:58:36	**Data File**	HJ-02.D
Type	Sample	**Name**	HJ-02
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	30757	3017	0.038 mg/L
Chloroform			3017	ND mg/L
Benzene			3017	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 7:11:28	**Data File**	HJ-03.D
Type	Sample	**Name**	HJ-03
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	22100	2795	0.028 mg/L
Chloroform			2795	ND mg/L
Benzene			2795	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 7:24:23	**Data File**	HJ-04.D
Type	Sample	**Name**	HJ-04
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	22315	2272	0.037 mg/L
Chloroform			2272	ND mg/L
Benzene	5.999	9645	2272	0.013 mg/L

DCM







Generated at 오전 11:02 on 2026-04-10

Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 7:37:12	**Data File**	HJ-05.D
Type	Sample	**Name**	HJ-05
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	14274	2237	0.021 mg/L
Chloroform			2237	ND mg/L
Benzene			2237	ND mg/L

DCM







Generated at 오전 11:02 on 2026-04-10

Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 7:50:07	**Data File**	HJ-06.D
Type	Sample	**Name**	HJ-06
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	21199	2866	0.025 mg/L
Chloroform			2866	ND mg/L
Benzene			2866	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 8:02:59	**Data File**	HJ-07.D
Type	Sample	**Name**	HJ-07
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	24565	2952	0.030 mg/L
Chloroform			2952	ND mg/L
Benzene	5.999	43421	2952	0.033 mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 8:15:49	**Data File**	HJ-08.D
Type	Sample	**Name**	HJ-08
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	41093	3021	0.054 mg/L
Chloroform			3021	ND mg/L
Benzene			3021	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 8:28:40	**Data File**	HJ-09.D
Type	Sample	**Name**	HJ-09
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	9833	2374	0.010 mg/L
Chloroform			2374	ND mg/L
Benzene			2374	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 8:41:34	**Data File**	HJ-10.D
Type	Sample	**Name**	HJ-10
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	10838	2480	0.011 mg/L
Chloroform	5.720	9034	2480	0.014 mg/L
Benzene			2480	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 8:54:28	**Data File**	HJ-11.D
Type	Sample	**Name**	HJ-11
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	18772	2289	0.029 mg/L
Chloroform	5.720	51938	2289	0.103 mg/L
Benzene			2289	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 9:07:17	**Data File**	HJ-12.D
Type	Sample	**Name**	HJ-12
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	20521	2087	0.037 mg/L
Chloroform	5.720	12843	2087	0.025 mg/L
Benzene			2087	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 9:20:10	**Data File**	HJ-13.D
Type	Sample	**Name**	HJ-13
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	10916	1949	0.017 mg/L
Chloroform			1949	ND mg/L
Benzene			1949	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 9:33:04	**Data File**	HJ-14.D
Type	Sample	**Name**	HJ-14
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	24378	2082	0.046 mg/L
Chloroform			2082	ND mg/L
Benzene			2082	ND mg/L

DCM







Generated at 오전 11:02 on 2026-04-10

Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 9:45:53	**Data File**	HJ-15.D
Type	Sample	**Name**	HJ-15
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	9553	1878	0.015 mg/L
Chloroform	5.720	11632	1878	0.026 mg/L
Benzene			1878	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 9:58:43	**Data File**	HJ-16.D
Type	Sample	**Name**	HJ-16
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	26898	2150	0.049 mg/L
Chloroform	5.720	5118	2150	0.008 mg/L
Benzene			2150	ND mg/L

DCM







Generated at 오전 11:02 on 2026-04-10

Chloroform







Benzene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 10:11:32	**Data File**	HJ-17.D
Type	Sample	**Name**	HJ-17
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	24627	1606	0.063 mg/L
Chloroform			1606	ND mg/L
Benzene			1606	ND mg/L

DCM







Generated at 오전 11:02 on 2026-04-10

Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 10:24:21	**Data File**	MBK-3.D
Type	Blank	**Name**	MBK-3
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.148	4799	2965	0.000 mg/L
Chloroform	5.720	4150	2965	0.003 mg/L
Benzene			2965	ND mg/L

DCM







Chloroform







Benzene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\2. 2026년 2월\VOC10(HC 1.26, 1.27, YS 1.27,28, HJ 1.28)\QuantResults\B C D.batch.bin		
Analysis Time	2026-02-03 오후 2:46:36	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:02:56	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-03 오후 2:46:36	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-02 오후 10:37:11	**Data File**	CCV-3.D
Type	QC	**Name**	CCV-3
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
DCM	5.153	42576	2257	0.079 mg/L
Chloroform	5.720	53974	2257	0.109 mg/L
Benzene	5.999	89311	2257	0.081 mg/L

DCM







Chloroform







Benzene