Quantitative Analysis Results Summary Rep

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin				
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

1,1-DCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	4.996	9836	3470	0.003
VOC10 ST-02.D	Cal	4.991	28169	3531	0.017
VOC10 ST-03.D	Cal	4.996	69709	3013	0.058
VOC10 ST-04.D	Cal	4.996	121518	2910	0.109
VOC10 ST-05.D	Cal	4.996	251290	3467	0.193
VOC10 ST-06.D	Cal	4.991	389231	2618	0.401
BK-.D	Blank			3731	ND
FBK(RE 26-03-11).D	Blank			3376	ND
RE-07.D	Sample			3722	ND
RE-08.D	Sample			3864	ND
RE-17.D	Sample			3714	ND
RE-18.D	Sample			3529	ND
RE-19.D	Sample			3198	ND
RE-20.D	Sample			3463	ND
RE-21.D	Sample			3684	ND
RE-22.D	Sample			3312	ND
RE-23.D	Sample			3325	ND
MBK.D	Blank			3201	ND
-CCV-.D	QC	4.996	190064	2389	0.212

DCM

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.153	150772	3470	0.010
VOC10 ST-02.D	Cal	5.148	301498	3531	0.021
VOC10 ST-03.D	Cal	5.148	658113	3013	0.056
VOC10 ST-04.D	Cal	5.153	1113725	2910	0.100
VOC10 ST-05.D	Cal	5.153	2479132	3467	0.187
VOC10 ST-06.D	Cal	5.149	4039749	2618	0.406
BK-.D	Blank			3731	ND
FBK(RE 26-03-11).D	Blank			3376	ND
RE-07.D	Sample	5.153	0	3722	ND
RE-08.D	Sample			3864	ND
RE-17.D	Sample	5.148	0	3714	ND
RE-18.D	Sample			3529	ND
RE-19.D	Sample	5.147	379889	3198	0.030
RE-20.D	Sample			3463	ND
RE-21.D	Sample			3684	ND
RE-22.D	Sample			3312	ND
RE-23.D	Sample			3325	ND
MBK.D	Blank			3201	ND
-CCV-.D	QC	5.153	1999279	2389	0.219

Chloroform

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.720	7008	3470	0.009
VOC10 ST-02.D	Cal	5.720	16042	3531	0.020
VOC10 ST-03.D	Cal	5.715	38456	3013	0.054
VOC10 ST-04.D	Cal	5.720	68699	2910	0.100
VOC10 ST-05.D	Cal	5.720	158629	3467	0.195
VOC10 ST-06.D	Cal	5.715	247848	2618	0.402
BK-.D	Blank			3731	ND
FBK(RE 26-03-11).D	Blank			3376	ND
RE-07.D	Sample	5.720	0	3722	ND
RE-08.D	Sample			3864	ND
RE-17.D	Sample	5.720	0	3714	ND
RE-18.D	Sample			3529	ND
RE-19.D	Sample			3198	ND

Chloroform

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
RE-20.D	Sample	5.720	0	3463	ND
RE-21.D	Sample			3684	ND
RE-22.D	Sample			3312	ND
RE-23.D	Sample			3325	ND
MBK.D	Blank			3201	ND
-CCV-.D	QC	5.720	112170	2389	0.200

Carbon Chloride

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.922	8255	3470	0.004
VOC10 ST-02.D	Cal	5.922	24032	3531	0.018
VOC10 ST-03.D	Cal	5.922	60208	3013	0.058
VOC10 ST-04.D	Cal	5.922	105791	2910	0.107
VOC10 ST-05.D	Cal	5.922	222932	3467	0.192
VOC10 ST-06.D	Cal	5.922	350225	2618	0.402
BK-.D	Blank			3731	ND
FBK(RE 26-03-11).D	Blank			3376	ND
RE-07.D	Sample			3722	ND
RE-08.D	Sample			3864	ND
RE-17.D	Sample			3714	ND
RE-18.D	Sample			3529	ND
RE-19.D	Sample			3198	ND
RE-20.D	Sample			3463	ND
RE-21.D	Sample			3684	ND
RE-22.D	Sample			3312	ND
RE-23.D	Sample			3325	ND
MBK.D	Blank			3201	ND
-CCV-.D	QC	5.922	176398	2389	0.220

1,2-DCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.980	3395	3470	0.010
VOC10 ST-02.D	Cal	5.980	8217	3531	0.022
VOC10 ST-03.D	Cal	5.980	19443	3013	0.057
VOC10 ST-04.D	Cal	5.980	33979	2910	0.101
VOC10 ST-05.D	Cal	5.980	73493	3467	0.183
VOC10 ST-06.D	Cal	5.980	124420	2618	0.407
BK-.D	Blank			3731	ND
FBK(RE 26-03-11).D	Blank			3376	ND
RE-07.D	Sample			3722	ND
RE-08.D	Sample			3864	ND
RE-17.D	Sample			3714	ND
RE-18.D	Sample			3529	ND
RE-19.D	Sample			3198	ND
RE-20.D	Sample			3463	ND
RE-21.D	Sample			3684	ND
RE-22.D	Sample			3312	ND
RE-23.D	Sample			3325	ND
MBK.D	Blank			3201	ND
-CCV-.D	QC	5.980	57471	2389	0.207

Benzene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	5.993	15075	3470	0.008
VOC10 ST-02.D	Cal	5.993	37984	3531	0.019
VOC10 ST-03.D	Cal	5.993	92751	3013	0.055
VOC10 ST-04.D	Cal	5.993	168323	2910	0.103
VOC10 ST-05.D	Cal	5.993	375622	3467	0.193
VOC10 ST-06.D	Cal	5.993	590501	2618	0.402
BK-.D	Blank			3731	ND
FBK(RE 26-03-11).D	Blank			3376	ND
RE-07.D	Sample			3722	ND
RE-08.D	Sample			3864	ND
RE-17.D	Sample			3714	ND

Benzene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
RE-18.D	Sample			3529	ND
RE-19.D	Sample			3198	ND
RE-20.D	Sample			3463	ND
RE-21.D	Sample			3684	ND
RE-22.D	Sample			3312	ND
RE-23.D	Sample			3325	ND
MBK.D	Blank			3201	ND
-CCV-.D	QC	5.993	256722	2389	0.192

TCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	6.232	4655	3470	0.006
VOC10 ST-02.D	Cal	6.232	12247	3531	0.018
VOC10 ST-03.D	Cal	6.232	30171	3013	0.055
VOC10 ST-04.D	Cal	6.232	54908	2910	0.105
VOC10 ST-05.D	Cal	6.232	120315	3467	0.195
VOC10 ST-06.D	Cal	6.232	186114	2618	0.401
BK-.D	Blank			3731	ND
FBK(RE 26-03-11).D	Blank			3376	ND
RE-07.D	Sample			3722	ND
RE-08.D	Sample			3864	ND
RE-17.D	Sample			3714	ND
RE-18.D	Sample			3529	ND
RE-19.D	Sample			3198	ND
RE-20.D	Sample			3463	ND
RE-21.D	Sample			3684	ND
RE-22.D	Sample			3312	ND
RE-23.D	Sample			3325	ND
MBK.D	Blank			3201	ND
-CCV-.D	QC	6.232	82767	2389	0.195

PCE

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
VOC10 ST-01.D	Cal	7.145	4526	3470	0.005
VOC10 ST-02.D	Cal	7.145	12310	3531	0.018
VOC10 ST-03.D	Cal	7.145	30282	3013	0.055
VOC10 ST-04.D	Cal	7.145	55380	2910	0.107
VOC10 ST-05.D	Cal	7.145	118742	3467	0.193
VOC10 ST-06.D	Cal	7.145	185049	2618	0.401
BK-.D	Blank			3731	ND
FBK(RE 26-03-11).D	Blank			3376	ND
RE-07.D	Sample			3722	ND
RE-08.D	Sample			3864	ND
RE-17.D	Sample			3714	ND
RE-18.D	Sample			3529	ND
RE-19.D	Sample			3198	ND
RE-20.D	Sample			3463	ND
RE-21.D	Sample			3684	ND
RE-22.D	Sample			3312	ND
RE-23.D	Sample			3325	ND
MBK.D	Blank			3201	ND
-CCV-.D	QC	7.145	83851	2389	0.198

1,1-DCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-01.D	Calibration	1	x	0.010	9836
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-02.D	Calibration	2	x	0.020	28169
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-03.D	Calibration	3	x	0.050	69709
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-04.D	Calibration	4	x	0.100	121518
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-05.D	Calibration	5	x	0.200	251290
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\-CCV-.D	QC	5	x	0.200	190064
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-06.D	Calibration	6	x	0.400	389231

DCM



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-01.D	Calibration	1	x	0.010	150772
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-02.D	Calibration	2	x	0.020	301498
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-03.D	Calibration	3	x	0.050	658113
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-04.D	Calibration	4	x	0.100	1113725
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-05.D	Calibration	5	x	0.200	2479132
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\-CCV-.D	QC	5	x	0.200	1999279
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-06.D	Calibration	6	x	0.400	4039749

Chloroform



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-01.D	Calibration	1	x	0.010	7008
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-02.D	Calibration	2	x	0.020	16042
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-03.D	Calibration	3	x	0.050	38456
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-04.D	Calibration	4	x	0.100	68699
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-05.D	Calibration	5	x	0.200	158629
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\-CCV-.D	QC	5	x	0.200	112170
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-06.D	Calibration	6	x	0.400	247848

Carbon Chloride



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-01.D	Calibration	1	x	0.010	8255
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-02.D	Calibration	2	x	0.020	24032
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-03.D	Calibration	3	x	0.050	60208
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-04.D	Calibration	4	x	0.100	105791
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-05.D	Calibration	5	x	0.200	222932
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\-CCV-.D	QC	5	x	0.200	176398
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-06.D	Calibration	6	x	0.400	350225

1,2-DCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-01.D	Calibration	1	x	0.010	3395
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-02.D	Calibration	2	x	0.020	8217
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-03.D	Calibration	3	x	0.050	19443
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-04.D	Calibration	4	x	0.100	33979
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-05.D	Calibration	5	x	0.200	73493
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\-CCV-.D	QC	5	x	0.200	57471
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-06.D	Calibration	6	x	0.400	124420

Benzene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-01.D	Calibration	1	x	0.010	15075
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-02.D	Calibration	2	x	0.020	37984
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-03.D	Calibration	3	x	0.050	92751
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-04.D	Calibration	4	x	0.100	168323
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-05.D	Calibration	5	x	0.200	375622
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\-CCV-.D	QC	5	x	0.200	256722
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-06.D	Calibration	6	x	0.400	590501

TCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-01.D	Calibration	1	x	0.010	4655
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-02.D	Calibration	2	x	0.020	12247
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-03.D	Calibration	3	x	0.050	30171
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-04.D	Calibration	4	x	0.100	54908
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-05.D	Calibration	5	x	0.200	120315
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\-CCV-.D	QC	5	x	0.200	82767
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-06.D	Calibration	6	x	0.400	186114

PCE



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-01.D	Calibration	1	x	0.010	4526
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-02.D	Calibration	2	x	0.020	12310
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-03.D	Calibration	3	x	0.050	30282
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-04.D	Calibration	4	x	0.100	55380
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-05.D	Calibration	5	x	0.200	118742
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\-CCV-.D	QC	5	x	0.200	83851
D:\DATA\2026년\3. 2026년 3월 \VOC10(RE 3.11)\VOC10 ST-06.D	Calibration	6	x	0.400	185049

Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 6:16:47	**Data File**	VOC10 ST-01.D
Type	Cal	**Name**	VOC10 ST-01
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	9836	3470	0.003 mg/L
DCM	5.153	150772	3470	0.010 mg/L
Chloroform	5.720	7008	3470	0.009 mg/L
Carbon Chloride	5.922	8255	3470	0.004 mg/L
1,2-DCE	5.980	3395	3470	0.010 mg/L
Benzene	5.993	15075	3470	0.008 mg/L
TCE	6.232	4655	3470	0.006 mg/L
PCE	7.145	4526	3470	0.005 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 6:29:36	**Data File**	VOC10 ST-02.D
Type	Cal	**Name**	VOC10 ST-02
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.991	28169	3531	0.017 mg/L
DCM	5.148	301498	3531	0.021 mg/L
Chloroform	5.720	16042	3531	0.020 mg/L
Carbon Chloride	5.922	24032	3531	0.018 mg/L
1,2-DCE	5.980	8217	3531	0.022 mg/L
Benzene	5.993	37984	3531	0.019 mg/L
TCE	6.232	12247	3531	0.018 mg/L
PCE	7.145	12310	3531	0.018 mg/L

1,1-DCE







Generated at 오전 11:13 on 2026-04-10

DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 6:42:25	**Data File**	VOC10 ST-03.D
Type	Cal	**Name**	VOC10 ST-03
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	69709	3013	0.058 mg/L
DCM	5.148	658113	3013	0.056 mg/L
Chloroform	5.715	38456	3013	0.054 mg/L
Carbon Chloride	5.922	60208	3013	0.058 mg/L
1,2-DCE	5.980	19443	3013	0.057 mg/L
Benzene	5.993	92751	3013	0.055 mg/L
TCE	6.232	30171	3013	0.055 mg/L
PCE	7.145	30282	3013	0.055 mg/L

1,1-DCE







Generated at 오전 11:13 on 2026-04-10

Agilent Trusted Answers

DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 6:55:14	**Data File**	VOC10 ST-04.D
Type	Cal	**Name**	VOC10 ST-04
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	121518	2910	0.109 mg/L
DCM	5.153	1113725	2910	0.100 mg/L
Chloroform	5.720	68699	2910	0.100 mg/L
Carbon Chloride	5.922	105791	2910	0.107 mg/L
1,2-DCE	5.980	33979	2910	0.101 mg/L
Benzene	5.993	168323	2910	0.103 mg/L
TCE	6.232	54908	2910	0.105 mg/L
PCE	7.145	55380	2910	0.107 mg/L

1,1-DCE







Generated at 오전 11:13 on 2026-04-10

Agilent Trusted Answers

DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 7:08:03	**Data File**	VOC10 ST-05.D
Type	Cal	**Name**	VOC10 ST-05
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	251290	3467	0.193 mg/L
DCM	5.153	2479132	3467	0.187 mg/L
Chloroform	5.720	158629	3467	0.195 mg/L
Carbon Chloride	5.922	222932	3467	0.192 mg/L
1,2-DCE	5.980	73493	3467	0.183 mg/L
Benzene	5.993	375622	3467	0.193 mg/L
TCE	6.232	120315	3467	0.195 mg/L
PCE	7.145	118742	3467	0.193 mg/L

1,1-DCE







Generated at 오전 11:13 on 2026-04-10

DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 7:20:55	**Data File**	VOC10 ST-06.D
Type	Cal	**Name**	VOC10 ST-06
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.991	389231	2618	0.401 mg/L
DCM	5.149	4039749	2618	0.406 mg/L
Chloroform	5.715	247848	2618	0.402 mg/L
Carbon Chloride	5.922	350225	2618	0.402 mg/L
1,2-DCE	5.980	124420	2618	0.407 mg/L
Benzene	5.993	590501	2618	0.402 mg/L
TCE	6.232	186114	2618	0.401 mg/L
PCE	7.145	185049	2618	0.401 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 7:46:45	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3731	ND mg/L
DCM			3731	ND mg/L
Chloroform			3731	ND mg/L
Carbon Chloride			3731	ND mg/L
1,2-DCE			3731	ND mg/L
Benzene			3731	ND mg/L
TCE			3731	ND mg/L
PCE			3731	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 7:59:34	**Data File**	FBK(RE 26-03-11).D
Type	Blank	**Name**	FBK(RE 26-03-11)
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3376	ND mg/L
DCM			3376	ND mg/L
Chloroform			3376	ND mg/L
Carbon Chloride			3376	ND mg/L
1,2-DCE			3376	ND mg/L
Benzene			3376	ND mg/L
TCE			3376	ND mg/L
PCE			3376	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 8:12:26	**Data File**	RE-07.D
Type	Sample	**Name**	RE-07
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3722	ND mg/L
DCM	5.153	0	3722	ND mg/L
Chloroform	5.720	0	3722	ND mg/L
Carbon Chloride			3722	ND mg/L
1,2-DCE			3722	ND mg/L
Benzene			3722	ND mg/L
TCE			3722	ND mg/L
PCE			3722	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 8:25:19	**Data File**	RE-08.D
Type	Sample	**Name**	RE-08
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3864	ND mg/L
DCM			3864	ND mg/L
Chloroform			3864	ND mg/L
Carbon Chloride			3864	ND mg/L
1,2-DCE			3864	ND mg/L
Benzene			3864	ND mg/L
TCE			3864	ND mg/L
PCE			3864	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 8:38:09	**Data File**	RE-17.D
Type	Sample	**Name**	RE-17
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3714	ND mg/L
DCM	5.148	0	3714	ND mg/L
Chloroform	5.720	0	3714	ND mg/L
Carbon Chloride			3714	ND mg/L
1,2-DCE			3714	ND mg/L
Benzene			3714	ND mg/L
TCE			3714	ND mg/L
PCE			3714	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 8:51:02	**Data File**	RE-18.D
Type	Sample	**Name**	RE-18
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3529	ND mg/L
DCM			3529	ND mg/L
Chloroform			3529	ND mg/L
Carbon Chloride			3529	ND mg/L
1,2-DCE			3529	ND mg/L
Benzene			3529	ND mg/L
TCE			3529	ND mg/L
PCE			3529	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 9:04:07	**Data File**	RE-19.D
Type	Sample	**Name**	RE-19
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3198	ND mg/L
DCM	5.147	379889	3198	0.030 mg/L
Chloroform			3198	ND mg/L
Carbon Chloride			3198	ND mg/L
1,2-DCE			3198	ND mg/L
Benzene			3198	ND mg/L
TCE			3198	ND mg/L
PCE			3198	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 9:16:55	**Data File**	RE-20.D
Type	Sample	**Name**	RE-20
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3463	ND mg/L
DCM			3463	ND mg/L
Chloroform	5.720	0	3463	ND mg/L
Carbon Chloride			3463	ND mg/L
1,2-DCE			3463	ND mg/L
Benzene			3463	ND mg/L
TCE			3463	ND mg/L
PCE			3463	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 9:29:45	**Data File**	RE-21.D
Type	Sample	**Name**	RE-21
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3684	ND mg/L
DCM			3684	ND mg/L
Chloroform			3684	ND mg/L
Carbon Chloride			3684	ND mg/L
1,2-DCE			3684	ND mg/L
Benzene			3684	ND mg/L
TCE			3684	ND mg/L
PCE			3684	ND mg/L

1,1-DCE







Generated at 오전 11:13 on 2026-04-10

DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 9:42:55	**Data File**	RE-22.D
Type	Sample	**Name**	RE-22
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3312	ND mg/L
DCM			3312	ND mg/L
Chloroform			3312	ND mg/L
Carbon Chloride			3312	ND mg/L
1,2-DCE			3312	ND mg/L
Benzene			3312	ND mg/L
TCE			3312	ND mg/L
PCE			3312	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 9:55:44	**Data File**	RE-23.D
Type	Sample	**Name**	RE-23
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3325	ND mg/L
DCM			3325	ND mg/L
Chloroform			3325	ND mg/L
Carbon Chloride			3325	ND mg/L
1,2-DCE			3325	ND mg/L
Benzene			3325	ND mg/L
TCE			3325	ND mg/L
PCE			3325	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 10:08:37	**Data File**	MBK.D
Type	Blank	**Name**	MBK
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE			3201	ND mg/L
DCM			3201	ND mg/L
Chloroform			3201	ND mg/L
Carbon Chloride			3201	ND mg/L
1,2-DCE			3201	ND mg/L
Benzene			3201	ND mg/L
TCE			3201	ND mg/L
PCE			3201	ND mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\VOC10(RE 3.11)\QuantResults\VOC100.batch.bin		
Analysis Time	2026-03-17 오전 9:05:25	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오전 11:13:51	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-17 오전 9:05:25	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-12 오후 10:34:34	**Data File**	-CCV-.D
Type	QC	**Name**	-CCV-
Dil.	1	**Acq. Method File**	2025 VOCs(10) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
1,1-DCE	4.996	190064	2389	0.212 mg/L
DCM	5.153	1999279	2389	0.219 mg/L
Chloroform	5.720	112170	2389	0.200 mg/L
Carbon Chloride	5.922	176398	2389	0.220 mg/L
1,2-DCE	5.980	57471	2389	0.207 mg/L
Benzene	5.993	256722	2389	0.192 mg/L
TCE	6.232	82767	2389	0.195 mg/L
PCE	7.145	83851	2389	0.198 mg/L

1,1-DCE







DCM







Chloroform







Carbon Chloride







1,2-DCE







Benzene







TCE







PCE