Operator: *Lab.:*

Results

Results file: C:\Users\Public\Documents\Analytik Jena\ASpectPQ\ICP\Results\251201_VGA_TJ,IS-25-12-01 152948.tps

Instrument: PQ 9000 #13585100362BA0517 *Technique:* ICP-OES

Operator:

Comment:

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
method blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:30
As193.698						
Conc.1	0				mg/L	
Ints.	117	9	7.71			
Single values (Ints.): #1: 107 #2: 120 #3: 124						
Sb217.581						
Conc.1	0				mg/L	
Ints.	-81	41	50.99			
Single values (Ints.): #1: -60 #2: -54 #3: -128						
Se196.028						
Conc.1	0				mg/L	
Ints.	278	22	8.03			
Single values (Ints.): #1: 253 #2: 295 #3: 286						
ST1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:32
As193.698						
Conc.1	0.02				mg/L	
Ints.	507	45	8.92			
Single values (Ints.): #1: 503 #2: 463 #3: 554						
Sb217.581						
Conc.1	0.02				mg/L	
Ints.	137	14	10.10			
Single values (Ints.): #1: 143 #2: 121 #3: 146						
Se196.028						
Conc.1	0.02				mg/L	
Ints.	687	23	3.31			
Single values (Ints.): #1: 699 #2: 701 #3: 660						
ST2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:34
As193.698						
Conc.1	0.04				mg/L	
Ints.	918	39	4.27			
Single values (Ints.): #1: 924 #2: 954 #3: 876						
Sb217.581						
Conc.1	0.04				mg/L	
Ints.	444	19	4.23			
Single values (Ints.): #1: 464 #2: 441 #3: 426						
Se196.028						
Conc.1	0.04				mg/L	
Ints.	1014	45	4.46			
Single values (Ints.): #1: 992 #2: 984 #3: 1066						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
ST3	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:35
As193.698						
Conc.1	0.05				mg/L	
Ints.	1146	28	2.46			
Single values (Ints.): #1: 1171 #2: 1150 #3: 1115						
Sb217.581						
Conc.1	0.05				mg/L	
Ints.	592	52	8.73			
Single values (Ints.): #1: 539 #2: 594 #3: 642						
Se196.028						
Conc.1	0.05				mg/L	
Ints.	1194	20	1.67			
Single values (Ints.): #1: 1216 #2: 1191 #3: 1176						
ST4	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:37
As193.698						
Conc.1	0.1				mg/L	
Ints.	2067	36	1.76			
Single values (Ints.): #1: 2075 #2: 2100 #3: 2028						
Sb217.581						
Conc.1	0.1				mg/L	
Ints.	1264	25	1.95			
Single values (Ints.): #1: 1292 #2: 1247 #3: 1252						
Se196.028						
Conc.1	0.1				mg/L	
Ints.	2043	47	2.32			
Single values (Ints.): #1: 1988 #2: 2063 #3: 2076						
ST5	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:39
As193.698						
Conc.1	0.2				mg/L	
Ints.	4134	21	0.51			
Single values (Ints.): #1: 4127 #2: 4158 #3: 4118						
Sb217.581						
Conc.1	0.2				mg/L	
Ints.	2680	34	1.27			
Single values (Ints.): #1: 2674 #2: 2717 #3: 2650						
Se196.028						
Conc.1	0.2				mg/L	
Ints.	3888	27	0.71			
Single values (Ints.): #1: 3872 #2: 3920 #3: 3873						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*

Compute calib. *Method:* As193.698 *Date:* 2025-12-01 15:39

As193.698

y=a+bx a=112.98318 b=20026.528 Method SD: 0.0015061 mg/L

R²(adj.): 0.999429197 ; Slope: 20026.528 Ints./(mg/L) ; BEC: 0.4140190 mg/L



Compute calib. *Method:* Sb217.581 *Date:* 2025-12-01 15:39

Sb217.581

y=a+bx a=-111.3676 b=13910.830 Method SD: 0.0017029 mg/L

R²(adj.): 0.999270332 ; Slope: 13910.830 Ints./(mg/L) ; BEC: 1.1672760 mg/L



Compute calib. *Method:* Se196.028 *Date:* 2025-12-01 15:39

Se196.028

y=a+bx a=294.71161 b=17892.563 Method SD: 0.0016358 mg/L

R²(adj.): 0.999326678 ; Slope: 17892.563 Ints./(mg/L) ; BEC: 0.2310587 mg/L



Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
BLANK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:40
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0006	0.0008	148.40	0.0032	mg/L	
Conc.2	0.0087	0.0008	9.50	0.0032	mg/L	
Ints.	124	17	13.40			
Single values (Ints.): #1: 143 #2: 114 #3: 115						
Sb217.581						
Conc.1	0.0141	0.0021	15.23	0.0034	mg/L	
Conc.2	0.0065	0.0021	33.02	0.0034	mg/L	
Ints.	85	30	35.18			
Single values (Ints.): #1: 66 #2: 119 #3: 69						
Se196.028						
Conc.1	0.0012	0.0005	42.55	0.0035	mg/L	
Conc.2	-0.0043	0.0005	12.15	0.0035	mg/L	
Ints.	317	9	2.98			
Single values (Ints.): #1: 326 #2: 318 #3: 307						
Reag. blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:42
As193.698						
Conc.1	0.0000	0.0013	###.##	0.0032	mg/L	
Ints.	113	25	22.48			
Single values (Ints.): #1: 115 #2: 138 #3: 87						
Sb217.581						
Conc.1	0.0068	0.0005	6.91	0.0035	mg/L	
Ints.	-16	7	40.15			
Single values (Ints.): #1: -17 #2: -10 #3: -23						
Se196.028						
Conc.1	-0.0009	0.0020	230.62	0.0035	mg/L	<KAL
Ints.	279	37	13.11			<KAL
Single values (Ints.): #1: 237 #2: 297 #3: 303						
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:44
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0005	0.0009	172.62	0.0032	mg/L	
Conc.2	0.0005	0.0009	179.60	0.0032	mg/L	
Ints.	123	18	14.41			
Single values (Ints.): #1: 136 #2: 130 #3: 103						
Sb217.581						
Conc.1	0.0064	0.0018	27.96	0.0036	mg/L	
Conc.2	-0.0005	0.0018	384.56	0.0036	mg/L	
Ints.	-23	25	108.65			
Single values (Ints.): #1: -28 #2: 4 #3: -44						
Se196.028						
Conc.1	0.0003	0.0001	41.36	0.0035	mg/L	
Conc.2	0.0011	0.0001	9.12	0.0035	mg/L	
Ints.	299	2	0.62			
Single values (Ints.): #1: 297 #2: 300 #3: 300						

Operator: *Lab.:*

Line	Mean	SD	RSD[%]	CI	Unit	Rem.
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:45
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0008	0.0011	147.93	0.0032	mg/L	
Conc.2	0.0007	0.0011	151.89	0.0032	mg/L	
Ints.	128	23	17.68			
Single values (Ints.): #1: 133 #2: 149 #3: 104						
Sb217.581						
Conc.1	0.0057	0.0012	20.35	0.0036	mg/L	
Conc.2	-0.0012	0.0012	98.20	0.0036	mg/L	
Ints.	-33	16	49.03			
Single values (Ints.): #1: -17 #2: -33 #3: -49						
Se196.028						
Conc.1	0.0007	0.0006	86.72	0.0035	mg/L	
Conc.2	0.0015	0.0006	37.09	0.0035	mg/L	
Ints.	307	10	3.35			
Single values (Ints.): #1: 309 #2: 295 #3: 316						
FBK(11.12)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:47
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0005	0.0016	295.57	0.0032	mg/L	<KAL
Conc.2	0.0005	0.0016	307.12	0.0032	mg/L	<KAL
Ints.	124	31	25.45			<KAL
Single values (Ints.): #1: 90 #2: 153 #3: 128						
Sb217.581						
Conc.1	0.0063	0.0010	16.15	0.0036	mg/L	
Conc.2	-0.0005	0.0010	197.41	0.0036	mg/L	
Ints.	-24	14	60.27			
Single values (Ints.): #1: -40 #2: -15 #3: -16						
Se196.028						
Conc.1	-0.0011	0.0009	86.15	0.0035	mg/L	<KAL
Conc.2	-0.0002	0.0009	446.46	0.0035	mg/L	<KAL
Ints.	275	17	6.15			<KAL
Single values (Ints.): #1: 289 #2: 256 #3: 280						
FBK(11.13)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:49
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0001	0.0009	###.##	0.0032	mg/L	<KAL
Conc.2	0.0001	0.0009	###.##	0.0032	mg/L	<KAL
Ints.	115	19	16.16			<KAL
Single values (Ints.): #1: 129 #2: 94 #3: 121						
Sb217.581						
Conc.1	0.0046	0.0033	70.04	0.0036	mg/L	
Conc.2	-0.0022	0.0033	149.17	0.0036	mg/L	
Ints.	-47	45	96.94			
Single values (Ints.): #1: -62 #2: -83 #3: 4						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0017	0.0003	20.33	0.0035	mg/L	
Conc.2	0.0026	0.0003	13.42	0.0035	mg/L	
Ints.	325	6	1.92			
Single values (Ints.): #1: 329 #2: 318 #3: 329						
FBK(11.14)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:50
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0005	0.0012	247.56	0.0032	mg/L	<KAL
Conc.2	0.0005	0.0012	258.09	0.0032	mg/L	<KAL
Ints.	123	24	19.76			<KAL
Single values (Ints.): #1: 107 #2: 110 #3: 151						
Sb217.581						
Conc.1	0.0043	0.0028	64.55	0.0036	mg/L	
Conc.2	-0.0026	0.0028	108.15	0.0036	mg/L	
Ints.	-52	38	74.05			
Single values (Ints.): #1: -64 #2: -9 #3: -83						
Se196.028						
Conc.1	-0.0007	0.0017	235.65	0.0035	mg/L	<KAL
Conc.2	0.0002	0.0017	###.##	0.0035	mg/L	<KAL
Ints.	282	31	10.96			<KAL
Single values (Ints.): #1: 264 #2: 264 #3: 317						
FBK(11.18)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:52
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0004	0.0013	353.37	0.0032	mg/L	<KAL
Conc.2	0.0003	0.0013	374.18	0.0032	mg/L	<KAL
Ints.	120	25	21.15			<KAL
Single values (Ints.): #1: 118 #2: 96 #3: 147						
Sb217.581						
Conc.1	0.0045	0.0008	16.89	0.0036	mg/L	
Conc.2	-0.0023	0.0008	33.33	0.0036	mg/L	
Ints.	-48	11	22.05			
Single values (Ints.): #1: -40 #2: -45 #3: -60						
Se196.028						
Conc.1	-0.0008	0.0005	57.62	0.0035	mg/L	<KAL
Conc.2	0.0001	0.0005	631.86	0.0035	mg/L	<KAL
Ints.	280	8	2.99			<KAL
Single values (Ints.): #1: 286 #2: 271 #3: 284						
FBK(11.19)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:54
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0009	0.0013	145.61	0.0032	mg/L	
Conc.2	0.0009	0.0013	149.00	0.0032	mg/L	
Ints.	131	26	19.60			
Single values (Ints.): #1: 160 #2: 111 #3: 121						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0043	0.0019	44.04	0.0036	mg/L	
Conc.2	-0.0025	0.0019	74.51	0.0036	mg/L	
Ints.	-52	26	50.92			
Single values (Ints.): #1: -61 #2: -72 #3: -22						
Se196.028						
Conc.1	0.0003	0.0015	582.10	0.0035	mg/L	<KAL
Conc.2	0.0011	0.0015	131.85	0.0035	mg/L	<KAL
Ints.	299	27	9.03			<KAL
Single values (Ints.): #1: 302 #2: 271 #3: 325						
FBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:55
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0007	0.0019	263.07	0.0032	mg/L	<KAL
Conc.2	0.0007	0.0019	270.45	0.0032	mg/L	<KAL
Ints.	128	39	30.22			<KAL
Single values (Ints.): #1: 104 #2: 172 #3: 107						
Sb217.581						
Conc.1	0.0055	0.0039	70.46	0.0036	mg/L	
Conc.2	-0.0014	0.0039	284.35	0.0036	mg/L	
Ints.	-35	54	152.33			
Single values (Ints.): #1: 23 #2: -46 #3: -83						
Se196.028						
Conc.1	0.0007	0.0006	88.61	0.0035	mg/L	
Conc.2	0.0016	0.0006	39.22	0.0035	mg/L	
Ints.	307	11	3.63			
Single values (Ints.): #1: 295 #2: 310 #3: 317						
JH IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:57
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0038	0.0019	50.18	0.0033	mg/L	<KAL
Conc.2	-0.0038	0.0019	49.92	0.0033	mg/L	<KAL
Ints.	37	38	104.69			<KAL
Single values (Ints.): #1: 3 #2: 29 #3: 78						
Sb217.581						
Conc.1	0.0056	0.0020	36.38	0.0036	mg/L	
Conc.2	-0.0012	0.0020	163.31	0.0036	mg/L	
Ints.	-34	28	83.98			
Single values (Ints.): #1: -60 #2: -4 #3: -37						
Se196.028						
Conc.1	0.0154	0.0023	14.86	0.0033	mg/L	
Conc.2	0.0163	0.0023	14.05	0.0033	mg/L	
Ints.	570	41	7.18			
Single values (Ints.): #1: 523 #2: 596 #3: 593						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
JH OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 15:59
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0054	0.0010	18.23	0.0033	mg/L	<KAL
Conc.2	-0.0054	0.0010	18.16	0.0033	mg/L	<KAL
Ints.	5	20	359.21			<KAL
Single values (Ints.): #1: -3 #2: -9 #3: 28						
Sb217.581						
Conc.1	0.0049	0.0031	64.38	0.0036	mg/L	
Conc.2	-0.0020	0.0031	158.60	0.0036	mg/L	
Ints.	-44	44	99.37			
Single values (Ints.): #1: -19 #2: -18 #3: -94						
Se196.028						
Conc.1	0.0152	0.0010	6.41	0.0033	mg/L	
Conc.2	0.0161	0.0010	6.06	0.0033	mg/L	
Ints.	567	17	3.08			
Single values (Ints.): #1: 550 #2: 565 #3: 585						
CCV 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:00
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1013	0.0016	1.63	0.0028	mg/L	
Conc.2	0.1013	0.0016	1.63	0.0028	mg/L	
Ints.	2141	33	1.54			
Single values (Ints.): #1: 2114 #2: 2132 #3: 2178						
Sb217.581						
Conc.1	0.0944	0.0011	1.20	0.0031	mg/L	
Conc.2	0.0876	0.0011	1.29	0.0031	mg/L	
Ints.	1202	16	1.31			
Single values (Ints.): #1: 1219 #2: 1198 #3: 1188						
Se196.028						
Conc.1	0.0991	0.0014	1.41	0.0030	mg/L	
Conc.2	0.1000	0.0014	1.40	0.0030	mg/L	
Ints.	2068	25	1.21			
Single values (Ints.): #1: 2043 #2: 2069 #3: 2093						
WN IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:02
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0040	0.0010	24.36	0.0033	mg/L	<KAL
Conc.2	-0.0040	0.0010	24.24	0.0033	mg/L	<KAL
Ints.	33	19	58.21			<KAL
Single values (Ints.): #1: 56 #2: 23 #3: 21						
Sb217.581						
Conc.1	0.0093	0.0018	19.50	0.0035	mg/L	
Conc.2	0.0025	0.0018	73.27	0.0035	mg/L	
Ints.	18	25	139.47			
Single values (Ints.): #1: 47 #2: 7 #3: 1						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0052	0.0014	25.89	0.0034	mg/L	
Conc.2	0.0061	0.0014	22.15	0.0034	mg/L	
Ints.	389	24	6.26			
Single values (Ints.): #1: 366 #2: 414 #3: 386						
WN OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:04
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0039	0.0010	26.51	0.0033	mg/L	<KAL
Conc.2	-0.0039	0.0010	26.37	0.0033	mg/L	<KAL
Ints.	35	21	60.04			<KAL
Single values (Ints.): #1: 54 #2: 13 #3: 37						
Sb217.581						
Conc.1	0.0066	0.0046	68.51	0.0036	mg/L	
Conc.2	-0.0002	0.0046	###.##	0.0036	mg/L	
Ints.	-19	63	335.72			
Single values (Ints.): #1: -18 #2: 44 #3: -83						
Se196.028						
Conc.1	0.0043	0.0024	55.77	0.0034	mg/L	
Conc.2	0.0052	0.0024	46.22	0.0034	mg/L	
Ints.	371	43	11.52			
Single values (Ints.): #1: 396 #2: 397 #3: 322						
JJ IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:05
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0020	0.0014	69.28	0.0033	mg/L	<KAL
Conc.2	-0.0020	0.0014	68.60	0.0033	mg/L	<KAL
Ints.	73	28	38.15			<KAL
Single values (Ints.): #1: 41 #2: 85 #3: 92						
Sb217.581						
Conc.1	0.0050	0.0037	73.84	0.0036	mg/L	
Conc.2	-0.0018	0.0037	199.80	0.0036	mg/L	
Ints.	-42	51	121.99			
Single values (Ints.): #1: 17 #2: -70 #3: -73						
Se196.028						
Conc.1	0.0016	0.0011	67.03	0.0035	mg/L	
Conc.2	0.0025	0.0011	42.94	0.0035	mg/L	
Ints.	323	19	5.86			
Single values (Ints.): #1: 329 #2: 338 #3: 302						
JJ OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:07
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0034	0.0006	17.88	0.0033	mg/L	<KAL
Conc.2	-0.0034	0.0006	17.78	0.0033	mg/L	<KAL
Ints.	44	12	27.73			<KAL
Single values (Ints.): #1: 58 #2: 39 #3: 35						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0065	0.0010	14.90	0.0036	mg/L	
Conc.2	-0.0003	0.0010	296.40	0.0036	mg/L	
Ints.	-21	13	64.49			
Single values (Ints.): #1: -20 #2: -35 #3: -8						
Se196.028						
Conc.1	0.0000	0.0018	###.##	0.0035	mg/L	<KAL
Conc.2	0.0009	0.0018	207.75	0.0035	mg/L	<KAL
Ints.	295	33	11.21			<KAL
Single values (Ints.): #1: 302 #2: 259 #3: 324						
CJ IN	*Method:*	As,Sb,Se(1)	Orig.	*Date:*		2025-12-01 16:09
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0033	0.0007	21.87	0.0033	mg/L	<KAL
Conc.2	-0.0033	0.0007	21.74	0.0033	mg/L	<KAL
Ints.	48	14	30.09			<KAL
Single values (Ints.): #1: 33 #2: 48 #3: 62						
Sb217.581						
Conc.1	0.0061	0.0011	17.85	0.0036	mg/L	
Conc.2	-0.0008	0.0011	144.53	0.0036	mg/L	
Ints.	-27	15	56.33			
Single values (Ints.): #1: -13 #2: -24 #3: -43						
Se196.028						
Conc.1	0.0012	0.0028	235.80	0.0035	mg/L	
Conc.2	0.0021	0.0028	135.59	0.0035	mg/L	
Ints.	316	51	16.00			
Single values (Ints.): #1: 326 #2: 361 #3: 261						
CJ OUT	*Method:*	As,Sb,Se(1)	Orig.	*Date:*		2025-12-01 16:10
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0023	0.0027	116.70	0.0033	mg/L	<KAL
Conc.2	-0.0023	0.0027	115.69	0.0033	mg/L	<KAL
Ints.	67	53	78.93			<KAL
Single values (Ints.): #1: 111 #2: 8 #3: 83						
Sb217.581						
Conc.1	0.0035	0.0007	21.06	0.0036	mg/L	
Conc.2	-0.0033	0.0007	22.03	0.0036	mg/L	
Ints.	-63	10	16.29			
Single values (Ints.): #1: -62 #2: -73 #3: -53						
Se196.028						
Conc.1	-0.0002	0.0010	572.45	0.0035	mg/L	<KAL
Conc.2	0.0007	0.0010	140.56	0.0035	mg/L	<KAL
Ints.	292	18	6.14			<KAL
Single values (Ints.): #1: 278 #2: 285 #3: 312						

Operator: *Lab.:*

Line	Mean	SD	RSD[%]	CI	Unit	Rem.
IS IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:12
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0030	0.0022	73.27	0.0033	mg/L	<KAL
Conc.2	-0.0030	0.0022	72.79	0.0033	mg/L	<KAL
Ints.	53	44	82.58			<KAL
Single values (Ints.): #1: 101 #2: 16 #3: 42						
Sb217.581						
Conc.1	0.0223	0.0041	18.44	0.0033	mg/L	
Conc.2	0.0154	0.0041	26.60	0.0033	mg/L	
Ints.	198	57	28.80			
Single values (Ints.): #1: 153 #2: 179 #3: 262						
Se196.028						
Conc.1	-0.0003	0.0015	482.42	0.0035	mg/L	<KAL
Conc.2	0.0006	0.0015	272.61	0.0035	mg/L	<KAL
Ints.	289	28	9.55			<KAL
Single values (Ints.): #1: 311 #2: 298 #3: 258						
IS OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:14
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0035	0.0029	82.79	0.0033	mg/L	<KAL
Conc.2	-0.0035	0.0029	82.32	0.0033	mg/L	<KAL
Ints.	43	58	135.76			<KAL
Single values (Ints.): #1: 78 #2: 75 #3: -24						
Sb217.581						
Conc.1	0.0262	0.0013	4.86	0.0033	mg/L	
Conc.2	0.0194	0.0013	6.57	0.0033	mg/L	
Ints.	254	18	6.99			
Single values (Ints.): #1: 258 #2: 234 #3: 269						
Se196.028						
Conc.1	-0.0009	0.0005	59.25	0.0035	mg/L	<KAL
Conc.2	0.0000	0.0005	###.##	0.0035	mg/L	<KAL
Ints.	279	9	3.35			<KAL
Single values (Ints.): #1: 269 #2: 288 #3: 279						
ISW IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:15
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0003	0.0008	245.87	0.0032	mg/L	<KAL
Conc.2	0.0003	0.0008	262.47	0.0032	mg/L	<KAL
Ints.	119	16	13.03			<KAL
Single values (Ints.): #1: 101 #2: 128 #3: 129						
Sb217.581						
Conc.1	0.0050	0.0018	36.05	0.0036	mg/L	
Conc.2	-0.0018	0.0018	99.18	0.0036	mg/L	
Ints.	-42	25	60.27			
Single values (Ints.): #1: -53 #2: -13 #3: -59						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0017	0.0019	113.47	0.0035	mg/L	
Conc.2	0.0026	0.0019	74.09	0.0035	mg/L	
Ints.	325	34	10.43			
Single values (Ints.): #1: 298 #2: 314 #3: 362						
CCV 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:17
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1012	0.0015	1.51	0.0028	mg/L	
Conc.2	0.1012	0.0015	1.51	0.0028	mg/L	
Ints.	2140	31	1.43			
Single values (Ints.): #1: 2157 #2: 2158 #3: 2104						
Sb217.581						
Conc.1	0.0953	0.0033	3.45	0.0031	mg/L	
Conc.2	0.0885	0.0033	3.71	0.0031	mg/L	
Ints.	1215	46	3.76			
Single values (Ints.): #1: 1193 #2: 1184 #3: 1267						
Se196.028						
Conc.1	0.1000	0.0025	2.48	0.0031	mg/L	
Conc.2	0.1009	0.0025	2.46	0.0031	mg/L	
Ints.	2084	44	2.13			
Single values (Ints.): #1: 2079 #2: 2043 #3: 2131						
ISW OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:19
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0009	0.0006	68.60	0.0033	mg/L	<KAL
Conc.2	-0.0010	0.0006	67.18	0.0033	mg/L	<KAL
Ints.	94	13	13.80			<KAL
Single values (Ints.): #1: 100 #2: 103 #3: 79						
Sb217.581						
Conc.1	0.0103	0.0021	20.77	0.0035	mg/L	
Conc.2	0.0035	0.0021	61.69	0.0035	mg/L	
Ints.	32	30	93.36			
Single values (Ints.): #1: 32 #2: 62 #3: 2						
Se196.028						
Conc.1	0.0000	0.0009	###.##	0.0035	mg/L	<KAL
Conc.2	0.0009	0.0009	100.72	0.0035	mg/L	<KAL
Ints.	295	17	5.65			<KAL
Single values (Ints.): #1: 283 #2: 314 #3: 289						
DS IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:20
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0041	0.0011	26.00	0.0033	mg/L	<KAL
Conc.2	-0.0041	0.0011	25.88	0.0033	mg/L	<KAL
Ints.	31	21	69.39			<KAL
Single values (Ints.): #1: 39 #2: 47 #3: 7						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0086	0.0020	23.33	0.0035	mg/L	
Conc.2	0.0017	0.0020	115.17	0.0035	mg/L	
Ints.	8	28	357.00			
Single values (Ints.): #1: 1 #2: -16 #3: 38						
Se196.028						
Conc.1	0.0025	0.0021	83.08	0.0035	mg/L	
Conc.2	0.0034	0.0021	61.51	0.0035	mg/L	
Ints.	340	38	11.05			
Single values (Ints.): #1: 364 #2: 297 #3: 359						
DS OUT	*Method:*	As,Sb,Se(1)	Orig.	*Date:*		2025-12-01 16:22
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0029	0.0015	52.05	0.0033	mg/L	<KAL
Conc.2	-0.0030	0.0015	51.70	0.0033	mg/L	<KAL
Ints.	54	31	56.81			<KAL
Single values (Ints.): #1: 80 #2: 20 #3: 62						
Sb217.581						
Conc.1	0.0061	0.0016	25.82	0.0036	mg/L	
Conc.2	-0.0007	0.0016	213.19	0.0036	mg/L	
Ints.	-27	22	82.21			
Single values (Ints.): #1: -34 #2: -2 #3: -44						
Se196.028						
Conc.1	-0.0005	0.0007	144.12	0.0035	mg/L	<KAL
Conc.2	0.0004	0.0007	168.71	0.0035	mg/L	<KAL
Ints.	286	12	4.31			<KAL
Single values (Ints.): #1: 275 #2: 284 #3: 300						
CSJ	*Method:*	As,Sb,Se(1)	Orig.	*Date:*		2025-12-01 16:24
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0018	0.0016	88.43	0.0033	mg/L	<KAL
Conc.2	-0.0018	0.0016	87.47	0.0033	mg/L	<KAL
Ints.	77	32	42.14			<KAL
Single values (Ints.): #1: 95 #2: 95 #3: 39						
Sb217.581						
Conc.1	0.0044	0.0028	63.60	0.0036	mg/L	
Conc.2	-0.0024	0.0028	115.83	0.0036	mg/L	
Ints.	-50	39	77.97			
Single values (Ints.): #1: -24 #2: -95 #3: -31						
Se196.028						
Conc.1	0.0000	0.0029	###.##	0.0035	mg/L	
Conc.2	0.0009	0.0029	322.71	0.0035	mg/L	
Ints.	295	51	17.44			
Single values (Ints.): #1: 351 #2: 249 #3: 284						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
JRJ	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:25
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0024	0.0010	39.78	0.0033	mg/L	<KAL
Conc.2	-0.0024	0.0010	39.45	0.0033	mg/L	<KAL
Ints.	65	19	29.48			<KAL
Single values (Ints.): #1: 43 #2: 78 #3: 74						
Sb217.581						
Conc.1	0.0013	0.0012	89.27	0.0036	mg/L	
Conc.2	-0.0055	0.0012	21.92	0.0036	mg/L	
Ints.	-93	17	18.05			
Single values (Ints.): #1: -73 #2: -102 #3: -103						
Se196.028						
Conc.1	0.0009	0.0003	28.86	0.0035	mg/L	
Conc.2	0.0018	0.0003	14.51	0.0035	mg/L	
Ints.	311	5	1.49			
Single values (Ints.): #1: 310 #2: 316 #3: 306						
ASI IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:27
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0002	0.0007	412.10	0.0032	mg/L	<KAL
Conc.2	0.0001	0.0007	467.48	0.0032	mg/L	<KAL
Ints.	116	14	11.96			<KAL
Single values (Ints.): #1: 112 #2: 132 #3: 105						
Sb217.581						
Conc.1	0.0034	0.0037	110.19	0.0036	mg/L	
Conc.2	-0.0035	0.0037	106.57	0.0036	mg/L	
Ints.	-65	51	79.61			
Single values (Ints.): #1: -10 #2: -112 #3: -72						
Se196.028						
Conc.1	0.0005	0.0009	186.94	0.0035	mg/L	<KAL
Conc.2	0.0013	0.0009	63.76	0.0035	mg/L	<KAL
Ints.	303	15	5.06			<KAL
Single values (Ints.): #1: 286 #2: 316 #3: 306						
ASI OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:29
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0015	0.0009	61.63	0.0033	mg/L	<KAL
Conc.2	-0.0015	0.0009	60.83	0.0033	mg/L	<KAL
Ints.	83	19	22.69			<KAL
Single values (Ints.): #1: 61 #2: 92 #3: 95						
Sb217.581						
Conc.1	0.0052	0.0036	69.34	0.0036	mg/L	
Conc.2	-0.0016	0.0036	229.48	0.0036	mg/L	
Ints.	-38	51	131.75			
Single values (Ints.): #1: -63 #2: -72 #3: 20						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0011	0.0005	50.50	0.0035	mg/L	
Conc.2	0.0020	0.0005	27.69	0.0035	mg/L	
Ints.	314	10	3.10			
Single values (Ints.): #1: 308 #2: 325 #3: 308						
AST IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:30
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0018	0.0012	63.93	0.0033	mg/L	<KAL
Conc.2	-0.0018	0.0012	63.23	0.0033	mg/L	<KAL
Ints.	77	23	30.06			<KAL
Single values (Ints.): #1: 70 #2: 58 #3: 103						
Sb217.581						
Conc.1	0.0061	0.0025	40.44	0.0036	mg/L	
Conc.2	-0.0008	0.0025	325.18	0.0036	mg/L	
Ints.	-27	34	127.22			
Single values (Ints.): #1: -58 #2: -32 #3: 9						
Se196.028						
Conc.1	0.0041	0.0010	24.85	0.0035	mg/L	
Conc.2	0.0049	0.0010	20.40	0.0035	mg/L	
Ints.	367	18	4.92			
Single values (Ints.): #1: 384 #2: 371 #3: 348						
AST OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:32
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0005	0.0013	286.65	0.0032	mg/L	<KAL
Conc.2	-0.0005	0.0013	274.77	0.0032	mg/L	<KAL
Ints.	104	27	25.57			<KAL
Single values (Ints.): #1: 81 #2: 133 #3: 98						
Sb217.581						
Conc.1	0.0031	0.0041	130.25	0.0036	mg/L	
Conc.2	-0.0037	0.0041	110.80	0.0036	mg/L	
Ints.	-68	57	84.03			
Single values (Ints.): #1: -96 #2: -2 #3: -105						
Se196.028						
Conc.1	-0.0019	0.0008	41.83	0.0035	mg/L	<KAL
Conc.2	-0.0010	0.0008	77.55	0.0035	mg/L	<KAL
Ints.	260	14	5.53			<KAL
Single values (Ints.): #1: 264 #2: 244 #3: 272						
CCV 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:34
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1028	0.0008	0.76	0.0028	mg/L	
Conc.2	0.1028	0.0008	0.76	0.0028	mg/L	
Ints.	2171	16	0.72			
Single values (Ints.): #1: 2183 #2: 2176 #3: 2154						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0949	0.0010	1.02	0.0031	mg/L	
Conc.2	0.0881	0.0010	1.10	0.0031	mg/L	
Ints.	1209	13	1.11			
Single values (Ints.): #1: 1224 #2: 1197 #3: 1207						
Se196.028						
Conc.1	0.0991	0.0035	3.49	0.0030	mg/L	
Conc.2	0.1000	0.0035	3.46	0.0030	mg/L	
Ints.	2068	62	2.99			
Single values (Ints.): #1: 2009 #2: 2132 #3: 2063						
PT IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:35
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0056	0.0016	29.15	0.0033	mg/L	<KAL
Conc.2	-0.0056	0.0016	29.04	0.0033	mg/L	<KAL
Ints.	1	33	###.##			<KAL
Single values (Ints.): #1: 36 #2: -4 #3: -28						
Sb217.581						
Conc.1	0.0310	0.0009	2.82	0.0032	mg/L	
Conc.2	0.0241	0.0009	3.61	0.0032	mg/L	
Ints.	319	12	3.80			
Single values (Ints.): #1: 317 #2: 332 #3: 308						
Se196.028						
Conc.1	0.0177	0.0013	7.26	0.0033	mg/L	
Conc.2	0.0186	0.0013	6.92	0.0033	mg/L	
Ints.	611	23	3.76			
Single values (Ints.): #1: 634 #2: 588 #3: 611						
PT OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:37
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0049	0.0006	11.35	0.0033	mg/L	<KAL
Conc.2	-0.0049	0.0006	11.30	0.0033	mg/L	<KAL
Ints.	15	11	73.91			<KAL
Single values (Ints.): #1: 18 #2: 24 #3: 3						
Sb217.581						
Conc.1	0.0233	0.0003	1.47	0.0033	mg/L	
Conc.2	0.0165	0.0003	2.07	0.0033	mg/L	
Ints.	213	5	2.23			
Single values (Ints.): #1: 208 #2: 215 #3: 217						
Se196.028						
Conc.1	0.0180	0.0019	10.57	0.0033	mg/L	
Conc.2	0.0189	0.0019	10.07	0.0033	mg/L	
Ints.	617	34	5.52			
Single values (Ints.): #1: 582 #2: 650 #3: 619						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
PT OUT1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:39
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0048	0.0012	24.71	0.0033	mg/L	<KAL
Conc.2	-0.0048	0.0012	24.60	0.0033	mg/L	<KAL
Ints.	16	24	146.93			<KAL
Single values (Ints.): #1: 1 #2: 4 #3: 44						
Sb217.581						
Conc.1	0.0239	0.0032	13.42	0.0033	mg/L	
Conc.2	0.0171	0.0032	18.78	0.0033	mg/L	
Ints.	222	45	20.16			
Single values (Ints.): #1: 221 #2: 266 #3: 177						
Se196.028						
Conc.1	0.0171	0.0017	10.19	0.0033	mg/L	
Conc.2	0.0179	0.0017	9.69	0.0033	mg/L	
Ints.	600	31	5.19			
Single values (Ints.): #1: 596 #2: 571 #3: 633						
GM IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:41
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0015	0.0013	89.23	0.0033	mg/L	<KAL
Conc.2	-0.0015	0.0013	88.03	0.0033	mg/L	<KAL
Ints.	84	26	31.18			<KAL
Single values (Ints.): #1: 60 #2: 112 #3: 80						
Sb217.581						
Conc.1	0.0036	0.0016	44.69	0.0036	mg/L	
Conc.2	-0.0033	0.0016	48.37	0.0036	mg/L	
Ints.	-62	22	35.61			
Single values (Ints.): #1: -50 #2: -48 #3: -87						
Se196.028						
Conc.1	0.0006	0.0020	329.90	0.0035	mg/L	<KAL
Conc.2	0.0015	0.0020	134.93	0.0035	mg/L	<KAL
Ints.	306	36	11.84			<KAL
Single values (Ints.): #1: 298 #2: 274 #3: 345						
GM OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:42
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0038	0.0016	41.50	0.0033	mg/L	<KAL
Conc.2	-0.0038	0.0016	41.28	0.0033	mg/L	<KAL
Ints.	38	31	83.47			<KAL
Single values (Ints.): #1: 72 #2: 30 #3: 11						
Sb217.581						
Conc.1	0.0038	0.0028	74.39	0.0036	mg/L	
Conc.2	-0.0030	0.0028	95.02	0.0036	mg/L	
Ints.	-58	40	68.26			
Single values (Ints.): #1: -15 #2: -66 #3: -93						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	-0.0010	0.0007	74.24	0.0035	mg/L	<KAL
Conc.2	-0.0001	0.0007	###.##	0.0035	mg/L	<KAL
Ints.	278	13	4.56			<KAL
Single values (Ints.): #1: 292 #2: 273 #3: 268						
IS 1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:44
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0004	0.0005	135.91	0.0032	mg/L	
Conc.2	0.0004	0.0005	143.62	0.0032	mg/L	
Ints.	120	10	8.42			
Single values (Ints.): #1: 131 #2: 119 #3: 111						
Sb217.581						
Conc.1	0.0054	0.0030	55.20	0.0036	mg/L	
Conc.2	-0.0015	0.0030	201.70	0.0036	mg/L	
Ints.	-37	41	111.99			
Single values (Ints.): #1: -83 #2: -3 #3: -25						
Se196.028						
Conc.1	-0.0017	0.0021	120.86	0.0035	mg/L	<KAL
Conc.2	-0.0008	0.0021	251.28	0.0035	mg/L	<KAL
Ints.	264	37	13.97			<KAL
Single values (Ints.): #1: 298 #2: 270 #3: 225						
IS 2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:46
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0023	0.0001	4.52	0.0033	mg/L	<KAL
Conc.2	-0.0023	0.0001	4.48	0.0033	mg/L	<KAL
Ints.	67	2	3.06			<KAL
Single values (Ints.): #1: 70 #2: 66 #3: 67						
Sb217.581						
Conc.1	0.0063	0.0003	3.97	0.0036	mg/L	
Conc.2	-0.0005	0.0003	52.49	0.0036	mg/L	
Ints.	-23	4	15.23			
Single values (Ints.): #1: -27 #2: -20 #3: -22						
Se196.028						
Conc.1	-0.0008	0.0017	222.33	0.0035	mg/L	<KAL
Conc.2	0.0001	0.0017	###.##	0.0035	mg/L	<KAL
Ints.	281	31	10.97			<KAL
Single values (Ints.): #1: 312 #2: 281 #3: 250						
IS 3	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:47
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0000	0.0013	###.##	0.0032	mg/L	
Conc.2	0.0000	0.0013	###.##	0.0032	mg/L	
Ints.	114	26	22.79			
Single values (Ints.): #1: 119 #2: 137 #3: 86						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0038	0.0022	58.09	0.0036	mg/L	
Conc.2	-0.0031	0.0022	70.75	0.0036	mg/L	
Ints.	-59	30	51.20			
Single values (Ints.): #1: -73 #2: -80 #3: -24						
Se196.028						
Conc.1	-0.0013	0.0015	118.13	0.0035	mg/L	<KAL
Conc.2	-0.0004	0.0015	386.94	0.0035	mg/L	<KAL
Ints.	272	27	9.91			<KAL
Single values (Ints.): #1: 300 #2: 269 #3: 246						
IS 4	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:49
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0029	0.0009	32.11	0.0033	mg/L	<KAL
Conc.2	-0.0029	0.0009	31.89	0.0033	mg/L	<KAL
Ints.	56	18	33.15			<KAL
Single values (Ints.): #1: 77 #2: 45 #3: 45						
Sb217.581						
Conc.1	0.0064	0.0008	12.50	0.0036	mg/L	
Conc.2	-0.0004	0.0008	194.52	0.0036	mg/L	
Ints.	-22	11	50.52			
Single values (Ints.): #1: -12 #2: -20 #3: -34						
Se196.028						
Conc.1	-0.0008	0.0006	74.87	0.0035	mg/L	<KAL
Conc.2	0.0001	0.0006	497.10	0.0035	mg/L	<KAL
Ints.	281	10	3.67			<KAL
Single values (Ints.): #1: 292 #2: 278 #3: 273						
CCV 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:51
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1021	0.0010	0.94	0.0028	mg/L	
Conc.2	0.1020	0.0010	0.94	0.0028	mg/L	
Ints.	2157	19	0.89			
Single values (Ints.): #1: 2162 #2: 2136 #3: 2174						
Sb217.581						
Conc.1	0.0924	0.0018	1.99	0.0031	mg/L	
Conc.2	0.0856	0.0018	2.15	0.0031	mg/L	
Ints.	1174	26	2.18			
Single values (Ints.): #1: 1153 #2: 1167 #3: 1203						
Se196.028						
Conc.1	0.0997	0.0019	1.89	0.0031	mg/L	
Conc.2	0.1006	0.0019	1.87	0.0031	mg/L	
Ints.	2078	34	1.62			
Single values (Ints.): #1: 2104 #2: 2040 #3: 2090						

Operator: *Lab.:*

Line	Mean	SD	RSD[%]	CI	Unit	Rem.
IS 5	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:52
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0080	0.0012	15.30	0.0031	mg/L	
Conc.2	0.0080	0.0012	15.34	0.0031	mg/L	
Ints.	274	25	8.99			
Single values (Ints.): #1: 285 #2: 246 #3: 291						
Sb217.581						
Conc.1	0.0088	0.0005	5.60	0.0035	mg/L	
Conc.2	0.0020	0.0005	24.60	0.0035	mg/L	
Ints.	12	7	59.04			
Single values (Ints.): #1: 11 #2: 19 #3: 5						
Se196.028						
Conc.1	-0.0007	0.0006	83.58	0.0035	mg/L	<KAL
Conc.2	0.0001	0.0006	436.56	0.0035	mg/L	<KAL
Ints.	281	11	3.95			<KAL
Single values (Ints.): #1: 293 #2: 270 #3: 281						
IS 6	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:54
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0015	0.0016	105.19	0.0033	mg/L	<KAL
Conc.2	-0.0015	0.0016	103.80	0.0033	mg/L	<KAL
Ints.	83	32	37.99			<KAL
Single values (Ints.): #1: 63 #2: 119 #3: 66						
Sb217.581						
Conc.1	0.0070	0.0014	20.42	0.0035	mg/L	
Conc.2	0.0002	0.0014	721.67	0.0035	mg/L	
Ints.	-14	20	146.98			
Single values (Ints.): #1: 7 #2: -33 #3: -14						
Se196.028						
Conc.1	0.0007	0.0024	357.10	0.0035	mg/L	
Conc.2	0.0016	0.0024	154.69	0.0035	mg/L	
Ints.	307	43	14.10			
Single values (Ints.): #1: 353 #2: 268 #3: 300						
IS 7	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:56
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0013	0.0007	53.03	0.0033	mg/L	<KAL
Conc.2	-0.0013	0.0007	52.22	0.0033	mg/L	<KAL
Ints.	87	14	15.87			<KAL
Single values (Ints.): #1: 102 #2: 85 #3: 74						
Sb217.581						
Conc.1	0.0037	0.0020	56.14	0.0036	mg/L	
Conc.2	-0.0032	0.0020	64.47	0.0036	mg/L	
Ints.	-61	29	47.07			
Single values (Ints.): #1: -60 #2: -32 #3: -89						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	-0.0026	0.0010	38.26	0.0036	mg/L	<KAL
Conc.2	-0.0017	0.0010	58.02	0.0036	mg/L	<KAL
Ints.	248	18	7.18			<KAL
Single values (Ints.): #1: 245 #2: 267 #3: 232						
IS 8	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:57
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0019	0.0015	81.62	0.0032	mg/L	
Conc.2	0.0018	0.0015	82.51	0.0032	mg/L	
Ints.	150	30	20.21			
Single values (Ints.): #1: 124 #2: 143 #3: 184						
Sb217.581						
Conc.1	0.0052	0.0013	25.55	0.0036	mg/L	
Conc.2	-0.0016	0.0013	80.24	0.0036	mg/L	
Ints.	-39	18	46.85			
Single values (Ints.): #1: -58 #2: -38 #3: -21						
Se196.028						
Conc.1	-0.0012	0.0011	90.07	0.0035	mg/L	<KAL
Conc.2	-0.0003	0.0011	352.53	0.0035	mg/L	<KAL
Ints.	273	19	7.01			<KAL
Single values (Ints.): #1: 257 #2: 268 #3: 295						
IS 9	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 16:59
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0019	0.0019	99.86	0.0033	mg/L	<KAL
Conc.2	-0.0019	0.0019	98.83	0.0033	mg/L	<KAL
Ints.	75	38	51.52			<KAL
Single values (Ints.): #1: 83 #2: 33 #3: 108						
Sb217.581						
Conc.1	0.0080	0.0025	31.39	0.0035	mg/L	
Conc.2	0.0011	0.0025	218.83	0.0035	mg/L	
Ints.	-0	35	###.##			
Single values (Ints.): #1: -20 #2: -21 #3: 40						
Se196.028						
Conc.1	-0.0024	0.0017	69.66	0.0036	mg/L	<KAL
Conc.2	-0.0015	0.0017	111.09	0.0036	mg/L	<KAL
Ints.	252	30	11.74			<KAL
Single values (Ints.): #1: 226 #2: 284 #3: 246						
IS 10	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 17:01
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0000	0.0007	###.##	0.0032	mg/L	<KAL
Conc.2	-0.0000	0.0007	###.##	0.0032	mg/L	<KAL
Ints.	112	15	13.05			<KAL
Single values (Ints.): #1: 104 #2: 129 #3: 104						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0108	0.0023	21.50	0.0035	mg/L	
Conc.2	0.0039	0.0023	58.92	0.0035	mg/L	
Ints.	38	32	84.12			
Single values (Ints.): #1: 1 #2: 61 #3: 53						
Se196.028						
Conc.1	-0.0023	0.0019	83.46	0.0036	mg/L	<KAL
Conc.2	-0.0014	0.0019	135.24	0.0036	mg/L	<KAL
Ints.	253	35	13.64			<KAL
Single values (Ints.): #1: 291 #2: 246 #3: 223						
IS 11	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 17:02
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0006	0.0017	275.61	0.0032	mg/L	<KAL
Conc.2	-0.0006	0.0017	267.12	0.0032	mg/L	<KAL
Ints.	100	35	34.52			<KAL
Single values (Ints.): #1: 130 #2: 109 #3: 62						
Sb217.581						
Conc.1	0.0026	0.0018	68.21	0.0036	mg/L	
Conc.2	-0.0043	0.0018	41.26	0.0036	mg/L	
Ints.	-76	24	32.33			
Single values (Ints.): #1: -103 #2: -55 #3: -69						
Se196.028						
Conc.1	0.0039	0.0023	58.89	0.0035	mg/L	
Conc.2	0.0048	0.0023	47.92	0.0035	mg/L	
Ints.	364	41	11.21			
Single values (Ints.): #1: 317 #2: 386 #3: 389						
IS 12	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 17:04
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0014	0.0013	90.89	0.0033	mg/L	<KAL
Conc.2	-0.0014	0.0013	89.63	0.0033	mg/L	<KAL
Ints.	84	26	30.75			<KAL
Single values (Ints.): #1: 111 #2: 59 #3: 83						
Sb217.581						
Conc.1	0.0083	0.0009	10.73	0.0035	mg/L	
Conc.2	0.0015	0.0009	60.37	0.0035	mg/L	
Ints.	4	12	296.05			
Single values (Ints.): #1: 17 #2: 2 #3: -7						
Se196.028						
Conc.1	-0.0002	0.0027	###.##	0.0035	mg/L	<KAL
Conc.2	0.0007	0.0027	392.20	0.0035	mg/L	<KAL
Ints.	291	49	16.85			<KAL
Single values (Ints.): #1: 247 #2: 344 #3: 283						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
CCV 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-01 17:06
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1008	0.0005	0.45	0.0028	mg/L	
Conc.2	0.1008	0.0005	0.45	0.0028	mg/L	
Ints.	2131	9	0.42			
Single values (Ints.): #1: 2122 #2: 2140 #3: 2132						
Sb217.581						
Conc.1	0.0937	0.0006	0.68	0.0031	mg/L	
Conc.2	0.0869	0.0006	0.74	0.0031	mg/L	
Ints.	1192	9	0.75			
Single values (Ints.): #1: 1200 #2: 1194 #3: 1182						
Se196.028						
Conc.1	0.0989	0.0023	2.28	0.0030	mg/L	
Conc.2	0.0998	0.0023	2.26	0.0030	mg/L	
Ints.	2065	40	1.95			
Single values (Ints.): #1: 2094 #2: 2019 #3: 2081						