Operator: *Lab.:*

Results

Results file: C:\Users\Public\Documents\Analytik Jena\ASpectPQ\ICP\Results\251204_VGA_GH,MCC,B-25-12-04 151753.tps

Instrument: PQ 9000 #13585100362BA0517 *Technique:* ICP-OES

Operator:

Comment:

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
method blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:19
As193.698						
Conc.1	0				mg/L	
Ints.	71	26	36.31			
Single values (Ints.): #1: 77 #2: 42 #3: 92						
Sb217.581						
Conc.1	0				mg/L	
Ints.	-10	30	313.16			
Single values (Ints.): #1: -26 #2: -29 #3: 25						
Se196.028						
Conc.1	0				mg/L	
Ints.	304	29	9.54			
Single values (Ints.): #1: 277 #2: 301 #3: 334						
ST1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:21
As193.698						
Conc.1	0.02				mg/L	
Ints.	448	11	2.55			
Single values (Ints.): #1: 460 #2: 446 #3: 437						
Sb217.581						
Conc.1	0.02				mg/L	
Ints.	142	57	39.83			
Single values (Ints.): #1: 91 #2: 133 #3: 203						
Se196.028						
Conc.1	0.02				mg/L	
Ints.	599	9	1.46			
Single values (Ints.): #1: 603 #2: 605 #3: 589						
ST2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:22
As193.698						
Conc.1	0.04				mg/L	
Ints.	827	18	2.19			
Single values (Ints.): #1: 839 #2: 806 #3: 836						
Sb217.581						
Conc.1	0.04				mg/L	
Ints.	401	21	5.16			
Single values (Ints.): #1: 425 #2: 390 #3: 389						
Se196.028						
Conc.1	0.04				mg/L	
Ints.	930	26	2.82			
Single values (Ints.): #1: 946 #2: 899 #3: 944						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
ST3	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:24
As193.698						
Conc.1	0.05				mg/L	
Ints.	1045	28	2.64			
Single values (Ints.): #1: 1075 #2: 1020 #3: 1040						
Sb217.581						
Conc.1	0.05				mg/L	
Ints.	585	15	2.56			
Single values (Ints.): #1: 602 #2: 573 #3: 580						
Se196.028						
Conc.1	0.05				mg/L	
Ints.	1116	72	6.44			
Single values (Ints.): #1: 1034 #2: 1149 #3: 1166						
ST4	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:26
As193.698						
Conc.1	0.1				mg/L	
Ints.	1958	54	2.76			
Single values (Ints.): #1: 1909 #2: 1947 #3: 2016						
Sb217.581						
Conc.1	0.1				mg/L	
Ints.	1201	46	3.80			
Single values (Ints.): #1: 1239 #2: 1214 #3: 1151						
Se196.028						
Conc.1	0.1				mg/L	
Ints.	1955	17	0.89			
Single values (Ints.): #1: 1954 #2: 1973 #3: 1938						
ST5	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:27
As193.698						
Conc.1	0.2				mg/L	
Ints.	3885	29	0.74			
Single values (Ints.): #1: 3854 #2: 3910 #3: 3892						
Sb217.581						
Conc.1	0.2				mg/L	
Ints.	2511	62	2.46			
Single values (Ints.): #1: 2444 #2: 2566 #3: 2524						
Se196.028						
Conc.1	0.2				mg/L	
Ints.	3642	34	0.93			
Single values (Ints.): #1: 3609 #2: 3641 #3: 3676						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*

Compute calib. *Method:* As193.698 *Date:* 2025-12-04 15:28

As193.698

y=a+bx a=70.578249 b=19049.670 Method SD: 0.0007715 mg/L

R²(adj.): 0.999850183 ; Slope: 19049.670 Ints./(mg/L) ; BEC: 0.3569424 mg/L



Compute calib. *Method:* Sb217.581 *Date:* 2025-12-04 15:28

Sb217.581

y=a+bx a=-73.01745 b=12853.715 Method SD: 0.0034551 mg/L

R²(adj.): 0.997000237 ; Slope: 12853.715 Ints./(mg/L) ; BEC: 1.0324130 mg/L



Compute calib. *Method:* Se196.028 *Date:* 2025-12-04 15:28

Se196.028

y=a+bx a=276.38512 b=16799.323 Method SD: 0.0010863 mg/L

R²(adj.): 0.999702979 ; Slope: 16799.323 Ints./(mg/L) ; BEC: 0.2293889 mg/L



Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
BLANK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:29
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0018	0.0016	90.49	0.0016	mg/L	
Conc.2	0.0113	0.0016	14.35	0.0016	mg/L	
Ints.	105	31	29.51			
Single values (Ints.): #1: 137 #2: 75 #3: 102						
Sb217.581						
Conc.1	0.0129	0.0014	10.67	0.0070	mg/L	
Conc.2	0.0064	0.0014	21.44	0.0070	mg/L	
Ints.	92	18	19.11			
Single values (Ints.): #1: 111 #2: 76 #3: 90						
Se196.028						
Conc.1	0.0017	0.0015	89.03	0.0023	mg/L	
Conc.2	-0.0034	0.0015	44.57	0.0023	mg/L	
Ints.	305	26	8.46			
Single values (Ints.): #1: 322 #2: 319 #3: 276						
Reag. blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:31
As193.698						
Conc.1	0.0004	0.0011	318.08	0.0017	mg/L	<KAL
Ints.	77	22	27.91			<KAL
Single values (Ints.): #1: 81 #2: 54 #3: 97						
Sb217.581						
Conc.1	0.0058	0.0012	19.89	0.0072	mg/L	
Ints.	2	15	767.76			
Single values (Ints.): #1: -13 #2: 1 #3: 17						
Se196.028						
Conc.1	0.0019	0.0017	89.98	0.0023	mg/L	<KAL
Ints.	309	29	9.48			<KAL
Single values (Ints.): #1: 276 #2: 319 #3: 332						
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:32
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0006	0.0012	187.48	0.0017	mg/L	
Conc.2	0.0003	0.0012	420.98	0.0017	mg/L	
Ints.	83	23	27.70			
Single values (Ints.): #1: 106 #2: 82 #3: 60						
Sb217.581						
Conc.1	0.0055	0.0021	38.92	0.0072	mg/L	
Conc.2	-0.0003	0.0021	650.58	0.0072	mg/L	
Ints.	-2	28	###.##			
Single values (Ints.): #1: 23 #2: -32 #3: 2						
Se196.028						
Conc.1	0.0027	0.0011	39.82	0.0023	mg/L	
Conc.2	0.0008	0.0011	135.08	0.0023	mg/L	
Ints.	323	18	5.70			
Single values (Ints.): #1: 322 #2: 304 #3: 341						

Operator: *Lab.:*

Line	Mean	SD	RSD[%]	CI	Unit	Rem.
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:34
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0008	0.0015	173.42	0.0017	mg/L	
Conc.2	0.0005	0.0015	300.88	0.0017	mg/L	
Ints.	87	28	32.09			
Single values (Ints.): #1: 111 #2: 56 #3: 93						
Sb217.581						
Conc.1	0.0017	0.0007	38.72	0.0074	mg/L	
Conc.2	-0.0041	0.0007	16.26	0.0074	mg/L	
Ints.	-51	9	16.88			
Single values (Ints.): #1: -52 #2: -59 #3: -42						
Se196.028						
Conc.1	0.0013	0.0006	45.55	0.0023	mg/L	
Conc.2	-0.0006	0.0006	95.99	0.0023	mg/L	
Ints.	298	10	3.37			
Single values (Ints.): #1: 310 #2: 295 #3: 291						
FBK(GH)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:36
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0023	0.0007	29.77	0.0016	mg/L	
Conc.2	0.0019	0.0007	35.36	0.0016	mg/L	
Ints.	114	13	11.26			
Single values (Ints.): #1: 99 #2: 123 #3: 119						
Sb217.581						
Conc.1	0.0010	0.0028	296.65	0.0074	mg/L	<KAL
Conc.2	-0.0049	0.0028	57.74	0.0074	mg/L	<KAL
Ints.	-61	36	59.58			<KAL
Single values (Ints.): #1: -100 #2: -29 #3: -54						
Se196.028						
Conc.1	0.0035	0.0014	39.06	0.0023	mg/L	
Conc.2	0.0016	0.0014	87.58	0.0023	mg/L	
Ints.	335	23	6.85			
Single values (Ints.): #1: 339 #2: 310 #3: 356						
FBK(MCC)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:37
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0024	0.0030	125.78	0.0016	mg/L	<KAL
Conc.2	0.0020	0.0030	147.70	0.0016	mg/L	<KAL
Ints.	116	58	49.46			<KAL
Single values (Ints.): #1: 59 #2: 174 #3: 116						
Sb217.581						
Conc.1	0.0037	0.0023	62.66	0.0073	mg/L	
Conc.2	-0.0022	0.0023	107.07	0.0073	mg/L	
Ints.	-26	30	115.15			
Single values (Ints.): #1: -9 #2: -8 #3: -60						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0014	0.0015	110.54	0.0023	mg/L	
Conc.2	-0.0006	0.0015	265.31	0.0023	mg/L	
Ints.	299	25	8.48			
Single values (Ints.): #1: 319 #2: 271 #3: 308						
GH1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:39
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0003	0.0012	379.85	0.0017	mg/L	<KAL
Conc.2	-0.0007	0.0012	177.82	0.0017	mg/L	<KAL
Ints.	65	23	35.13			<KAL
Single values (Ints.): #1: 90 #2: 59 #3: 45						
Sb217.581						
Conc.1	0.0055	0.0019	33.66	0.0072	mg/L	
Conc.2	-0.0003	0.0019	551.74	0.0072	mg/L	
Ints.	-2	24	###.##			
Single values (Ints.): #1: 16 #2: -29 #3: 6						
Se196.028						
Conc.1	0.0025	0.0011	45.74	0.0023	mg/L	
Conc.2	0.0005	0.0011	215.28	0.0023	mg/L	
Ints.	318	19	5.95			
Single values (Ints.): #1: 335 #2: 297 #3: 321						
GH2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:41
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0021	0.0023	110.76	0.0016	mg/L	
Conc.2	0.0017	0.0023	134.02	0.0016	mg/L	
Ints.	110	43	39.48			
Single values (Ints.): #1: 129 #2: 60 #3: 140						
Sb217.581						
Conc.1	0.0012	0.0024	194.23	0.0074	mg/L	<KAL
Conc.2	-0.0046	0.0024	52.51	0.0074	mg/L	<KAL
Ints.	-57	31	54.30			<KAL
Single values (Ints.): #1: -92 #2: -44 #3: -35						
Se196.028						
Conc.1	0.0028	0.0018	66.36	0.0023	mg/L	
Conc.2	0.0008	0.0018	219.91	0.0023	mg/L	
Ints.	323	31	9.57			
Single values (Ints.): #1: 299 #2: 358 #3: 313						
GH3	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:42
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0002	0.0026	###.##	0.0017	mg/L	<KAL
Conc.2	-0.0001	0.0026	###.##	0.0017	mg/L	<KAL
Ints.	75	50	67.68			<KAL
Single values (Ints.): #1: 72 #2: 25 #3: 126						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	-0.0006	0.0024	432.95	0.0074	mg/L	<KAL
Conc.2	-0.0064	0.0024	38.22	0.0074	mg/L	<KAL
Ints.	-80	31	39.14			<KAL
Single values (Ints.): #1: -114 #2: -52 #3: -76						
Se196.028						
Conc.1	0.0013	0.0007	53.60	0.0023	mg/L	
Conc.2	-0.0007	0.0007	100.38	0.0023	mg/L	
Ints.	298	11	3.82			
Single values (Ints.): #1: 285 #2: 307 #3: 301						
GH4	*Method:*	As,Sb,Se(1)	Orig.	*Date:*		2025-12-04 15:44
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0029	0.0020	69.42	0.0016	mg/L	
Conc.2	0.0025	0.0020	79.14	0.0016	mg/L	
Ints.	126	38	30.48			
Single values (Ints.): #1: 169 #2: 112 #3: 96						
Sb217.581						
Conc.1	0.0009	0.0006	65.61	0.0074	mg/L	
Conc.2	-0.0049	0.0006	12.57	0.0074	mg/L	
Ints.	-61	8	12.97			
Single values (Ints.): #1: -57 #2: -55 #3: -70						
Se196.028						
Conc.1	0.0028	0.0022	79.40	0.0023	mg/L	
Conc.2	0.0009	0.0022	256.11	0.0023	mg/L	
Ints.	324	37	11.57			
Single values (Ints.): #1: 367 #2: 302 #3: 302						
GH5	*Method:*	As,Sb,Se(1)	Orig.	*Date:*		2025-12-04 15:46
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0020	0.0007	34.81	0.0016	mg/L	
Conc.2	0.0016	0.0007	42.44	0.0016	mg/L	
Ints.	108	13	12.13			
Single values (Ints.): #1: 98 #2: 123 #3: 104						
Sb217.581						
Conc.1	-0.0006	0.0030	485.38	0.0074	mg/L	<KAL
Conc.2	-0.0065	0.0030	46.84	0.0074	mg/L	<KAL
Ints.	-81	39	47.97			<KAL
Single values (Ints.): #1: -104 #2: -36 #3: -103						
Se196.028						
Conc.1	0.0015	0.0015	101.47	0.0023	mg/L	
Conc.2	-0.0004	0.0015	339.29	0.0023	mg/L	
Ints.	301	25	8.43			
Single values (Ints.): #1: 318 #2: 272 #3: 314						

Operator: *Lab.:*

Line	Mean	SD	RSD[%]	CI	Unit	Rem.
GH7	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:47
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0005	0.0003	69.90	0.0017	mg/L	
Conc.2	0.0001	0.0003	254.27	0.0017	mg/L	
Ints.	80	7	8.19			
Single values (Ints.): #1: 84 #2: 72 #3: 84						
Sb217.581						
Conc.1	0.0011	0.0026	225.95	0.0074	mg/L	
Conc.2	-0.0047	0.0026	54.67	0.0074	mg/L	
Ints.	-58	33	56.49			
Single values (Ints.): #1: -22 #2: -87 #3: -66						
Se196.028						
Conc.1	0.0019	0.0038	201.72	0.0023	mg/L	
Conc.2	-0.0000	0.0038	###.##	0.0023	mg/L	
Ints.	308	65	20.96			
Single values (Ints.): #1: 371 #2: 242 #3: 313						
CCV 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:49
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1003	0.0039	3.86	0.0014	mg/L	
Conc.2	0.0999	0.0039	3.87	0.0014	mg/L	
Ints.	1981	74	3.72			
Single values (Ints.): #1: 1898 #2: 2010 #3: 2036						
Sb217.581						
Conc.1	0.0943	0.0010	1.11	0.0064	mg/L	
Conc.2	0.0884	0.0010	1.19	0.0064	mg/L	
Ints.	1139	13	1.18			
Single values (Ints.): #1: 1151 #2: 1140 #3: 1125						
Se196.028						
Conc.1	0.0991	0.0019	1.92	0.0020	mg/L	
Conc.2	0.0971	0.0019	1.96	0.0020	mg/L	
Ints.	1941	32	1.65			
Single values (Ints.): #1: 1967 #2: 1905 #3: 1951						
GH8	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:51
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0020	0.0009	46.67	0.0016	mg/L	
Conc.2	0.0016	0.0009	56.90	0.0016	mg/L	
Ints.	108	18	16.27			
Single values (Ints.): #1: 112 #2: 89 #3: 124						
Sb217.581						
Conc.1	0.0079	0.0020	25.47	0.0072	mg/L	
Conc.2	0.0021	0.0020	96.45	0.0072	mg/L	
Ints.	29	26	89.96			
Single values (Ints.): #1: 1 #2: 52 #3: 34						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0064	0.0023	35.11	0.0023	mg/L	
Conc.2	0.0045	0.0023	50.32	0.0023	mg/L	
Ints.	384	38	9.84			
Single values (Ints.): #1: 365 #2: 428 #3: 359						
MCC	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:52
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0748	0.0006	0.75	0.0014	mg/L	
Conc.2	0.0745	0.0006	0.75	0.0014	mg/L	
Ints.	1496	11	0.71			
Single values (Ints.): #1: 1484 #2: 1504 #3: 1501						
Sb217.581						
Conc.1	0.0009	0.0018	207.44	0.0074	mg/L	<KAL
Conc.2	-0.0050	0.0018	36.86	0.0074	mg/L	<KAL
Ints.	-62	23	38.02			<KAL
Single values (Ints.): #1: -62 #2: -85 #3: -38						
Se196.028						
Conc.1	0.0020	0.0021	104.42	0.0023	mg/L	
Conc.2	0.0001	0.0021	###.##	0.0023	mg/L	
Ints.	310	35	11.26			
Single values (Ints.): #1: 348 #2: 279 #3: 302						
B1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:54
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0012	0.0008	65.14	0.0016	mg/L	
Conc.2	0.0009	0.0008	91.19	0.0016	mg/L	
Ints.	94	15	16.40			
Single values (Ints.): #1: 76 #2: 104 #3: 103						
Sb217.581						
Conc.1	0.0023	0.0018	76.47	0.0073	mg/L	
Conc.2	-0.0035	0.0018	50.28	0.0073	mg/L	
Ints.	-43	23	52.54			
Single values (Ints.): #1: -65 #2: -20 #3: -45						
Se196.028						
Conc.1	0.0017	0.0008	45.86	0.0023	mg/L	
Conc.2	-0.0003	0.0008	305.40	0.0023	mg/L	
Ints.	305	13	4.26			
Single values (Ints.): #1: 315 #2: 290 #3: 309						
B2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:56
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0023	0.0003	12.40	0.0016	mg/L	
Conc.2	0.0020	0.0003	14.67	0.0016	mg/L	
Ints.	115	5	4.76			
Single values (Ints.): #1: 117 #2: 108 #3: 119						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	-0.0009	0.0016	168.85	0.0075	mg/L	<KAL
Conc.2	-0.0068	0.0016	23.55	0.0075	mg/L	<KAL
Ints.	-85	21	24.08			<KAL
Single values (Ints.): #1: -98 #2: -96 #3: -61						
Se196.028						
Conc.1	0.0047	0.0019	40.11	0.0023	mg/L	
Conc.2	0.0027	0.0019	68.50	0.0023	mg/L	
Ints.	355	31	8.87			
Single values (Ints.): #1: 342 #2: 332 #3: 391						
CCV 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-04 15:57
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1006	0.0018	1.75	0.0014	mg/L	
Conc.2	0.1002	0.0018	1.76	0.0014	mg/L	
Ints.	1986	34	1.69			
Single values (Ints.): #1: 2011 #2: 1948 #3: 2000						
Sb217.581						
Conc.1	0.0934	0.0021	2.25	0.0063	mg/L	
Conc.2	0.0876	0.0021	2.40	0.0063	mg/L	
Ints.	1128	27	2.40			
Single values (Ints.): #1: 1106 #2: 1119 #3: 1158						
Se196.028						
Conc.1	0.0981	0.0012	1.23	0.0020	mg/L	
Conc.2	0.0962	0.0012	1.26	0.0020	mg/L	
Ints.	1925	20	1.05			
Single values (Ints.): #1: 1946 #2: 1906 #3: 1922						