Operator: *Lab.:*

Results

Results file: C:\Users\Public\Documents\Analytik Jena\ASpectPQ\ICP\Results\251217_VGA_B,HSDS-25-12-17 151911.tps

Instrument: PQ 9000 #13585100362BA0517 *Technique:* ICP-OES

Operator:

Comment:

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
method blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:20
As193.698						
Conc.1	0				mg/L	
Ints.	113	36	32.13			
Single values (Ints.): #1: 151 #2: 79 #3: 109						
Sb217.581						
Conc.1	0				mg/L	
Ints.	-33	62	186.72			
Single values (Ints.): #1: -35 #2: 30 #3: -95						
Se196.028						
Conc.1	0				mg/L	
Ints.	280	35	12.53			
Single values (Ints.): #1: 320 #2: 255 #3: 265						
ST1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:22
As193.698						
Conc.1	0.02				mg/L	
Ints.	437	18	4.13			
Single values (Ints.): #1: 442 #2: 417 #3: 453						
Sb217.581						
Conc.1	0.02				mg/L	
Ints.	129	12	9.62			
Single values (Ints.): #1: 144 #2: 122 #3: 123						
Se196.028						
Conc.1	0.02				mg/L	
Ints.	612	18	2.99			
Single values (Ints.): #1: 623 #2: 591 #3: 621						
ST2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:23
As193.698						
Conc.1	0.04				mg/L	
Ints.	840	20	2.42			
Single values (Ints.): #1: 864 #2: 826 #3: 831						
Sb217.581						
Conc.1	0.04				mg/L	
Ints.	406	9	2.29			
Single values (Ints.): #1: 412 #2: 395 #3: 411						
Se196.028						
Conc.1	0.04				mg/L	
Ints.	879	22	2.50			
Single values (Ints.): #1: 857 #2: 901 #3: 879						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
ST3	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:25
As193.698						
Conc.1	0.05				mg/L	
Ints.	1040	21	2.05			
Single values (Ints.): #1: 1044 #2: 1017 #3: 1059						
Sb217.581						
Conc.1	0.05				mg/L	
Ints.	580	23	4.04			
Single values (Ints.): #1: 606 #2: 561 #3: 573						
Se196.028						
Conc.1	0.05				mg/L	
Ints.	1101	3	0.29			
Single values (Ints.): #1: 1097 #2: 1101 #3: 1104						
ST4	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:27
As193.698						
Conc.1	0.1				mg/L	
Ints.	1937	11	0.57			
Single values (Ints.): #1: 1945 #2: 1941 #3: 1924						
Sb217.581						
Conc.1	0.1				mg/L	
Ints.	1226	17	1.40			
Single values (Ints.): #1: 1245 #2: 1212 #3: 1221						
Se196.028						
Conc.1	0.1				mg/L	
Ints.	1910	14	0.73			
Single values (Ints.): #1: 1919 #2: 1916 #3: 1893						
ST5	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:28
As193.698						
Conc.1	0.2				mg/L	
Ints.	3880	57	1.48			
Single values (Ints.): #1: 3846 #2: 3848 #3: 3946						
Sb217.581						
Conc.1	0.2				mg/L	
Ints.	2522	8	0.30			
Single values (Ints.): #1: 2527 #2: 2514 #3: 2527						
Se196.028						
Conc.1	0.2				mg/L	
Ints.	3559	25	0.71			
Single values (Ints.): #1: 3553 #2: 3538 #3: 3587						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*

Compute calib. *Method:* As193.698 *Date:* 2025-12-17 15:29

As193.698

y=a+bx a=83.234067 b=18898.040 Method SD: 0.0014911 mg/L

R²(adj.): 0.999440460 ; Slope: 18898.040 Ints./(mg/L) ; BEC: 0.3695446 mg/L



Compute calib. *Method:* Sb217.581 *Date:* 2025-12-17 15:29

Sb217.581

y=a+bx a=-84.73765 b=13020.690 Method SD: 0.0029718 mg/L

R²(adj.): 0.997779829 ; Slope: 13020.690 Ints./(mg/L) ; BEC: 1.1662814 mg/L



Compute calib. *Method:* Se196.028 *Date:* 2025-12-17 15:29

Se196.028

y=a+bx a=266.90273 b=16437.455 Method SD: 0.0015706 mg/L

R²(adj.): 0.999379276 ; Slope: 16437.455 Ints./(mg/L) ; BEC: 0.2082499 mg/L



Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
BLANK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:30
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0009	0.0023	270.42	0.0032	mg/L	
Conc.2	0.0099	0.0023	23.19	0.0032	mg/L	
Ints.	99	43	43.77			
Single values (Ints.): #1: 134 #2: 51 #3: 113						
Sb217.581						
Conc.1	0.0150	0.0043	28.54	0.0060	mg/L	
Conc.2	0.0241	0.0043	17.71	0.0060	mg/L	
Ints.	110	56	50.48			
Single values (Ints.): #1: 166 #2: 109 #3: 55						
Se196.028						
Conc.1	0.0031	0.0016	50.48	0.0033	mg/L	
Conc.2	0.0110	0.0016	14.41	0.0033	mg/L	
Ints.	318	26	8.16			
Single values (Ints.): #1: 346 #2: 316 #3: 294						
Reag. blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:32
As193.698						
Conc.1	0.0017	0.0016	95.09	0.0032	mg/L	
Ints.	116	31	26.59			
Single values (Ints.): #1: 150 #2: 91 #3: 105						
Sb217.581						
Conc.1	0.0055	0.0033	61.23	0.0062	mg/L	
Ints.	-14	43	315.77			
Single values (Ints.): #1: -19 #2: -54 #3: 32						
Se196.028						
Conc.1	0.0019	0.0021	111.16	0.0033	mg/L	<KAL
Ints.	298	35	11.74			<KAL
Single values (Ints.): #1: 262 #2: 332 #3: 301						
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:33
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0001	0.0020	###.##	0.0032	mg/L	
Conc.2	-0.0016	0.0020	126.94	0.0032	mg/L	
Ints.	85	39	45.37			
Single values (Ints.): #1: 115 #2: 99 #3: 42						
Sb217.581						
Conc.1	0.0023	0.0019	81.14	0.0063	mg/L	
Conc.2	-0.0032	0.0019	58.84	0.0063	mg/L	
Ints.	-55	24	44.09			
Single values (Ints.): #1: -29 #2: -78 #3: -58						
Se196.028						
Conc.1	-0.0004	0.0008	220.16	0.0034	mg/L	<KAL
Conc.2	-0.0023	0.0008	36.00	0.0034	mg/L	<KAL
Ints.	261	14	5.20			<KAL
Single values (Ints.): #1: 273 #2: 263 #3: 246						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:35
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0000	0.0012	###.##	0.0032	mg/L	<KAL
Conc.2	-0.0018	0.0012	66.04	0.0032	mg/L	<KAL
Ints.	82	22	26.65			<KAL
Single values (Ints.): #1: 87 #2: 101 #3: 58						
Sb217.581						
Conc.1	0.0049	0.0027	54.51	0.0062	mg/L	
Conc.2	-0.0006	0.0027	481.05	0.0062	mg/L	
Ints.	-21	35	165.58			
Single values (Ints.): #1: -60 #2: 6 #3: -9						
Se196.028						
Conc.1	0.0036	0.0003	8.43	0.0033	mg/L	
Conc.2	0.0017	0.0003	17.84	0.0033	mg/L	
Ints.	327	5	1.54			
Single values (Ints.): #1: 326 #2: 332 #3: 322						
FBK(B 12.16)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:37
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0004	0.0003	84.12	0.0032	mg/L	
Conc.2	-0.0014	0.0003	22.00	0.0032	mg/L	
Ints.	90	6	6.26			
Single values (Ints.): #1: 94 #2: 92 #3: 83						
Sb217.581						
Conc.1	0.0050	0.0009	18.58	0.0062	mg/L	
Conc.2	-0.0004	0.0009	209.60	0.0062	mg/L	
Ints.	-20	12	61.99			
Single values (Ints.): #1: -9 #2: -16 #3: -33						
Se196.028						
Conc.1	0.0028	0.0020	71.59	0.0033	mg/L	
Conc.2	0.0009	0.0020	224.18	0.0033	mg/L	
Ints.	313	33	10.58			
Single values (Ints.): #1: 344 #2: 278 #3: 317						
B2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:38
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0023	0.0019	82.73	0.0032	mg/L	
Conc.2	0.0006	0.0019	331.11	0.0032	mg/L	
Ints.	126	36	28.21			
Single values (Ints.): #1: 154 #2: 138 #3: 86						
Sb217.581						
Conc.1	0.0027	0.0021	78.26	0.0063	mg/L	
Conc.2	-0.0028	0.0021	76.31	0.0063	mg/L	
Ints.	-50	27	55.18			
Single values (Ints.): #1: -76 #2: -21 #3: -53						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0022	0.0020	89.71	0.0033	mg/L	
Conc.2	0.0003	0.0020	716.83	0.0033	mg/L	
Ints.	303	32	10.67			
Single values (Ints.): #1: 325 #2: 318 #3: 266						
JG IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:40
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0008	0.0005	65.35	0.0032	mg/L	
Conc.2	-0.0009	0.0005	54.26	0.0032	mg/L	
Ints.	98	10	9.78			
Single values (Ints.): #1: 87 #2: 103 #3: 104						
Sb217.581						
Conc.1	0.0029	0.0019	65.98	0.0063	mg/L	
Conc.2	-0.0026	0.0019	73.53	0.0063	mg/L	
Ints.	-47	25	52.15			
Single values (Ints.): #1: -45 #2: -24 #3: -73						
Se196.028						
Conc.1	0.0028	0.0010	35.20	0.0033	mg/L	
Conc.2	0.0008	0.0010	115.66	0.0033	mg/L	
Ints.	312	16	5.11			
Single values (Ints.): #1: 299 #2: 307 #3: 330						
JG1 OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:42
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0000	0.0004	###.##	0.0032	mg/L	
Conc.2	-0.0017	0.0004	22.97	0.0032	mg/L	
Ints.	84	7	8.67			
Single values (Ints.): #1: 92 #2: 82 #3: 78						
Sb217.581						
Conc.1	0.0020	0.0010	49.74	0.0063	mg/L	
Conc.2	-0.0034	0.0010	29.36	0.0063	mg/L	
Ints.	-58	13	22.44			
Single values (Ints.): #1: -43 #2: -64 #3: -68						
Se196.028						
Conc.1	0.0014	0.0021	147.32	0.0034	mg/L	<KAL
Conc.2	-0.0005	0.0021	454.40	0.0034	mg/L	<KAL
Ints.	291	35	12.06			<KAL
Single values (Ints.): #1: 265 #2: 331 #3: 277						
JG2 OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:43
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0008	0.0020	248.08	0.0032	mg/L	<KAL
Conc.2	-0.0009	0.0020	217.42	0.0032	mg/L	<KAL
Ints.	98	37	38.07			<KAL
Single values (Ints.): #1: 56 #2: 127 #3: 112						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0024	0.0024	100.35	0.0063	mg/L	
Conc.2	-0.0030	0.0024	80.91	0.0063	mg/L	
Ints.	-53	32	59.92			
Single values (Ints.): #1: -27 #2: -43 #3: -89						
Se196.028						
Conc.1	0.0015	0.0013	82.28	0.0034	mg/L	
Conc.2	-0.0004	0.0013	341.12	0.0034	mg/L	
Ints.	292	21	7.15			
Single values (Ints.): #1: 309 #2: 299 #3: 269						
JG3 OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:45
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0001	0.0021	###.##	0.0032	mg/L	<KAL
Conc.2	-0.0016	0.0021	124.97	0.0032	mg/L	<KAL
Ints.	84	39	46.01			<KAL
Single values (Ints.): #1: 45 #2: 86 #3: 122						
Sb217.581						
Conc.1	0.0015	0.0011	71.40	0.0063	mg/L	
Conc.2	-0.0039	0.0011	28.32	0.0063	mg/L	
Ints.	-65	14	22.29			
Single values (Ints.): #1: -81 #2: -54 #3: -59						
Se196.028						
Conc.1	0.0026	0.0009	34.74	0.0033	mg/L	
Conc.2	0.0006	0.0009	139.52	0.0033	mg/L	
Ints.	309	15	4.72			
Single values (Ints.): #1: 321 #2: 312 #3: 293						
IN IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:47
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0003	0.0007	257.30	0.0032	mg/L	<KAL
Conc.2	-0.0020	0.0007	36.20	0.0032	mg/L	<KAL
Ints.	78	14	17.46			<KAL
Single values (Ints.): #1: 73 #2: 93 #3: 68						
Sb217.581						
Conc.1	0.0038	0.0009	24.68	0.0063	mg/L	
Conc.2	-0.0016	0.0009	57.56	0.0063	mg/L	
Ints.	-35	12	34.97			
Single values (Ints.): #1: -49 #2: -31 #3: -25						
Se196.028						
Conc.1	0.0017	0.0018	108.54	0.0033	mg/L	
Conc.2	-0.0003	0.0018	688.86	0.0033	mg/L	
Ints.	294	30	10.05			
Single values (Ints.): #1: 273 #2: 282 #3: 328						

Operator: *Lab.:*

Line	Mean	SD	RSD[%]	CI	Unit	Rem.
IN	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:48
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0007	0.0003	43.13	0.0032	mg/L	<KAL
Conc.2	-0.0024	0.0003	11.97	0.0032	mg/L	<KAL
Ints.	71	5	7.56			<KAL
Single values (Ints.): #1: 77 #2: 69 #3: 67						
Sb217.581						
Conc.1	0.0042	0.0031	75.13	0.0063	mg/L	
Conc.2	-0.0013	0.0031	242.43	0.0063	mg/L	
Ints.	-31	41	133.23			
Single values (Ints.): #1: -77 #2: -14 #3: -1						
Se196.028						
Conc.1	0.0032	0.0003	10.29	0.0033	mg/L	
Conc.2	0.0013	0.0003	25.89	0.0033	mg/L	
Ints.	319	5	1.69			
Single values (Ints.): #1: 313 #2: 324 #3: 321						
YR	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:50
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0008	0.0009	114.81	0.0032	mg/L	<KAL
Conc.2	-0.0025	0.0009	35.75	0.0032	mg/L	<KAL
Ints.	69	17	24.43			<KAL
Single values (Ints.): #1: 54 #2: 66 #3: 87						
Sb217.581						
Conc.1	0.0021	0.0036	169.63	0.0063	mg/L	<KAL
Conc.2	-0.0033	0.0036	109.72	0.0063	mg/L	<KAL
Ints.	-57	47	83.16			<KAL
Single values (Ints.): #1: -88 #2: -2 #3: -80						
Se196.028						
Conc.1	0.0023	0.0012	53.29	0.0033	mg/L	
Conc.2	0.0004	0.0012	330.07	0.0033	mg/L	
Ints.	304	20	6.58			
Single values (Ints.): #1: 287 #2: 326 #3: 301						
OUT	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:52
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0008	0.0005	56.45	0.0032	mg/L	<KAL
Conc.2	-0.0025	0.0005	18.24	0.0032	mg/L	<KAL
Ints.	68	9	12.84			<KAL
Single values (Ints.): #1: 58 #2: 69 #3: 76						
Sb217.581						
Conc.1	0.0034	0.0045	132.84	0.0063	mg/L	
Conc.2	-0.0021	0.0045	212.58	0.0063	mg/L	
Ints.	-41	58	141.32			
Single values (Ints.): #1: 25 #2: -82 #3: -67						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0006	0.0017	282.38	0.0034	mg/L	
Conc.2	-0.0013	0.0017	129.49	0.0034	mg/L	
Ints.	277	28	10.11			
Single values (Ints.): #1: 309 #2: 263 #3: 258						
CCV 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2025-12-17 15:53
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1003	0.0020	2.02	0.0028	mg/L	
Conc.2	0.0986	0.0020	2.06	0.0028	mg/L	
Ints.	1979	38	1.94			
Single values (Ints.): #1: 1967 #2: 1948 #3: 2022						
Sb217.581						
Conc.1	0.0945	0.0033	3.45	0.0055	mg/L	
Conc.2	0.0891	0.0033	3.66	0.0055	mg/L	
Ints.	1146	42	3.70			
Single values (Ints.): #1: 1131 #2: 1194 #3: 1113						
Se196.028						
Conc.1	0.0997	0.0010	1.01	0.0029	mg/L	
Conc.2	0.0978	0.0010	1.03	0.0029	mg/L	
Ints.	1905	17	0.87			
Single values (Ints.): #1: 1911 #2: 1919 #3: 1887						