Operator: *Lab.:*

Results

Results file: C:\Users\Public\Documents\Analytik Jena\ASpectPQ\ICP\Results\260210_KDJR-26-02-10 93914.tps

Instrument: PQ 9000 #13585100362BA0517 *Technique:* ICP-OES

Operator:

Comment:

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
method blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:40
As193.698						
Conc.1	0				mg/L	
Ints.	21	19	90.71			
Single values (Ints.): #1: 30 #2: 34 #3: -1						
Sb217.581						
Conc.1	0				mg/L	
Ints.	26	20	78.97			
Single values (Ints.): #1: 25 #2: 47 #3: 6						
Se196.028						
Conc.1	0				mg/L	
Ints.	183	43	23.22			
Single values (Ints.): #1: 135 #2: 204 #3: 212						
ST1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:41
As193.698						
Conc.1	0.02				mg/L	
Ints.	475	16	3.45			
Single values (Ints.): #1: 456 #2: 481 #3: 486						
Sb217.581						
Conc.1	0.02				mg/L	
Ints.	223	27	12.01			
Single values (Ints.): #1: 238 #2: 192 #3: 239						
Se196.028						
Conc.1	0.02				mg/L	
Ints.	549	12	2.16			
Single values (Ints.): #1: 563 #2: 544 #3: 541						
ST2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:43
As193.698						
Conc.1	0.04				mg/L	
Ints.	879	18	2.08			
Single values (Ints.): #1: 896 #2: 860 #3: 881						
Sb217.581						
Conc.1	0.04				mg/L	
Ints.	504	17	3.39			
Single values (Ints.): #1: 484 #2: 512 #3: 515						
Se196.028						
Conc.1	0.04				mg/L	
Ints.	892	9	1.03			
Single values (Ints.): #1: 891 #2: 883 #3: 901						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
ST3	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:45
As193.698						
Conc.1	0.05				mg/L	
Ints.	1147	57	4.94			
Single values (Ints.): #1: 1099 #2: 1133 #3: 1210						
Sb217.581						
Conc.1	0.05				mg/L	
Ints.	677	16	2.41			
Single values (Ints.): #1: 664 #2: 671 #3: 695						
Se196.028						
Conc.1	0.05				mg/L	
Ints.	1157	31	2.72			
Single values (Ints.): #1: 1150 #2: 1131 #3: 1192						
ST4	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:47
As193.698						
Conc.1	0.1				mg/L	
Ints.	2351	25	1.05			
Single values (Ints.): #1: 2365 #2: 2323 #3: 2366						
Sb217.581						
Conc.1	0.1				mg/L	
Ints.	1451	20	1.40			
Single values (Ints.): #1: 1428 #2: 1468 #3: 1457						
Se196.028						
Conc.1	0.1				mg/L	
Ints.	2182	45	2.07			
Single values (Ints.): #1: 2226 #2: 2136 #3: 2186						
ST5	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:48
As193.698						
Conc.1	0.2				mg/L	
Ints.	4609	27	0.58			
Single values (Ints.): #1: 4604 #2: 4638 #3: 4585						
Sb217.581						
Conc.1	0.2				mg/L	
Ints.	3002	84	2.81			
Single values (Ints.): #1: 2910 #2: 3075 #3: 3020						
Se196.028						
Conc.1	0.2				mg/L	
Ints.	4136	63	1.53			
Single values (Ints.): #1: 4133 #2: 4074 #3: 4200						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*

Compute calib. *Method:* As193.698 *Date:* 2026-02-10 9:49

As193.698

y=a+bx a=3.7761214 b=23071.564 Method SD: 0.0014490 mg/L

R²(adj.): 0.999471603 ; Slope: 23071.564 Ints./(mg/L) ; BEC: 0.2989297 mg/L



Compute calib. *Method:* Sb217.581 *Date:* 2026-02-10 9:49

Sb217.581

y=a+bx a=-56.78279 b=15177.216 Method SD: 0.0034329 mg/L

R²(adj.): 0.997038611 ; Slope: 15177.216 Ints./(mg/L) ; BEC: 0.7908925 mg/L



Compute calib. *Method:* Se196.028 *Date:* 2026-02-10 9:49

Se196.028

y=a+bx a=154.01191 b=19940.316 Method SD: 0.0019010 mg/L

R²(adj.): 0.999090750 ; Slope: 19940.316 Ints./(mg/L) ; BEC: 0.2331631 mg/L



Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
BLANK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:50
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0024	0.0014	58.99	0.0031	mg/L	
Conc.2	0.0024	0.0014	58.99	0.0031	mg/L	
Ints.	60	33	55.26			
Single values (Ints.): #1: 86 #2: 23 #3: 70						
Blank unit mismatch! (#1392)						
Sb217.581						
Conc.1	0.0103	0.0008	7.56	0.0071	mg/L	
Conc.2	0.0103	0.0008	7.56	0.0071	mg/L	
Ints.	99	12	11.89			
Single values (Ints.): #1: 103 #2: 108 #3: 86						
Blank unit mismatch! (#1392)						
Se196.028						
Conc.1	0.0023	0.0017	73.87	0.0040	mg/L	
Conc.2	0.0023	0.0017	73.87	0.0040	mg/L	
Ints.	200	34	17.05			
Single values (Ints.): #1: 210 #2: 162 #3: 228						
Blank unit mismatch! (#1392)						
Reag. blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:52
As193.698						
Conc.1	0.0019	0.0008	39.39	0.0031	mg/L	
Ints.	49	18	36.34			
Single values (Ints.): #1: 40 #2: 37 #3: 69						
Sb217.581						
Conc.1	0.0021	0.0022	104.32	0.0073	mg/L	<KAL
Ints.	-25	33	128.93			<KAL
Single values (Ints.): #1: -61 #2: -18 #3: 3						
Se196.028						
Conc.1	0.0017	0.0011	65.52	0.0041	mg/L	
Ints.	188	22	11.77			
Single values (Ints.): #1: 213 #2: 176 #3: 174						
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:54
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0014	0.0022	160.23	0.0031	mg/L	
Conc.2	-0.0006	0.0022	378.99	0.0031	mg/L	
Ints.	35	51	143.11			
Single values (Ints.): #1: 50 #2: -21 #3: 77						
Sb217.581						
Conc.1	0.0017	0.0006	36.38	0.0073	mg/L	
Conc.2	-0.0004	0.0006	171.67	0.0073	mg/L	
Ints.	-31	9	30.51			
Single values (Ints.): #1: -27 #2: -42 #3: -24						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0011	0.0012	106.27	0.0041	mg/L	
Conc.2	-0.0006	0.0012	214.48	0.0041	mg/L	
Ints.	177	24	13.58			
Single values (Ints.): #1: 204 #2: 162 #3: 163						
FBK(KD)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:55
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0039	0.0016	40.88	0.0031	mg/L	
Conc.2	0.0020	0.0016	81.22	0.0031	mg/L	
Ints.	94	37	39.24			
Single values (Ints.): #1: 53 #2: 104 #3: 125						
Sb217.581						
Conc.1	0.0008	0.0007	93.61	0.0074	mg/L	
Conc.2	-0.0013	0.0007	55.06	0.0074	mg/L	
Ints.	-45	11	24.10			
Single values (Ints.): #1: -36 #2: -57 #3: -43						
Se196.028						
Conc.1	0.0031	0.0004	11.93	0.0040	mg/L	
Conc.2	0.0014	0.0004	26.24	0.0040	mg/L	
Ints.	216	7	3.42			
Single values (Ints.): #1: 215 #2: 209 #3: 224						
FBK(JR)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:57
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0025	0.0007	25.97	0.0031	mg/L	
Conc.2	0.0006	0.0007	110.21	0.0031	mg/L	
Ints.	63	15	24.40			
Single values (Ints.): #1: 65 #2: 77 #3: 46						
Sb217.581						
Conc.1	0.0023	0.0016	72.12	0.0073	mg/L	
Conc.2	0.0002	0.0016	870.06	0.0073	mg/L	
Ints.	-23	25	109.42			
Single values (Ints.): #1: -28 #2: 4 #3: -44						
Se196.028						
Conc.1	0.0020	0.0004	21.95	0.0040	mg/L	
Conc.2	0.0003	0.0004	133.32	0.0040	mg/L	
Ints.	194	9	4.56			
Single values (Ints.): #1: 184 #2: 198 #3: 201						
KD1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 9:59
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0037	0.0005	14.15	0.0031	mg/L	
Conc.2	0.0018	0.0005	29.56	0.0031	mg/L	
Ints.	90	12	13.56			
Single values (Ints.): #1: 103 #2: 80 #3: 87						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0030	0.0019	61.53	0.0073	mg/L	
Conc.2	0.0010	0.0019	191.55	0.0073	mg/L	
Ints.	-11	28	269.89			
Single values (Ints.): #1: -35 #2: 21 #3: -18						
Se196.028						
Conc.1	0.0013	0.0011	80.91	0.0041	mg/L	
Conc.2	-0.0004	0.0011	289.47	0.0041	mg/L	
Ints.	180	21	11.83			
Single values (Ints.): #1: 175 #2: 162 #3: 204						
KD2	*Method:*	As,Sb,Se(1)	Orig.	*Date:*		2026-02-10 10:00
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0017	0.0009	57.04	0.0031	mg/L	
Conc.2	-0.0003	0.0009	318.96	0.0031	mg/L	
Ints.	42	22	51.89			
Single values (Ints.): #1: 44 #2: 19 #3: 62						
Sb217.581						
Conc.1	0.0007	0.0019	273.69	0.0074	mg/L	
Conc.2	-0.0014	0.0019	143.10	0.0074	mg/L	
Ints.	-46	29	64.11			
Single values (Ints.): #1: -13 #2: -54 #3: -70						
Se196.028						
Conc.1	0.0009	0.0005	57.06	0.0041	mg/L	
Conc.2	-0.0008	0.0005	71.11	0.0041	mg/L	
Ints.	173	11	6.18			
Single values (Ints.): #1: 182 #2: 161 #3: 175						
KD3	*Method:*	As,Sb,Se(1)	Orig.	*Date:*		2026-02-10 10:02
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0027	0.0008	29.45	0.0031	mg/L	
Conc.2	0.0007	0.0008	110.30	0.0031	mg/L	
Ints.	65	18	27.74			
Single values (Ints.): #1: 83 #2: 47 #3: 65						
Sb217.581						
Conc.1	0.0018	0.0028	158.45	0.0073	mg/L	<KAL
Conc.2	-0.0003	0.0028	962.69	0.0073	mg/L	<KAL
Ints.	-30	43	143.15			<KAL
Single values (Ints.): #1: -79 #2: -12 #3: 1						
Se196.028						
Conc.1	-0.0001	0.0004	438.75	0.0041	mg/L	<KAL
Conc.2	-0.0018	0.0004	22.38	0.0041	mg/L	<KAL
Ints.	152	8	5.23			<KAL
Single values (Ints.): #1: 160 #2: 153 #3: 144						

Operator: *Lab.:*

Line	Mean	SD	RSD[%]	CI	Unit	Rem.
KD4	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 10:04
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0022	0.0008	34.99	0.0031	mg/L	
Conc.2	0.0002	0.0008	315.71	0.0031	mg/L	
Ints.	54	18	32.55			
Single values (Ints.): #1: 35 #2: 59 #3: 69						
Sb217.581						
Conc.1	0.0024	0.0009	36.23	0.0073	mg/L	
Conc.2	0.0004	0.0009	246.74	0.0073	mg/L	
Ints.	-20	13	66.65			
Single values (Ints.): #1: -10 #2: -15 #3: -35						
Se196.028						
Conc.1	0.0019	0.0014	75.93	0.0041	mg/L	
Conc.2	0.0002	0.0014	680.62	0.0041	mg/L	
Ints.	192	29	15.02			
Single values (Ints.): #1: 176 #2: 175 #3: 225						
JR1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 10:12
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0014	0.0004	32.23	0.0031	mg/L	
Conc.2	-0.0006	0.0004	79.26	0.0031	mg/L	
Ints.	36	10	28.82			
Single values (Ints.): #1: 31 #2: 48 #3: 29						
Sb217.581						
Conc.1	0.0027	0.0004	16.09	0.0073	mg/L	
Conc.2	0.0007	0.0004	66.96	0.0073	mg/L	
Ints.	-15	7	43.00			
Single values (Ints.): #1: -17 #2: -8 #3: -21						
Se196.028						
Conc.1	0.0006	0.0015	231.94	0.0041	mg/L	<KAL
Conc.2	-0.0011	0.0015	139.61	0.0041	mg/L	<KAL
Ints.	167	29	17.64			<KAL
Single values (Ints.): #1: 149 #2: 201 #3: 150						
JR2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 10:24
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0057	0.0013	22.58	0.0032	mg/L	<KAL
Conc.2	-0.0077	0.0013	16.87	0.0032	mg/L	<KAL
Ints.	-129	30	23.24			<KAL
Single values (Ints.): #1: -106 #2: -117 #3: -163						
Sb217.581						
Conc.1	-0.0014	0.0015	107.62	0.0074	mg/L	<KAL
Conc.2	-0.0035	0.0015	43.72	0.0074	mg/L	<KAL
Ints.	-78	23	29.53			<KAL
Single values (Ints.): #1: -56 #2: -102 #3: -76						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0698	0.0016	2.34	0.0034	mg/L	
Conc.2	0.0682	0.0016	2.40	0.0034	mg/L	
Ints.	1547	33	2.11			
Single values (Ints.): #1: 1520 #2: 1537 #3: 1583						
JR3	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 10:26
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0029	0.0022	76.27	0.0031	mg/L	
Conc.2	0.0009	0.0022	237.74	0.0031	mg/L	
Ints.	70	50	72.15			
Single values (Ints.): #1: 73 #2: 119 #3: 18						
Sb217.581						
Conc.1	0.0018	0.0021	115.04	0.0073	mg/L	
Conc.2	-0.0002	0.0021	949.28	0.0073	mg/L	
Ints.	-29	32	111.87			
Single values (Ints.): #1: -49 #2: 8 #3: -46						
Se196.028						
Conc.1	0.0012	0.0006	49.28	0.0041	mg/L	
Conc.2	-0.0005	0.0006	115.93	0.0041	mg/L	
Ints.	178	12	6.56			
Single values (Ints.): #1: 181 #2: 165 #3: 187						
JR4	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 10:27
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0034	0.0006	17.77	0.0031	mg/L	
Conc.2	0.0015	0.0006	40.91	0.0031	mg/L	
Ints.	83	14	16.97			
Single values (Ints.): #1: 88 #2: 94 #3: 67						
Sb217.581						
Conc.1	0.0012	0.0009	75.33	0.0073	mg/L	
Conc.2	-0.0008	0.0009	109.72	0.0073	mg/L	
Ints.	-38	14	36.73			
Single values (Ints.): #1: -37 #2: -25 #3: -53						
Se196.028						
Conc.1	0.0026	0.0008	31.91	0.0040	mg/L	
Conc.2	0.0009	0.0008	92.28	0.0040	mg/L	
Ints.	206	16	8.00			
Single values (Ints.): #1: 219 #2: 187 #3: 211						
JR5	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 10:29
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0015	0.0009	59.30	0.0031	mg/L	
Conc.2	-0.0004	0.0009	223.59	0.0031	mg/L	
Ints.	39	21	53.60			
Single values (Ints.): #1: 15 #2: 53 #3: 50						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0019	0.0021	111.18	0.0073	mg/L	<KAL
Conc.2	-0.0002	0.0021	###.##	0.0073	mg/L	<KAL
Ints.	-28	32	116.73			<KAL
Single values (Ints.): #1: -64 #2: -17 #3: -2						
Se196.028						
Conc.1	0.0004	0.0010	285.14	0.0041	mg/L	
Conc.2	-0.0013	0.0010	76.40	0.0041	mg/L	
Ints.	161	20	12.61			
Single values (Ints.): #1: 156 #2: 184 #3: 144						
JR6	*Method:*	As,Sb,Se(1)	Orig.	*Date:*	2026-02-10 10:31	
Pre-DF: 1	*Wt.[g]:*	*Vol.[mL]:*	100		*Blank corr.:*	on
As193.698						
Conc.1	0.0026	0.0005	19.58	0.0031	mg/L	
Conc.2	0.0007	0.0005	76.47	0.0031	mg/L	
Ints.	64	12	18.43			
Single values (Ints.): #1: 53 #2: 76 #3: 64						
Sb217.581						
Conc.1	0.0000	0.0021	###.##	0.0074	mg/L	<KAL
Conc.2	-0.0020	0.0021	101.03	0.0074	mg/L	<KAL
Ints.	-56	31	55.37			<KAL
Single values (Ints.): #1: -28 #2: -90 #3: -50						
Se196.028						
Conc.1	0.0028	0.0008	28.59	0.0040	mg/L	
Conc.2	0.0011	0.0008	71.27	0.0040	mg/L	
Ints.	210	16	7.66			
Single values (Ints.): #1: 220 #2: 220 #3: 192						
JR7	*Method:*	As,Sb,Se(1)	Orig.	*Date:*	2026-02-10 10:32	
Pre-DF: 1	*Wt.[g]:*	*Vol.[mL]:*	100		*Blank corr.:*	on
As193.698						
Conc.1	0.0029	0.0015	51.29	0.0031	mg/L	
Conc.2	0.0010	0.0015	152.42	0.0031	mg/L	
Ints.	71	35	48.58			
Single values (Ints.): #1: 75 #2: 35 #3: 105						
Sb217.581						
Conc.1	0.0006	0.0042	646.92	0.0074	mg/L	
Conc.2	-0.0014	0.0042	294.12	0.0074	mg/L	
Ints.	-47	63	135.10			
Single values (Ints.): #1: 16 #2: -47 #3: -111						
Se196.028						
Conc.1	0.0013	0.0006	49.65	0.0041	mg/L	
Conc.2	-0.0004	0.0006	149.12	0.0041	mg/L	
Ints.	179	13	7.01			
Single values (Ints.): #1: 193 #2: 177 #3: 168						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
JR8	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 10:34
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0029	0.0013	42.63	0.0031	mg/L	
Conc.2	0.0010	0.0013	126.76	0.0031	mg/L	
Ints.	71	29	40.38			
Single values (Ints.): #1: 104 #2: 59 #3: 51						
Sb217.581						
Conc.1	0.0017	0.0014	83.95	0.0073	mg/L	
Conc.2	-0.0004	0.0014	335.49	0.0073	mg/L	
Ints.	-32	21	66.53			
Single values (Ints.): #1: -51 #2: -34 #3: -9						
Se196.028						
Conc.1	0.0026	0.0012	46.21	0.0040	mg/L	
Conc.2	0.0009	0.0012	133.21	0.0040	mg/L	
Ints.	206	24	11.60			
Single values (Ints.): #1: 190 #2: 194 #3: 233						
CCV ST4 0.1mg/L	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-02-10 10:36
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0969	0.0021	2.15	0.0027	mg/L	
Conc.2	0.0949	0.0021	2.20	0.0027	mg/L	
Ints.	2239	48	2.15			
Single values (Ints.): #1: 2233 #2: 2290 #3: 2194						
Sb217.581						
Conc.1	0.0920	0.0014	1.49	0.0063	mg/L	
Conc.2	0.0900	0.0014	1.53	0.0063	mg/L	
Ints.	1340	21	1.55			
Single values (Ints.): #1: 1317 #2: 1357 #3: 1345						
Se196.028						
Conc.1	0.0937	0.0006	0.59	0.0035	mg/L	
Conc.2	0.0920	0.0006	0.60	0.0035	mg/L	
Ints.	2023	11	0.55			
Single values (Ints.): #1: 2034 #2: 2023 #3: 2012						