Operator: *Lab.:*

Results

Results file: C:\Users\Public\Documents\Analytik Jena\ASpectPQ\ICP\Results\260320 HS BI-26-03-20 115208.tps

Instrument: PQ 9000 #13585100362BA0517 *Technique:* ICP-OES

Operator:

Comment:

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
method blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 11:53
As193.698						
Conc.1	0				mg/L	
Ints.	100	19	18.79			
Single values (Ints.): #1: 99 #2: 82 #3: 120						
Sb217.581						
Conc.1	0				mg/L	
Ints.	-28	18	64.93			
Single values (Ints.): #1: -46 #2: -9 #3: -30						
Se196.028						
Conc.1	0				mg/L	
Ints.	202	14	7.06			
Single values (Ints.): #1: 186 #2: 214 #3: 204						
ST1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 11:54
As193.698						
Conc.1	0.02				mg/L	
Ints.	399	23	5.80			
Single values (Ints.): #1: 426 #2: 388 #3: 384						
Sb217.581						
Conc.1	0.02				mg/L	
Ints.	158	24	15.28			
Single values (Ints.): #1: 131 #2: 177 #3: 167						
Se196.028						
Conc.1	0.02				mg/L	
Ints.	491	23	4.78			
Single values (Ints.): #1: 464 #2: 501 #3: 508						
ST2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 11:56
As193.698						
Conc.1	0.04				mg/L	
Ints.	723	22	2.99			
Single values (Ints.): #1: 724 #2: 700 #3: 743						
Sb217.581						
Conc.1	0.04				mg/L	
Ints.	383	31	8.09			
Single values (Ints.): #1: 349 #2: 410 #3: 389						
Se196.028						
Conc.1	0.04				mg/L	
Ints.	783	7	0.92			
Single values (Ints.): #1: 788 #2: 775 #3: 787						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
ST3	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 11:58
As193.698						
Conc.1	0.05				mg/L	
Ints.	879	34	3.84			
Single values (Ints.): #1: 843 #2: 886 #3: 910						
Sb217.581						
Conc.1	0.05				mg/L	
Ints.	495	31	6.26			
Single values (Ints.): #1: 502 #2: 461 #3: 522						
Se196.028						
Conc.1	0.05				mg/L	
Ints.	901	34	3.81			
Single values (Ints.): #1: 863 #2: 929 #3: 912						
ST4	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 11:59
As193.698						
Conc.1	0.1				mg/L	
Ints.	1705	48	2.80			
Single values (Ints.): #1: 1707 #2: 1657 #3: 1752						
Sb217.581						
Conc.1	0.1				mg/L	
Ints.	1062	27	2.54			
Single values (Ints.): #1: 1088 #2: 1062 #3: 1034						
Se196.028						
Conc.1	0.1				mg/L	
Ints.	1556	16	1.05			
Single values (Ints.): #1: 1552 #2: 1542 #3: 1574						
ST5	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:01
As193.698						
Conc.1	0.2				mg/L	
Ints.	3317	61	1.84			
Single values (Ints.): #1: 3381 #2: 3310 #3: 3260						
Sb217.581						
Conc.1	0.2				mg/L	
Ints.	2196	47	2.16			
Single values (Ints.): #1: 2165 #2: 2251 #3: 2174						
Se196.028						
Conc.1	0.2				mg/L	
Ints.	2953	36	1.22			
Single values (Ints.): #1: 2985 #2: 2960 #3: 2914						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*

Compute calib. *Method:* As193.698 *Date:* 2026-03-20 12:02

As193.698

y=a+bx a=82.970087 b=16161.200 Method SD: 0.0007376 mg/L

R²(adj.): 0.999863069 ; Slope: 16161.200 Ints./(mg/L) ; BEC: 0.3031980 mg/L



Compute calib. *Method:* Sb217.581 *Date:* 2026-03-20 12:02

Sb217.581

y=a+bx a=-55.79847 b=11221.529 Method SD: 0.0015176 mg/L

R²(adj.): 0.999420453 ; Slope: 11221.529 Ints./(mg/L) ; BEC: 0.9137123 mg/L



Compute calib. *Method:* Se196.028 *Date:* 2026-03-20 12:02

Se196.028

y=a+bx a=213.97353 b=13662.687 Method SD: 0.0013306 mg/L

R²(adj.): 0.999554428 ; Slope: 13662.687 Ints./(mg/L) ; BEC: 0.2113210 mg/L



Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
BLANK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:03
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0011	0.0009	84.06	0.0016	mg/L	<KAL
Conc.2	-0.0044	0.0009	21.55	0.0016	mg/L	<KAL
Ints.	65	15	23.72			<KAL
Single values (Ints.): #1: 49 #2: 80 #3: 65						
Sb217.581						
Conc.1	0.0103	0.0003	3.34	0.0031	mg/L	
Conc.2	0.0053	0.0003	6.50	0.0031	mg/L	
Ints.	60	4	6.44			
Single values (Ints.): #1: 64 #2: 61 #3: 56						
Se196.028						
Conc.1	-0.0005	0.0012	238.00	0.0029	mg/L	<KAL
Conc.2	-0.0053	0.0012	23.19	0.0029	mg/L	<KAL
Ints.	207	17	8.07			<KAL
Single values (Ints.): #1: 198 #2: 226 #3: 197						
Reag. blank	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:04
As193.698						
Conc.1	-0.0004	0.0010	257.04	0.0016	mg/L	<KAL
Ints.	77	16	20.49			<KAL
Single values (Ints.): #1: 78 #2: 61 #3: 92						
Sb217.581						
Conc.1	0.0047	0.0008	17.02	0.0032	mg/L	
Ints.	-3	9	283.63			
Single values (Ints.): #1: -6 #2: -11 #3: 7						
Se196.028						
Conc.1	-0.0026	0.0018	70.29	0.0029	mg/L	<KAL
Ints.	179	25	13.82			<KAL
Single values (Ints.): #1: 207 #2: 159 #3: 170						
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:06
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0003	0.0009	270.00	0.0016	mg/L	<KAL
Conc.2	0.0000	0.0009	###.##	0.0016	mg/L	<KAL
Ints.	77	15	19.16			<KAL
Single values (Ints.): #1: 93 #2: 75 #3: 64						
Sb217.581						
Conc.1	0.0034	0.0031	91.89	0.0032	mg/L	
Conc.2	-0.0013	0.0031	236.26	0.0032	mg/L	
Ints.	-18	35	194.56			
Single values (Ints.): #1: -1 #2: -58 #3: 5						
Se196.028						
Conc.1	-0.0016	0.0028	172.29	0.0029	mg/L	<KAL
Conc.2	0.0009	0.0028	305.96	0.0029	mg/L	<KAL
Ints.	191	39	20.24			<KAL
Single values (Ints.): #1: 191 #2: 230 #3: 153						

Operator: *Lab.:*

Line	Mean	SD	RSD[%]	CI	Unit	Rem.
MBK	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:08
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0013	0.0012	90.97	0.0016	mg/L	<KAL
Conc.2	-0.0010	0.0012	126.53	0.0016	mg/L	<KAL
Ints.	61	20	32.42			<KAL
Single values (Ints.): #1: 70 #2: 75 #3: 38						
Sb217.581						
Conc.1	0.0032	0.0004	12.73	0.0032	mg/L	
Conc.2	-0.0015	0.0004	26.33	0.0032	mg/L	
Ints.	-20	5	22.23			
Single values (Ints.): #1: -23 #2: -15 #3: -23						
Se196.028						
Conc.1	-0.0003	0.0018	658.04	0.0029	mg/L	<KAL
Conc.2	0.0023	0.0018	79.27	0.0029	mg/L	<KAL
Ints.	210	25	11.83			<KAL
Single values (Ints.): #1: 219 #2: 182 #3: 230						
FBK(HS)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:09
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0024	0.0009	36.52	0.0016	mg/L	<KAL
Conc.2	-0.0020	0.0009	43.49	0.0016	mg/L	<KAL
Ints.	45	14	31.18			<KAL
Single values (Ints.): #1: 31 #2: 59 #3: 43						
Sb217.581						
Conc.1	0.0032	0.0012	36.69	0.0032	mg/L	
Conc.2	-0.0014	0.0012	82.43	0.0032	mg/L	
Ints.	-19	13	68.99			
Single values (Ints.): #1: -28 #2: -4 #3: -26						
Se196.028						
Conc.1	-0.0021	0.0011	50.23	0.0029	mg/L	<KAL
Conc.2	0.0004	0.0011	246.24	0.0029	mg/L	<KAL
Ints.	185	15	7.94			<KAL
Single values (Ints.): #1: 169 #2: 199 #3: 186						
FBK(BI)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:11
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0006	0.0006	105.00	0.0016	mg/L	<KAL
Conc.2	-0.0002	0.0006	314.85	0.0016	mg/L	<KAL
Ints.	74	10	13.08			<KAL
Single values (Ints.): #1: 70 #2: 85 #3: 67						
Sb217.581						
Conc.1	0.0047	0.0036	76.26	0.0032	mg/L	
Conc.2	0.0000	0.0036	###.##	0.0032	mg/L	
Ints.	-3	40	###.##			
Single values (Ints.): #1: -18 #2: -34 #3: 43						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Se196.028						
Conc.1	0.0003	0.0015	464.87	0.0029	mg/L	<KAL
Conc.2	0.0029	0.0015	51.84	0.0029	mg/L	<KAL
Ints.	218	21	9.39			<KAL
Single values (Ints.): #1: 208 #2: 242 #3: 206						
HS	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:13
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0028	0.0013	47.17	0.0016	mg/L	<KAL
Conc.2	-0.0024	0.0013	54.54	0.0016	mg/L	<KAL
Ints.	38	21	56.79			<KAL
Single values (Ints.): #1: 55 #2: 45 #3: 14						
Sb217.581						
Conc.1	0.0028	0.0011	39.59	0.0032	mg/L	
Conc.2	-0.0019	0.0011	58.35	0.0032	mg/L	
Ints.	-24	12	50.81			
Single values (Ints.): #1: -36 #2: -26 #3: -11						
Se196.028						
Conc.1	0.0036	0.0013	37.31	0.0028	mg/L	
Conc.2	0.0062	0.0013	21.79	0.0028	mg/L	
Ints.	263	18	6.99			
Single values (Ints.): #1: 246 #2: 283 #3: 261						
BI 1	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:14
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.0009	0.0011	119.99	0.0016	mg/L	
Conc.2	0.0013	0.0011	84.48	0.0016	mg/L	
Ints.	98	17	17.93			
Single values (Ints.): #1: 99 #2: 79 #3: 114						
Sb217.581						
Conc.1	0.0008	0.0023	302.02	0.0033	mg/L	<KAL
Conc.2	-0.0039	0.0023	58.23	0.0033	mg/L	<KAL
Ints.	-47	26	54.34			<KAL
Single values (Ints.): #1: -72 #2: -21 #3: -48						
Se196.028						
Conc.1	0.0040	0.0011	27.81	0.0028	mg/L	
Conc.2	0.0065	0.0011	16.84	0.0028	mg/L	
Ints.	268	15	5.60			
Single values (Ints.): #1: 285 #2: 258 #3: 261						
BI 2	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:16
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	-0.0008	0.0018	231.17	0.0016	mg/L	<KAL
Conc.2	-0.0004	0.0018	462.09	0.0016	mg/L	<KAL
Ints.	71	28	40.08			<KAL
Single values (Ints.): #1: 51 #2: 58 #3: 103						

Operator: *Lab.:*

Line	*Mean*	*SD*	*RSD[%]*	*CI*	*Unit*	*Rem.*
Sb217.581						
Conc.1	0.0027	0.0030	112.04	0.0032	mg/L	<KAL
Conc.2	-0.0020	0.0030	149.09	0.0032	mg/L	<KAL
Ints.	-26	34	130.80			<KAL
Single values (Ints.): #1: -57 #2: 10 #3: -31						
Se196.028						
Conc.1	0.0012	0.0011	93.38	0.0028	mg/L	
Conc.2	0.0038	0.0011	29.65	0.0028	mg/L	
Ints.	230	15	6.63			
Single values (Ints.): #1: 238 #2: 213 #3: 240						
CCV(ST4 0.1mg/L)	*Method:*	As,Sb,Se(1)		Orig.	*Date:*	2026-03-20 12:18
Pre-DF: 1	*Wt.[g]:*		*Vol.[mL]:*	100		*Blank corr.:* on
As193.698						
Conc.1	0.1012	0.0027	2.65	0.0014	mg/L	
Conc.2	0.1015	0.0027	2.64	0.0014	mg/L	
Ints.	1718	43	2.52			
Single values (Ints.): #1: 1672 #2: 1724 #3: 1758						
Sb217.581						
Conc.1	0.0970	0.0024	2.45	0.0028	mg/L	
Conc.2	0.0923	0.0024	2.58	0.0028	mg/L	
Ints.	1033	27	2.59			
Single values (Ints.): #1: 1007 #2: 1061 #3: 1031						
Se196.028						
Conc.1	0.1038	0.0017	1.66	0.0025	mg/L	
Conc.2	0.1064	0.0017	1.62	0.0025	mg/L	
Ints.	1633	24	1.44			
Single values (Ints.): #1: 1647 #2: 1645 #3: 1605						