Quantitative Analysis Results Summary Rep

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin				
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

Naphthalene

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
Naphthalene ST-01.D	Cal	14.430	1001	5006	0.000
Naphthalene ST-02.D	Cal	14.430	5056	3652	0.007
Naphthalene ST-03.D	Cal	14.430	10325	5277	0.010
Naphthalene ST-04.D	Cal	14.430	56606	6266	0.050
Naphthalene ST-05.D	Cal	14.430	96866	5363	0.100
BK-.D	Blank	14.425	618	6408	0.000
FBK(KD 26-01-15).D	Blank	14.436	428	6063	0.000
KD-01.D	Sample	14.436	298	9170	0.000
MBK.D	Blank	14.436	221	6628	0.000
CCV N.D	QC	14.436	61694	6735	0.050

Naphthalene



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\Naphthalene ST-01.D	Calibration	1	x	0.001	1001
D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\Naphthalene ST-02.D	Calibration	2	x	0.005	5056
D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\Naphthalene ST-03.D	Calibration	3	x	0.010	10325
D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\Naphthalene ST-04.D	Calibration	4	x	0.050	56606
D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\CCV N.D	QC	4	x	0.050	61694
D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\Naphthalene ST-05.D	Calibration	5	x	0.100	96866

Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오전 9:30:02	**Data File**	Naphthalene ST-01.D
Type	Cal	**Name**	Naphthalene ST-01
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.430	1001	5006	0.000 mg/L

Naphthalene







Generated at 오후 3:05 on 2026-04-10

Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오전 9:52:51	**Data File**	Naphthalene ST-02.D
Type	Cal	**Name**	Naphthalene ST-02
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.430	5056	3652	0.007 mg/L

Naphthalene







Generated at 오후 3:05 on 2026-04-10

Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오전 10:15:42	**Data File**	Naphthalene ST-03.D
Type	Cal	**Name**	Naphthalene ST-03
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.430	10325	5277	0.010 mg/L

Naphthalene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오전 10:38:33	**Data File**	Naphthalene ST-04.D
Type	Cal	**Name**	Naphthalene ST-04
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.430	56606	6266	0.050 mg/L

Naphthalene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오전 11:01:25	**Data File**	Naphthalene ST-05.D
Type	Cal	**Name**	Naphthalene ST-05
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.430	96866	5363	0.100 mg/L

Naphthalene







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오후 1:48:34	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.425	618	6408	0.000 mg/L

Naphthalene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오후 2:08:51	**Data File**	FBK(KD 26-01-15).D
Type	Blank	**Name**	FBK(KD 26-01-15)
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.436	428	6063	0.000 mg/L

Naphthalene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오후 2:29:06	**Data File**	KD-01.D
Type	Sample	**Name**	KD-01
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.436	298	9170	0.000 mg/L

Naphthalene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오후 3:09:40	**Data File**	MBK.D
Type	Blank	**Name**	MBK
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.436	221	6628	0.000 mg/L

Naphthalene







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\VOC2(KD 1.15)\QuantResults\N A .batch.bin		
Analysis Time	2026-01-21 오후 4:55:50	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 3:05:42	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-21 오후 4:55:50	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-01-21 오후 3:29:54	**Data File**	CCV N.D
Type	QC	**Name**	CCV N
Dil.	1	**Acq. Method File**	VOCs(2) NEW

Sample Chromatogram





Name	RT	Resp.	ISTD Resp	Final Units Conc.
Naphthalene	14.436	61694	6735	0.050 mg/L

Naphthalene







Generated at 오후 3:05 on 2026-04-10