Quantitative Analysis Results Summary Rep

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin				
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

Acrylamide

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
ACRYL ST-01.D	Cal	4.074	1302	25513	0.007
ACRYL ST-02.D	Cal	4.074	2475	24411	0.011
ACRYL ST-03.D	Cal	4.074	3838	24635	0.015
ACRYL ST-04.D	Cal	4.068	5292	25225	0.019
ACRYL ST-05.D	Cal	4.068	10428	25594	0.033
BK-.D	Blank			25358	ND
FBK(RE 26-03-04).D	Blank			25260	ND
FBK(KD 26-03-04).D	Blank			25227	ND
RE-01.D	Sample			23496	ND
RE-02.D	Sample			23897	ND
RE-03.D	Sample			24079	ND
RE-04.D	Sample			24172	ND
RE-09.D	Sample			24545	ND
RE-10.D	Sample			23728	ND
RE-11.D	Sample			25957	ND
RE-12.D	Sample			25025	ND
KD-01.D	Sample			18775	ND
KD-02.D	Sample			25580	ND
KD-03.D	Sample			24374	ND
MBK.D	Blank			26814	ND
CCV.D	QC	4.068	4860	24295	0.018

Acrylamide



Cal. Path	Type	Level	Enable	Conc.	Response
\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\ACRYL ST-01.D	Calibration	1	x	3.000	1302
\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\ACRYL ST-02.D	Calibration	2	x	5.000	2475
\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\ACRYL ST-03.D	Calibration	3	x	8.000	3838
\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\CCV.D	QC	3	x	8.000	4860
\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\ACRYL ST-04.D	Calibration	4	x	10.000	5292
\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\ACRYL ST-05.D	Calibration	5	x	16.000	10428

Quant Sample Report

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 2:44:28	**Data File**	ACRYL ST-01.D
Type	Cal	**Name**	ACRYL ST-01
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.074	1302	25513	0.007 mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 2:53:14	**Data File**	ACRYL ST-02.D
Type	Cal	**Name**	ACRYL ST-02
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.074	2475	24411	0.011 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:01:59	**Data File**	ACRYL ST-03.D
Type	Cal	**Name**	ACRYL ST-03
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.074	3838	24635	0.015 mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:10:44	**Data File**	ACRYL ST-04.D
Type	Cal	**Name**	ACRYL ST-04
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.068	5292	25225	0.019 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:19:31	**Data File**	ACRYL ST-05.D
Type	Cal	**Name**	ACRYL ST-05
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.068	10428	25594	0.033 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:38:19	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			25358	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:47:04	**Data File**	FBK(RE 26-03-04).D
Type	Blank	**Name**	FBK(RE 26-03-04)
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			25260	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 3:55:50	**Data File**	FBK(KD 26-03-04).D
Type	Blank	**Name**	FBK(KD 26-03-04)
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			25227	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:04:37	**Data File**	RE-01.D
Type	Sample	**Name**	RE-01
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			23496	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:13:21	**Data File**	RE-02.D
Type	Sample	**Name**	RE-02
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			23897	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:22:07	**Data File**	RE-03.D
Type	Sample	**Name**	RE-03
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			24079	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:30:53	**Data File**	RE-04.D
Type	Sample	**Name**	RE-04
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			24172	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:39:38	**Data File**	RE-09.D
Type	Sample	**Name**	RE-09
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			24545	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:48:24	**Data File**	RE-10.D
Type	Sample	**Name**	RE-10
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			23728	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 4:57:07	**Data File**	RE-11.D
Type	Sample	**Name**	RE-11
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			25957	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:05:52	**Data File**	RE-12.D
Type	Sample	**Name**	RE-12
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			25025	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:14:38	**Data File**	KD-01.D
Type	Sample	**Name**	KD-01
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			18775	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:23:23	**Data File**	KD-02.D
Type	Sample	**Name**	KD-02
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			25580	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:32:08	**Data File**	KD-03.D
Type	Sample	**Name**	KD-03
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			24374	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:40:55	**Data File**	MBK.D
Type	Blank	**Name**	MBK
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			26814	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	\\192.168.254.121\data\2026\3. 2026-03\ACRYL(RE 3.4 KD 3.4)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-03-10 오후 4:38:15	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:14:59	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:38:15	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-06 오후 5:49:42	**Data File**	CCV.D
Type	QC	**Name**	CCV
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.068	4860	24295	0.018 mg/L

Acrylamide