Quantitative Analysis Results Summary Rep

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin				
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

Acrylamide

Data File	Type	RT	Resp.	ISTD Resp	Final Conc.
ACRYL ST-01.d	Cal	4.861	1794	10752	0.008
ACRYL ST-02.d	Cal	4.861	3172	9572	0.013
ACRYL ST-03.d	Cal	4.861	3512	9384	0.014
ACRYL ST-04.d	Cal	4.856	7755	9862	0.027
ACRYL ST-05.d	Cal	4.856	8299	9313	0.030
ACRYL ST-06.d	Cal	4.856	12188	9578	0.042
BK-.d	Blank			8371	ND
FBK(KD 26-02-04).d	Blank			8603	ND
FBK(JR 26-02-05).d	Blank			8884	ND
JR-01.d	Sample			4233	ND
JR-02.d	Sample			9307	ND
JR-03.d	Sample			5678	ND
JR-04.d	Sample			12073	ND
JR-05.d	Sample			5440	ND
JR-06.d	Sample			19792	ND
JR-07.d	Sample			25074	ND
JR-08.d	Sample			14099	ND
KD-01.d	Sample			13696	ND
KD-02.d	Sample			11967	ND
KD-04.d	Sample			9282	ND
MBK.d	Blank			10435	ND
-C-C-V-.d	QC	4.861	4915	11248	0.016

Acrylamide



Cal. Path	Type	Level	Enable	Conc.	Response
V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\ACRYL ST-01.d	Calibration	1	x	3.000	1794
V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\ACRYL ST-02.d	Calibration	2		5.000	3172
V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\ACRYL ST-03.d	Calibration	3	x	8.000	3512
V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\ACRYL ST-04.d	Calibration	4		10.000	7755
V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\-C-C-V-.d	QC	4	x	10.000	4915
V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\ACRYL ST-05.d	Calibration	5	x	16.000	8299
V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\ACRYL ST-06.d	Calibration	6	x	20.000	12188

Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 6:14:03	**Data File**	ACRYL ST-01.d
Type	Cal	**Name**	ACRYL ST-01
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.861	1794	10752	0.008 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 6:23:12	**Data File**	ACRYL ST-02.d
Type	Cal	**Name**	ACRYL ST-02
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.861	3172	9572	0.013 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 6:32:21	**Data File**	ACRYL ST-03.d
Type	Cal	**Name**	ACRYL ST-03
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.861	3512	9384	0.014 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 6:41:30	**Data File**	ACRYL ST-04.d
Type	Cal	**Name**	ACRYL ST-04
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.856	7755	9862	0.027 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 6:50:43	**Data File**	ACRYL ST-05.d
Type	Cal	**Name**	ACRYL ST-05
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.856	8299	9313	0.030 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 6:59:54	**Data File**	ACRYL ST-06.d
Type	Cal	**Name**	ACRYL ST-06
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.856	12188	9578	0.042 mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 7:18:06	**Data File**	BK-.d
Type	Blank	**Name**	BK-
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			8371	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 7:27:18	**Data File**	FBK(KD 26-02-04).d
Type	Blank	**Name**	FBK(KD 26-02-04)
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			8603	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 7:36:26	**Data File**	FBK(JR 26-02-05).d
Type	Blank	**Name**	FBK(JR 26-02-05)
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			8884	ND mg/L

Acrylamide







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 7:45:33	**Data File**	JR-01.d
Type	Sample	**Name**	JR-01
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			4233	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 7:54:43	**Data File**	JR-02.d
Type	Sample	**Name**	JR-02
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			9307	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 8:03:52	**Data File**	JR-03.d
Type	Sample	**Name**	JR-03
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			5678	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 8:13:02	**Data File**	JR-04.d
Type	Sample	**Name**	JR-04
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			12073	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 8:22:15	**Data File**	JR-05.d
Type	Sample	**Name**	JR-05
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			5440	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 8:31:24	**Data File**	JR-06.d
Type	Sample	**Name**	JR-06
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			19792	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 8:40:36	**Data File**	JR-07.d
Type	Sample	**Name**	JR-07
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			25074	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 8:49:44	**Data File**	JR-08.d
Type	Sample	**Name**	JR-08
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			14099	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 8:58:54	**Data File**	KD-01.d
Type	Sample	**Name**	KD-01
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			13696	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 9:08:08	**Data File**	KD-02.d
Type	Sample	**Name**	KD-02
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			11967	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 9:17:19	**Data File**	KD-04.d
Type	Sample	**Name**	KD-04
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			9282	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 9:26:29	**Data File**	MBK.d
Type	Blank	**Name**	MBK
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide			10435	ND mg/L

Acrylamide







Quant Sample Report

Batch Data Path	V:\2026\2026-02\ACRYL(KD 2.4, JR 2.5)\QuantResults\ACRYL.batch.bin		
Analysis Time	2026-02-10 오전 11:52:14	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:17:03	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-02-10 오전 11:52:13	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-02-09 오후 9:44:44	**Data File**	-C-C-V-.d
Type	QC	**Name**	-C-C-V-
Dil.	0.002	**Acq. Method File**	VOCs(Acrylamide)

Sample Chromatogram



Name	RT	Resp.	ISTD Resp	Final Units Conc.
Acrylamide	4.861	4915	11248	0.016 mg/L

Acrylamide