Quantitative Analysis Results Summary Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin				
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

Epichlorohydrin

Data File	Type	RT	Resp.	Final Conc.
EPIC ST-01.D	Cal	6.290	2175	0.200
EPIC ST-02.D	Cal	6.290	4118	0.487
EPIC ST-03.D	Cal	6.291	7781	1.028
EPIC ST-04.D	Cal	6.290	18244	2.573
EPIC ST-05.D	Cal	6.291	34427	4.962
BK-.D	Blank			ND
FBK(PL 25-12-30).D	Blank			ND
PL-01.D	Sample	6.291	0	ND
PL-02.D	Sample	6.291	0	ND
MBK-1.D	Blank			ND
CCV-1.D	QC	6.291	20340	2.882

Epichlorohydrin



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\EPIC ST-01.D	Calibration	1	x	0.250	2175
D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\EPIC ST-02.D	Calibration	2	x	0.500	4118
D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\EPIC ST-03.D	Calibration	3	x	1.000	7781
D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\EPIC ST-04.D	Calibration	4	x	2.500	18244
D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\CCV-1.D	QC	4	x	2.500	20340
D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\EPIC ST-05.D	Calibration	5	x	5.000	34427

Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 2:22:04	**Data File**	EPIC ST-01.D
Type	Cal	**Name**	EPIC ST-01
Dil.	1	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Name	RT	Resp.	Final Conc.	Units
Epichlorohydrin	6.290	2175	0.200	mg/L

Epichlorohydrin







Quant Sample Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 2:32:18	**Data File**	EPIC ST-02.D
Type	Cal	**Name**	EPIC ST-02
Dil.	1	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Name	RT	Resp.	Final Conc.	Units
Epichlorohydrin	6.290	4118	0.487	mg/L

Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 2:42:29	**Data File**	EPIC ST-03.D
Type	Cal	**Name**	EPIC ST-03
Dil.	1	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Name	RT	Resp.	Final Conc.	Units
Epichlorohydrin	6.291	7781	1.028	mg/L

Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 2:52:43	**Data File**	EPIC ST-04.D
Type	Cal	**Name**	EPIC ST-04
Dil.	1	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Name	RT	Resp.	Final Conc.	Units
Epichlorohydrin	6.290	18244	2.573	mg/L

Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 3:02:56	**Data File**	EPIC ST-05.D
Type	Cal	**Name**	EPIC ST-05
Dil.	1	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Name	RT	Resp.	Final Conc.	Units
Epichlorohydrin	6.291	34427	4.962	mg/L

Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 3:23:23	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	0.002	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 3:33:35	**Data File**	FBK(PL 25-12-30).D
Type	Blank	**Name**	FBK(PL 25-12-30)
Dil.	0.002	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 3:43:49	**Data File**	PL-01.D
Type	Sample	**Name**	PL-01
Dil.	0.002	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Name	RT	Resp.	Final Conc.	Units
Epichlorohydrin	6.291	0	ND	mg/L

Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 3:54:03	**Data File**	PL-02.D
Type	Sample	**Name**	PL-02
Dil.	0.002	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Name	RT	Resp.	Final Conc.	Units
Epichlorohydrin	6.291	0	ND	mg/L

Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 4:04:19	**Data File**	MBK-1.D
Type	Blank	**Name**	MBK-1
Dil.	0.002	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Epichlorohydrin







Quant Sample Report

Batch Data Path	D:\DATA\2026년\1. 2026년 1월\EPIC(PL 12.30)\QuantResults\EPIC.batch.bin		
Analysis Time	2026-01-08 오후 1:42:40	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:48:04	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-01-08 오후 1:42:40	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-01-06 오후 4:14:29	**Data File**	CCV-1.D
Type	QC	**Name**	CCV-1
Dil.	1	**Acq. Method File**	VOC(EPIC) NEW

Sample Chromatogram



Name	RT	Resp.	Final Conc.	Units
Epichlorohydrin	6.291	20340	2.882	mg/L

Epichlorohydrin