Quantitative Analysis Results Summary Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin				
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

Org-P

Data File	Type	RT	Resp.	Final Conc.	Exp. Conc.
Org-P ST-01.D	Cal	10.306	1748749	0.3010	0.250
Org-P ST-02.D	Cal	10.579	5591566	0.5071	0.500
Org-P ST-03.D	Cal	10.307	10711498	0.7817	0.750
Org-P ST-04.D	Cal	10.305	13779367	0.9462	1.000
Org-P ST-05.D	Cal	10.301	23629059	1.4745	1.500
Org-P ST-06.D	Cal	10.300	31871910	1.9165	2.000
Org-P ST-07.D	Cal	10.299	53435494	3.0730	3.000
BK-.D	Blank	0.000	0	ND	
FBK(DL 26-02-26).D	Blank	0.000	0	ND	
FBK(LT 26-02-26).D	Blank	0.000	0	ND	
FBK(PL 26-02-27).D	Blank	0.000	0	ND	
DL-01.D	Sample	11.026	0	ND	
LT-01.D	Sample	10.013	0	ND	
LT-02.D	Sample	0.000	0	ND	
LT-03.D	Sample	9.945	0	ND	
PL-01.D	Sample	0.000	0	ND	
PL-02.D	Sample	0.000	0	ND	
MBK-1.D	Blank	0.000	0	ND	
-C-C-V-1.D	QC	10.294	28353500	1.7278	1.500

Org-P **Relative Standard Error** 9.9



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\Org-P ST-01.D	Calibration	1	x	0.250	1748749
D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\Org-P ST-02.D	Calibration	2	x	0.500	5591566
D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\Org-P ST-03.D	Calibration	3	x	0.750	10711498
D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\Org-P ST-04.D	Calibration	4	x	1.000	13779367
D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\Org-P ST-05.D	Calibration	5	x	1.500	23629059
D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\-C-C-V-1.D	QC	5	x	1.500	28353500
D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\Org-P ST-06.D	Calibration	6	x	2.000	31871910
D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\Org-P ST-07.D	Calibration	7	x	3.000	53435494

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 5:44:58	**Data File**	Org-P ST-01.D
Type	Cal	**Name**	Org-P ST-01
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.306	1748749	0.3010 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 6:03:33	**Data File**	Org-P ST-02.D
Type	Cal	**Name**	Org-P ST-02
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.579	5591566	0.5071 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 6:22:04	**Data File**	Org-P ST-03.D
Type	Cal	**Name**	Org-P ST-03
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.307	10711498	0.7817 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 6:40:36	**Data File**	Org-P ST-04.D
Type	Cal	**Name**	Org-P ST-04
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.305	13779367	0.9462 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 6:59:09	**Data File**	Org-P ST-05.D
Type	Cal	**Name**	Org-P ST-05
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.301	23629059	1.4745 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 7:17:44	**Data File**	Org-P ST-06.D
Type	Cal	**Name**	Org-P ST-06
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.300	31871910	1.9165 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 7:36:17	**Data File**	Org-P ST-07.D
Type	Cal	**Name**	Org-P ST-07
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.299	53435494	3.0730 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 8:13:24	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-03-04 오후 8:31:56	**Data File**	FBK(DL 26-02-26).D
Type	Blank	**Name**	FBK(DL 26-02-26)
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 8:50:32	**Data File**	FBK(LT 26-02-26).D
Type	Blank	**Name**	FBK(LT 26-02-26)
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 9:09:08	**Data File**	FBK(PL 26-02-27).D
Type	Blank	**Name**	FBK(PL 26-02-27)
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 9:27:43	**Data File**	DL-01.D
Type	Sample	**Name**	DL-01
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	11.026	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 9:46:18	**Data File**	LT-01.D
Type	Sample	**Name**	LT-01
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.013	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-03-04 오후 10:04:50	**Data File**	LT-02.D
Type	Sample	**Name**	LT-02
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 10:23:22	**Data File**	LT-03.D
Type	Sample	**Name**	LT-03
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	9.945	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 10:41:53	**Data File**	PL-01.D
Type	Sample	**Name**	PL-01
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 11:00:24	**Data File**	PL-02.D
Type	Sample	**Name**	PL-02
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-04 오후 11:18:56	**Data File**	MBK-1.D
Type	Blank	**Name**	MBK-1
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(DL 2.26 LT 2.26 PL 2.27)\QuantResults\ORGP.batch.bin		
Analysis Time	2026-03-05 오후 1:53:30	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:54	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-05 오후 1:53:30	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-05 오전 12:14:33	**Data File**	-C-C-V-1.D
Type	QC	**Name**	-C-C-V-1
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.294	28353500	1.7278 mg/L