Quantitative Analysis Results Summary Report

Agilent | Trusted Answers

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin				
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Analyze Quant Version**	10.2	**Report Quant Version**	10.2

Org-P

Data File	Type	RT	Resp.	Final Conc.	Exp. Conc.
Org-P ST-01.D	Cal	10.302	6722301	0.3005	0.250
Org-P ST-02.D	Cal	10.291	20131395	0.5642	0.500
Org-P ST-03.D	Cal	10.289	24911674	0.6582	0.750
Org-P ST-04.D	Cal	10.288	37142642	0.8987	1.000
Org-P ST-05.D	Cal	10.288	67274755	1.4912	1.500
Org-P ST-06.D	Cal	10.287	99401128	2.1229	2.000
Org-P ST-07.D	Cal	10.287	142193935	2.9643	3.000
BK-.D	Blank	0.000	0	ND	
FBK(RE 26-03-04).D	Blank	0.000	0	ND	
FBK(KD 26-03-04).D	Blank	0.000	0	ND	
RE-01.D	Sample	0.000	0	ND	
RE-02.D	Sample	0.000	0	ND	
RE-03.D	Sample	7.948	0	ND	
RE-04.D	Sample	0.000	0	ND	
RE-09.D	Sample	0.000	0	ND	
RE-10.D	Sample	0.000	0	ND	
RE-11.D	Sample	0.000	0	ND	
RE-12.D	Sample	0.000	0	ND	
KD-01.D	Sample	0.000	0	ND	
KD-02.D	Sample	0.000	0	ND	
KD-03.D	Sample	0.000	0	ND	
MBK.D	Blank	0.000	0	ND	
C-C-V.D	QC	10.276	58312285	1.3150	1.500

Org-P **Relative Standard Error** 13.2



Cal. Path	Type	Level	Enable	Conc.	Response
D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\Org-P ST-01.D	Calibration	1	x	0.250	6722301
D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\Org-P ST-02.D	Calibration	2	x	0.500	20131395
D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\Org-P ST-03.D	Calibration	3	x	0.750	24911674
D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\Org-P ST-04.D	Calibration	4	x	1.000	37142642
D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\Org-P ST-05.D	Calibration	5	x	1.500	67274755
D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\C-C-V.D	QC	5	x	1.500	58312285
D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\Org-P ST-06.D	Calibration	6	x	2.000	99401128
D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\Org-P ST-07.D	Calibration	7	x	3.000	142193935

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 12:33:29	**Data File**	Org-P ST-01.D
Type	Cal	**Name**	Org-P ST-01
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.302	6722301	0.3005 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 12:51:58	**Data File**	Org-P ST-02.D
Type	Cal	**Name**	Org-P ST-02
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.291	20131395	0.5642 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 1:10:28	**Data File**	Org-P ST-03.D
Type	Cal	**Name**	Org-P ST-03
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.289	24911674	0.6582 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 1:28:59	**Data File**	Org-P ST-04.D
Type	Cal	**Name**	Org-P ST-04
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.288	37142642	0.8987 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 1:47:29	**Data File**	Org-P ST-05.D
Type	Cal	**Name**	Org-P ST-05
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.288	67274755	1.4912 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 2:06:03	**Data File**	Org-P ST-06.D
Type	Cal	**Name**	Org-P ST-06
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.287	99401128	2.1229 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 2:24:34	**Data File**	Org-P ST-07.D
Type	Cal	**Name**	Org-P ST-07
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.287	142193935	2.9643 mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 3:01:39	**Data File**	BK-.D
Type	Blank	**Name**	BK-
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-03-07 오후 3:20:11	**Data File**	FBK(RE 26-03-04).D
Type	Blank	**Name**	FBK(RE 26-03-04)
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 3:38:43	**Data File**	FBK(KD 26-03-04).D
Type	Blank	**Name**	FBK(KD 26-03-04)
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-03-07 오후 3:57:15	**Data File**	RE-01.D
Type	Sample	**Name**	RE-01
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 4:15:47	**Data File**	RE-02.D
Type	Sample	**Name**	RE-02
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2
Acq. Date-Time	2026-03-07 오후 4:34:23	**Data File**	RE-03.D
Type	Sample	**Name**	RE-03
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	7.948	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 4:52:57	**Data File**	RE-04.D
Type	Sample	**Name**	RE-04
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 5:11:32	**Data File**	RE-09.D
Type	Sample	**Name**	RE-09
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 5:30:09	**Data File**	RE-10.D
Type	Sample	**Name**	RE-10
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 5:48:42	**Data File**	RE-11.D
Type	Sample	**Name**	RE-11
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 6:07:18	**Data File**	RE-12.D
Type	Sample	**Name**	RE-12
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 6:25:53	**Data File**	KD-01.D
Type	Sample	**Name**	KD-01
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 6:44:26	**Data File**	KD-02.D
Type	Sample	**Name**	KD-02
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 7:02:58	**Data File**	KD-03.D
Type	Sample	**Name**	KD-03
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 7:21:35	**Data File**	MBK.D
Type	Blank	**Name**	MBK
Dil.	0.002	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	0.000	0	ND mg/L

Quant Sample Report

Batch Data Path	D:\DATA\2026년\3. 2026년 3월\Org-P(RE 3.4 KD 3.4)\QuantResults\ORG-P.batch.bin		
Analysis Time	2026-03-10 오후 4:52:24	**Analyst Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Report Generation Time	2026-04-10 오후 2:43:26	**Report Generator Name**	DESKTOP-N82KMQD\수질분석센터 GC2
Calibration Last Update	2026-03-10 오후 4:52:24	**Batch State**	Processed
Analyze Quant Version	10.2	**Report Quant Version**	10.2

Acq. Date-Time	2026-03-07 오후 7:58:41	**Data File**	C-C-V.D
Type	QC	**Name**	C-C-V
Dil.	1	**Acq. Method File**	ORG(Org-P)



Name	RT	Resp.	Final Conc. Units
Org-P	10.276	58312285	1.3150 mg/L