SERION ELISA classic ESR117A

CANDIDA ALBICANS IgA

SMH.DK

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot SMH.DK IFU-Version 117-18

Verw. bis / Exp. 2019-12

05.01.2018

Prüfdatum /

Date of control



80

60

Verwendete Reagenzien / Reagents used	Standard							Standard Kur	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	SLH.DH	Ref Werte / Re	f. Values		Gültigke	itsbereich	/ Valid	dity Range	Parameter	Α	0,041
Standardserum / Standard serum	SMH.BO	OE	0,86		OD	0,43	-	1,46		В	1,136
Negativ Kontrolle / Negative control	SMH.BN									С	5,061
Konjugat / Conjugate	SIH.CE+	Units	s 48,5 U	/ml						D	3,987
Quantifizierungsgrenzen / Limits of quantific	U/ml	10	-	225							
Grenzwertbereich / Borderline range	U/ml	60	-	80							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																			
0,43	} -	0,47	0,48	- 0,53	0,54	- 0,58	0,59	- 0,64	0,65	- 0,69	0,70	- 0,74	0,75	- 0,80	0,81	- 0,85	0,86	U/ml	Interpretation
	<	0,54		< 0,61		< 0,67		< 0,73		< 0,80		< 0,86		< 0,93		< 0,99	< 1,03	< 60,0	neg
0,54	-	0,68	0,61	- 0,76	0,67	- 0,84	0,73	- 0,92	0,80	- 1,00	0,86	- 1,08	0,93	- 1,16	0,99	- 1,24	1,03 - 1,29	60,0 - 80,0	gw / borderline
	>	0,68		> 0,76		> 0,84		> 0,92		> 1,00		> 1,08		> 1,16		> 1,24	> 1,29	> 80,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum												
U/ml	0,86	0,87 - 0,94	0,95 - 1,01	1,02 - 1,09	1,10 - 1,16	1,17 - 1,24	1,25 - 1,31	1,32 - 1,39 1,40 - 1,	16 Interpretation			
< 60,0	< 1,03	< 1,08	< 1,17	< 1,26	< 1,35	< 1,44	< 1,53	< 1,62 < 1,	71 neg			
60,0 - 80,0	1,03 - 1,29	1,08 - 1,35	1,17 - 1,47	1,26 - 1,58	1,35 - 1,69	1,44 - 1,81	1,53 - 1,92		14 gw / borderline			
> 80,0	> 1,29	> 1,35	> 1,47	> 1,58	> 1,69	> 1,81	> 1,92	> 2,03 > 2,	14 pos			

Formeln für spezielle Auswertesysteme Special case formulas OD = 1,498 x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = 1,195 x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = exp(5,061-ln(3,946/(MV(Sample)\ x0,86/\ MV(STD)-0,041)-1)/1,136)$

Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(5.061-ln(3.946/(Sample*0.860/S-0.041)-1)/1.136)



Gültigkeitsbereich / Validity Range

0.430<=S1<=1.462



If OD Sample < Parameter A

if Ti < (0.041*(S1/0.860)) then Ti = (0.041+0.001)*(S1/0.860)



If OD Sample > Parameter D

if Ti>(3.987*(S1/0.860)) then Ti=(3.987-0.001)*(S1/0.860)



If OD Negative control < Parameter A

if NC1<(0.041*(S1/0.860)) then NCi=(0.041+0.001)*(S1/0.860)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg