SERION ELISA classic

ESR1363G

EPSTEIN-BARR VIRUS EA IgG

EK0134

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot Verw. bis / Exp.

EK0134 2021-10-31 IFU-Version 136-25

22.11.2019

!New!

Prüfdatum /

Date of control



10

| Verwendete Reagenzien / Reagents used | Standard | | | | | | | Standard Kui | Standard Kurve / Standard curve | | | |
|--|-----------|----------------|----------------|------|-----|----------|------------|--------------|---------------------------------|-----------|---|-------|
| Teststreifen / Antigen coated strips | ECK0271 | Ref Werte / Re | ef. Valu | es | 0 | Gültigke | itsbereich | / Vali | dity Range | Parameter | Α | 0,011 |
| Standardserum / Standard serum | ECK0269 | 0 | D 0,98 | | | OD | 0,49 | - | 1,67 | | В | 1,088 |
| Negativ Kontrolle / Negative control | ECK0267 | | | | | | | | | | С | 4,248 |
| Konjugat / Conjugate | KJK008+++ | Unit | ts 48,3 | U/ml | | | | | | | D | 2,431 |
| Quantifizierungsgrenzen / Limits of quantifi | U/ml | 2 | | | 400 | | | | | | | |
| Grenzwertbereich / Borderline range | U/ml | 1 | 0 | • | 15 | | | | | | | |

| OD Bereich / OD Range 405 nm, Standardserum / Standard serum | | | | | | | | | | | | | | | | | | | |
|--|-----|------|------|--------|------|--------|------|--------|------|--------|------|--------|------|--------|------|--------|-------------|------------|-------------------|
| 0,49 |) - | 0,54 | 0,55 | - 0,60 | 0,61 | - 0,66 | 0,67 | - 0,73 | 0,74 | - 0,79 | 0,80 | - 0,85 | 0,86 | - 0,91 | 0,92 | - 0,97 | 0,98 | U/ml | Interpretation |
| | < | 0,14 | | < 0,16 | | < 0,18 | | < 0,19 | | < 0,21 | | < 0,23 | | < 0,24 | | < 0,26 | < 0,27 | < 10, | 0 neg |
| 0,14 | - | 0,21 | 0,16 | - 0,23 | 0,18 | - 0,25 | 0,19 | - 0,28 | 0,21 | - 0,30 | 0,23 | - 0,33 | 0,24 | - 0,35 | 0,26 | - 0,38 | 0,27 - 0,39 | 10,0 - 15, | 0 gw / borderline |
| | > | 0,21 | | > 0,23 | | > 0,25 | | > 0,28 | | > 0,30 | | > 0,33 | | > 0,35 | | > 0,38 | > 0,39 | > 15, | 0 pos |

| OD Bereich / OD Range 405 nm, Standardserum / Standard serum | | | | | | | | | | | |
|--|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|-------------------------|------------------------|--|--|
| U/ml | 0,98 | 0,99 - 1,07 | 1,08 - 1,15 | 1,16 - 1,24 | 1,25 - 1,32 | 1,33 - 1,41 | 1,42 - 1,49 | 1,50 - 1,58 1,59 - 1,67 | Interpretation | | |
| < 10,0 | < 0,27 | < 0,28 | < 0,31 | < 0,33 | < 0,35 | < 0,38 | < 0,40 | < 0,42 < 0,45 | neg | | |
| 10,0 - 15,0 > 15,0 | 0,27 - 0,39 > 0,39 | 0,28 - 0,41 > 0,41 | 0,31 - 0,44 > 0,44 | 0,33 - 0,48 > 0.48 | 0,35 - 0,51 > 0,51 | 0,38 - 0,55 > 0,55 | 0,40 - 0,58 > 0,58 | | gw / borderline pos | | |

Formeln für spezielle Auswertesysteme Special case formulas

OD =

0,401 x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off

0,277 x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off OD =

Concentration= exp(4,248-In(2,42/(MV(Sample) x0,98/ MV(STD)-0,011)-1)/1,088)

15 Institut Virion\Serion GmbH

> Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(4.248-ln(2.420/(Sample*0.980/S-0.011)-1)/1.088)



Gültigkeitsbereich / Validity Range

0.490<=S1<=1.666



If OD Sample < Parameter A

if Ti<(0.011*(S1/0.980)) then Ti=(0.011+0.001)*(S1/0.980)



If OD Sample > Parameter D

if Ti>(2.431*(S1/0.980)) then Ti=(2.431-0.001)*(S1/0.980)



If OD Negative control < Parameter A

if NC1<(0.011*(S1/0.980)) then NCi=(0.011+0.001)*(S1/0.980)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg