SERION ELISA classic

ESR1371A

CHLAMYDIA PNEUMONIAE IgA

SBK.FR

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot SBK.FR IFU-Version 1371-5

Verw. bis / Exp. 2020-11

!New!

Prüfdatum /
Date of control

28.02.2019



10

Verwendete Reagenzien / Reagents used	Standard		Standard Kui	Standard Kurve / Standard curve							
Teststreifen / Antigen coated strips	SDI.BG	Ref Werte / Re	ef. Value	es	Gültigl	eitsbereich	ı / Vali	idity Range	Parameter	Α	0,078
Standardserum / Standard serum	SLI.CS	OI	D 0,99		OD	0,50	-	1,68		В	1,074
Negativ Kontrolle / Negative control	SLI.CR									С	5,061
Konjugat / Conjugate	SAK.EM++	Unit	s 35,5	U/ml						D	5,510
Quantifizierungsgrenzen / Limits of quantifi	U/ml	4	-	150							
Grenzwertbereich / Borderline range	U/ml	1	0 -	13							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																				
0,50	-	0,55	0,56	- 0,61	0,62	- 0,67	0,68	- 0,73	0,74	- 0,79	0,80	- 0,86	0,87	- 0,92	0,93	- 0,98	0,99	U/ml		Interpretation
	<	0,18		< 0,20		< 0,22		< 0,24		< 0,26		< 0,29		< 0,31		< 0,33	< 0,34	. <	10,0	neg
0,18	-	0,23	0,20	- 0,25	0,22	- 0,28	0,24	- 0,31	0,26	- 0,33	0,29	- 0,36	0,31	- 0,39	0,33	- 0,41	0,34 - 0,43	10,0	- 13,0	gw / borderline
	>	0,23		> 0,25		> 0,28		> 0,31		> 0,33		> 0,36		> 0,39		> 0,41	> 0,43	;	> 13,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum										
U/ml	0,99	1,00 - 1,08	1,09 - 1,16	1,17 - 1,25	1,26 - 1,34	1,35 - 1,42	1,43 - 1,51	1,52 - 1,60 1,61 - 1,68	Interpretation	
< 10,0	< 0,34	< 0,36	< 0,39	< 0,42	< 0,45	< 0,48	< 0,51	< 0,54 < 0,56	neg	
10,0 - 13,0 > 13,0	0,34 - 0,43 > 0.43	0,36 - 0,45 > 0,45	0,39 - 0,49 > 0,49	0,42 - 0,53 > 0,53	0,45 - 0,56 > 0,56	0,48 - 0,60 > 0,60	0,51 - 0,64 > 0,64	0,54 - 0,68	gw / borderline	

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,431** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,348** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = \exp(5,061 - \ln(5,432/(MV(Sample)\ x0,99/\ MV(STD) - 0,078) - 1)/1,074)$

13 Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(5.061-ln(5.432/(Sample*0.990/S-0.078)-1)/1.074)



Gültigkeitsbereich / Validity Range

0.495<=S1<=1.683



If OD Sample < Parameter A

if Ti<(0.078*(S1/0.990)) then Ti=(0.078+0.001)*(S1/0.990)



If OD Sample > Parameter D

if Ti>(5.510*(S1/0.990)) then Ti=(5.510-0.001)*(S1/0.990)



If OD Negative control < Parameter A

if NC1<(0.078*(S1/0.990)) then NCi=(0.078+0.001)*(S1/0.990)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg