SERION ELISA classic

ESR1051M

HERPES SIMPLEX VIRUS 1 IgM

EK0111

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot EK0111 IFU-Version 105-24

Verw. bis / Exp. 2021-08-31 !New!

07.10.2019

Prüfdatum /

Date of control



30

20

Verwendete Reagenzien / Reagents used	Standard							Standard Kur	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	ECK0170	Ref Werte / Re	f. Values		Gültigke	itsbereich	/ Valid	dity Range	Parameter	Α	-0,013
Standardserum / Standard serum	ECK0174	O	0,75		OD	0,38	-	1,28		В	0,855
Negativ Kontrolle / Negative control	ECK0173									С	4,496
Konjugat / Conjugate	KJK002++	Units	s 27,4 U	/ml						D	2,850
Quantifizierungsgrenzen / Limits of quantific	U/ml	10	-	500							
Grenzwertbereich / Borderline range	U/ml	20	-	30							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																			
0,38	-	0,41	0,42	- 0,46	0,47	- 0,51	0,52	- 0,55	0,56	- 0,60	0,61	- 0,65	0,66	- 0,69	0,70	- 0,74	0,75	U/ml	Interpretation
	<	0,32		< 0,36		< 0,40		< 0,43		< 0,47		< 0,51		< 0,55		< 0,59	< 0,61	< 20,0	neg
0,32	-	0,41	0,36	- 0,46	0,40	- 0,51	0,43	- 0,56	0,47	- 0,61	0,51	- 0,66	0,55	- 0,71	0,59	- 0,76	0,61 - 0,79	20,0 - 30,0	gw / borderline
	>	0,41		> 0,46		> 0,51		> 0,56		> 0,61		> 0,66		> 0,71		> 0,76	> 0,79	> 30,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum											
U/ml	0,75	0,76 - 0,82	0,83 - 0,88	0,89 - 0,95	0,96 - 1,01	1,02 - 1,08	1,09 - 1,14	1,15 - 1,21 1,22 - 1,28	Interpretation		
< 20,0	< 0,61	< 0,64	< 0,69	< 0,75	< 0,80	< 0,85	< 0,91	< 0,96 < 1,01	neg		
20,0 - 30,0 > 30,0	0,61 - 0,79 > 0,79	0,64 - 0,83 > 0,83	0,69 - 0,90 > 0,90	0,75 - 0,97 > 0,97	0,80 - 1,04 > 1,04	0,85 - 1,11 > 1,11	0,91 - 1,18 > 1.18		gw / borderline pos		

Formeln für spezielle Auswertesysteme Special case formulas OD = 1,058 x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = 0,811 x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = \exp(4,496 - \ln(2,863/(\text{MV(Sample}) \ x0,75/\ \text{MV(STD)} + 0,013) - 1)/0,855)$

Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19

D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(4.496-ln(2.863/(Sample*0.750/S+0.013)-1)/0.855)



Gültigkeitsbereich / Validity Range

0.375<=S1<=1.275



If OD Sample < Parameter A

if Ti<(-0.013*(S1/0.750)) then Ti=(-0.013+0.001)*(S1/0.750)



If OD Sample > Parameter D

if Ti>(2.850*(S1/0.750)) then Ti=(2.850-0.001)*(S1/0.750)



If OD Negative control < Parameter A

if NC1<(-0.013*(S1/0.750)) then NCi=(-0.013+0.001)*(S1/0.750)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg