SERION ELISA classic

ESR103G

MUMPS VIRUS / PAROTITIS / OREILLONS IgG

SAK.BX

Qualitätskontrollzertifikat / Quality Control Certificate

2020-12

Kitcharge / Lot SAK.BX IFU-Version 103-13

Verw. bis / Exp.

!New!

14.01.2019

Prüfdatum /

Date of control



Verwendete Reagenzien / Reagents used	Standard							Standard Ku	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	SKI.ET	Ref Werte / Re	f. Values		Gültigke	itsbereich	/ Vali	dity Range	Parameter	Α	-0,011
Standardserum / Standard serum	SMI.BM	00	0,97		OD	0,49	-	1,65		В	0,921
Negativ Kontrolle / Negative control	SMI.BL									С	6,350
Konjugat / Conjugate	SMI.CU+++	Units	s 415 U /m	ıl						D	2,288
Quantifizierungsgrenzen / Limits of quantifi	U/ml	30	-	2000							
Grenzwertbereich / Borderline range	U/ml	70	-	100							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																				
0,49	-	- 0,5	1 0,	,55	- 0,60	0,61	- 0,66	0,67	- 0,72	0,73	- 0,78	0,79	- 0,84	0,85	- 0,90	0,91	- 0,96	0,97	U/ml	Interpretation
	<	0,1	5 T		< 0,16		< 0.18		< 0.20		< 0,22	1	< 0.23		< 0.25		< 0,27	< 0,28	< 70,0	neg
0,15	· -	- 0,1	0,	,16	- 0,22	0,18	- 0,24	0,20	- 0,26	0,22	- 0,29	0,23	- 0,31	0,25	- 0,33	0,27	- 0,36	0,28 - 0,37		gw / borderline
	>	> 0,1)		> 0,22		> 0,24		> 0,26		> 0,29		> 0,31		> 0,33		> 0,36	> 0,37	> 100,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum											
U/mI	0,97	0,98 - 1,05	1,06 - 1,14 1,1	15 - 1,22	1,23 - 1,31	1,32 - 1,39	1,40 - 1,48	1,49 - 1,56 1,57 - 1,65	Interpretation		
< 70,0	< 0,28	< 0,29	< 0,32	< 0,34	< 0,37	< 0,39	< 0,42	< 0,44 < 0,47	neg		
70,0 - 100,0 > 100,0	0,28 - 0,37 > 0,37	0,29 - 0,39 > 0,39	0,32 - 0,42 0,3 > 0,42	34 - 0,45 > 0,45	0,37 - 0,49 > 0,49	0,39 - 0,52 > 0,52	0,42 - 0,55 > 0.55	0,44 - 0,58 0,47 - 0,61 > 0,58 > 0,61	gw / borderline		

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,384** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,288** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(6,35-ln(2,299/(MV(Sample) x0,97/ MV(STD)+0,011)-1)/0,921)

Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(6.350-ln(2.299/(Sample*0.970/S+0.011)-1)/0.921)



Gültigkeitsbereich / Validity Range

0.485<=S1<=1.649



If OD Sample < Parameter A

if Ti < (-0.011*(S1/0.970)) then Ti = (-0.011+0.001)*(S1/0.970)



If OD Sample > Parameter D

if Ti>(2.288*(S1/0.970)) then Ti=(2.288-0.001)*(S1/0.970)



If OD Negative control < Parameter A

if NC1<(-0.011*(S1/0.970)) then NCi=(-0.011+0.001)*(S1/0.970)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg