SERION ELISA classic ESR133M ENTEROVIRUS IgM EK0123

## **Qualitätskontrollzertifikat / Quality Control Certificate**

Kitcharge / Lot EK0123 IFU-Version 133-10

Verw. bis / Exp. 2021-08-31 !New!

28.10.2019

Prüfdatum /

Date of control



15

10

Verwendete Reagenzien / Reagents used	Standard							Standard Kui	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	ECK0193	Ref Werte / Ref	f. Values		Gültigke	itsbereich	/ Vali	dity Range	Parameter	Α	0,003
Standardserum / Standard serum	ECK0200	OD	0,80		OD	0,40	-	1,36		В	0,953
Negativ Kontrolle / Negative control	ECK0199									С	5,360
Konjugat / Conjugate	KJK010++	Units	45,5 U/	ml						D	4,266
Quantifizierungsgrenzen / Limits of quantific	U/ml	5	_	250							
Grenzwertbereich / Borderline range	U/ml	10	-	15							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																			
0,40	) -	0,44	0,45	- 0,49	0,50	- 0,54	0,55	- 0,59	0,60	- 0,64	0,65	- 0,69	0,70	- 0,74	0,75	- 0,79	0,80	U/ml	Interpretation
	<	0,12		< 0,13		< 0,14		< 0,16		< 0,17		< 0,18		< 0,20		< 0,21	< 0,22	< 10,	0 neg
0,12	<u> </u>	0,17	0,13	- 0,19	0,14	- 0,21	0,16	- 0,23	0,17	- 0,25	0,18	- 0,27	0,20	- 0,29	0,21	- 0,31	0,22 - 0,32	10,0 - 15,	gw / borderline
	>	0,17		> 0,19		> 0,21		> 0,23		> 0,25		> 0,27		> 0,29		> 0,31	> 0,32	> 15,	0 pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum											
U/ml	0,80	0,81 - 0,87	0,88 - 0,94	0,95 - 1,01	1,02 - 1,08	1,09 - 1,15	1,16 - 1,22	1,23 - 1,29 1,30 - 1,36	Interpretation		
< 10,0	< 0,22	< 0,23	< 0,25	< 0,27	< 0,29	< 0,31	< 0,33	< 0,35 < 0,37	neg		
10,0 - 15,0 > 15,0	0,22 - 0,32 > 0.32	0,23 - 0,34 > 0,34	0,25 - 0,36 > 0,36	0,27 - 0,39 > 0,39	0,29 - 0,42 > 0,42	0,31 - 0,45 > 0,45	0,33 - 0,48 > 0.48		gw / borderline pos		

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,398** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off

OD = **0,278** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(5,36-ln(4,263/(MV(Sample) x0,8/ MV(STD)-0,003)-1)/0,953)

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

# Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

#### 4PS- Formel / 4PS-formula

exp(5.360-ln(4.263/(Sample\*0.800/S-0.003)-1)/0.953)



#### Gültigkeitsbereich / Validity Range

0.400<=S1<=1.360



#### If OD Sample < Parameter A

if Ti < (0.003\*(S1/0.800)) then Ti = (0.003+0.001)\*(S1/0.800)



# If OD Sample > Parameter D

if Ti>(4.266\*(S1/0.800)) then Ti=(4.266-0.001)\*(S1/0.800)



### If OD Negative control < Parameter A

if NC1<(0.003\*(S1/0.800)) then NCi=(0.003+0.001)\*(S1/0.800)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg