SERION ELISA classic ESR138M YERSINIA IgM SCK.BX

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot SCK.BX IFU-Version 138-14

Verw. bis / Exp. 2021-02 !New! Prüfdatum /

15

10

Date of control

19.03.2019

Verwendete Reagenzien / Reagents used	Lot	Standard			Standard Kur	Standard Kurve / Standard curve					
Teststreifen / Antigen coated strips	SBK.DD	Ref Werte / Ref	. Values		Gültigkei	itsbereich	/ Valid	dity Range	Parameter	Α	0,028
Standardserum / Standard serum	SBK.CW	OD	0,91		OD	0,46	-	1,55		В	1,526
Negativ Kontrolle / Negative control	SBK.CV									С	3,954
Konjugat / Conjugate	SHI.FL+	Units	26,7 U/r	nl						D	3,364
Quantifizierungsgrenzen / Limits of quantific	U/ml	5	-	100							
Grenzwertbereich / Borderline range	U/ml	10	-	15							

OD B	D Bereich / OD Range 405 nm, Standardserum / Standard serum																		
0,46	; -	0,50	0,51	- 0,56	0,57	- 0,62	0,63	- 0,67	0,68	- 0,73	0,74	- 0,79	0,80	- 0,84	0,85	- 0,90	0,91	U/ml	Interpretation
	<	0,15		< 0,16		< 0,18		< 0,20		< 0,22		< 0,23		< 0,25		< 0,27	< 0,28	< 1	0,0 neg
0,15	; -	0,24	0,16	- 0,27	0,18	- 0,30	0,20	- 0,33	0,22	- 0,36	0,23	- 0,39	0,25	- 0,41	0,27	- 0,44	0,28 - 0,46	10,0 - 1	5,0 gw / borderline
	>	0,24		> 0,27		> 0,30		> 0,33		> 0,36		> 0,39		> 0,41		> 0,44	> 0,46	> 1	5,0 pos

OD Be	D Bereich / OD Range 405 nm, Standardserum / Standard serum																										
U/ml			(),91		0,92	- 0),99	1,00	- 1,0	7 1,0	3 -	1,15	1,16	- 1,23	1,24	- 1,31	1,32	-	1,39	1,40	-	1,47	1,48	-	1,55	Interpretation
	<	10,0		<	0,28		< 0),29		< 0,32	2	<	0,34		< 0,37		< 0,39		<	0,42		<	0,44		<	0,47	neg
10,0	- >	15,0 15,0	0,28		0,46 0,46	0,29),48),48	0,32	- 0,52 > 0,52			0,56 0,56	0,37	- 0,60 > 0,60	0,39	- 0,64 > 0,64	0,42	->	0,68 0,68	0,44		0,72 0,72	0,47	- >	0,76 0,76	gw / borderline pos

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,507** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,304** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(3,954-In(3,336/(MV(Sample) x0,91/ MV(STD)-0,028)-1)/1,526)

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(3.954-ln(3.336/(Sample*0.910/S-0.028)-1)/1.526)



Gültigkeitsbereich / Validity Range

0.455<=S1<=1.547



If OD Sample < Parameter A

if Ti < (0.028*(S1/0.910)) then Ti = (0.028+0.001)*(S1/0.910)



If OD Sample > Parameter D

if Ti>(3.364*(S1/0.910)) then Ti=(3.364-0.001)*(S1/0.910)



If OD Negative control < Parameter A

if NC1<(0.028*(S1/0.910)) then NCi=(0.028+0.001)*(S1/0.910)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg