virion\serion

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* RESP. SYNCYTIAL VIRUS IGM

BestNr. / Order No.	ESR113M
Chargen-Nr. / Lot No.	EK0032
Verfallsdatum / Expiry Date	2021-04
IFU-Version / IFU-Version	NEW! 113-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	SDK.AM	0,80	> 0,40
Cut off-Serum / Cut off serum	SDK.AN	0,30	0,15 - 0,60

Negativkontrolle / Negative Control	RESP. SYNCYTIAL VIRUS IgM	
Chargen-Nr. / Lot No.	SDK.AL	
Teststreifen / Strips	RESP. SYNCYTIAL VIRUS IgM	
Chargen-Nr. / Lot No.	SDK.AW	
Konjugat / Conjugate	IgM +	
Chargen-Nr. / Lot No.	SAK.EI+	

22.05.2019

Prüfdatum / Date of control

Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100

E-Mail: info@virion-serion.de Internet: www.virion-serion.de