SERION ELISA classic

ESR109M

CYTOMEGALOVIRUS IgM

EK0133

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot Verw. bis / Exp. EK0133 2021-01-31 IFU-Version 109-20

OD =

!New!

15.11.2019

Prüfdatum /

Date of control



15

10

Verwendete Reagenzien / Reagents used	Lot	Standard			Standard Kur	ve / Stand	dard curve				
Teststreifen / Antigen coated strips	Ref Werte / Ref	f. Values		Gültigke	itsbereich	/ Valid	lity Range	Parameter	Α	-0,134	
Standardserum / Standard serum	OD	0,87		OD 0,44 - 1			1,48		В	0,652	
Negativ Kontrolle / Negative control	SCK.BH									С	4,266
Konjugat / Conjugate	SAK.EI+	Units	29,9 U	/ml						D	2,639
Quantifizierungsgrenzen / Limits of quantific	U/ml	5	-	600							
Grenzwertbereich / Borderline range		U/ml	10	-	15						

OD B	OD Bereich / OD Range 405 nm, Standardserum / Standard serum																			
0,44		0,48	0,49	- 0,53	0,54	- 0,59	0,60	- 0,64	0,65	- 0,70	0,71	- 0,75	0,76	- 0,81	0,82	- 0,86	0,87	U/ml		Interpretation
	<	0,25		< 0.28		< 0.31		< 0.34		< 0.36	1	< 0.39		< 0.42		< 0.45	< 0,47		< 10.0	neg
0,25	i -	,	0,28		0,31	- 0,39	0,34	- 0,43	0,36	- 0,47	0,39	,	0,42	- 0,54	0,45	- 0,58	0,47 - 0,60	10,0		gw / borderline
	>	0,32		> 0,35		> 0,39		> 0,43		> 0,47		> 0,50		> 0,54		> 0,58	> 0,60			pos

OD Be	D Bereich / OD Range 405 nm, Standardserum / Standard serum																							
U/ml			0	,87	0,88	- 0,95	0,96	- 1,02	1,03	-	1,10	1,11	- 1,17	1,18	- 1,25	1,26	- 1,33	1,34	-	1,40	1,41	-	1,48	Interpretation
	<	10,0		< 0,47		< 0,49		< 0,53	3	<	0,58		< 0,62		< 0,66		< 0,70		<	0,74		<	0,78	neg
10,0	-	15,0	0,47	,	0,49	- 0,63	0,53	- 0,68			0,73	0,62	•	0,66	- 0,84	0,70	- 0,89		-	0,94	0,78	-		gw / borderline
	>	15,0		> 0,60		> 0,63		> 0,68	3	>	0,73		> 0,79		> 0,84		> 0,89		>	0,94		>	1,00	pos

Formeln für spezielle Auswertesysteme Special case formulas 0,693 x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off

OD = **0,539** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(4,266-In(2,773/(MV(Sample) x0,87/ MV(STD)+0,134)-1)/0,652)

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(4.266-ln(2.773/(Sample*0.870/S+0.134)-1)/0.652)



Gültigkeitsbereich / Validity Range

0.435<=S1<=1.479



If OD Sample < Parameter A

if Ti < (-0.134*(S1/0.870)) then Ti = (-0.134+0.001)*(S1/0.870)



If OD Sample > Parameter D

if Ti>(2.639*(S1/0.870)) then Ti=(2.639-0.001)*(S1/0.870)



If OD Negative control < Parameter A

if NC1<(-0.134*(S1/0.870)) then NCi=(-0.134+0.001)*(S1/0.870)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg