SERION ELISA classic ESR147G LEISHMANIA IgG EK0149

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot EK0149 IFU-Version 147-2

Verw. bis / Exp. 2021-10-31 !New! Prüfdatum /



15

10

Date of control

04.12.2019

Verwendete Reagenzien / Reagents used	Lot	Standard			Standard Kur	Standard Kurve / Standar					
Teststreifen / Antigen coated strips	ECK0295	Ref Werte / Ref	. Values		Gültigke	itsbereich	/ Valid	dity Range	Parameter	Α	0,003
Standardserum / Standard serum	ECK0284	OD	0,83		OD	0,42	-	1,41		В	1,016
Negativ Kontrolle / Negative control	ECK0283									С	4,501
Konjugat / Conjugate	KJK008+++	Units	47,0 U	/ml						D	2,432
Quantifizierungsgrenzen / Limits of quantific	U/ml	4	-	800							
Grenzwertbereich / Borderline range	U/ml	10	-	15							

OD E	OD Bereich / OD Range 405 nm, Standardserum / Standard serum																													
0,42	2 -	0,46	0,47	-	0,51	0,52	-	0,56	0,57	-	0,61	0,62	-	0,66	0,67	- 0,7	72	0,73	- 0),77	0,78	-	0,82		0,83	}	U/ml			Interpretation
		0.40			0.44		_	0.40			0.47			0.40		. 0.0	20						0.00	1		0.04			40.0	
	<	0,13		<	0,14		<	0,16		<	0,17		<	0,19		< 0,2	20		< (),22		<	0,23		<_	0,24		<	10,0	neg
0,13	3 -	0,18	0,14	-	0,20	0,16	-	0,22	0,17	-	0,24	0,19	-	0,26	0,20	- 0,2	28	0,22	- 0),31	0,23	-	0,33	0,24	-	0,34	10,0	-	15,0	gw / borderline
	>	0,18		>	0,20		>	0,22		>	0,24		>	0,26		> 0,2	28		> 0),31		>	0,33		>	0,34		>	15,0	pos

OD Bereic	DD Bereich / OD Range 405 nm, Standardserum / Standard serum														
U/ml		0,83	0,84 - 0,90	0,91 - 0,98	0,99 - 1,05	1,06 - 1,12	1,13 - 1,19	1,20 - 1,27	1,28 - 1,34 1,35 - 1,41	Interpretation					
<	10,0	< 0,24	< 0,25	< 0,27	< 0,29	< 0,31	< 0,34	< 0,36	< 0,38 < 0,40	neg					
10,0 -	15,0	, ,	0,25 - 0,36	0,27 - 0,39	0,29 - 0,42	-,-	0,34 - 0,48	0,36 - 0,51		gw / borderline					
>	15,0	> 0,34	> 0,36	> 0,39	> 0,42	> 0,45	> 0,48	> 0,51	> 0,54 > 0,57	pos					

Formeln für spezielle Auswertesysteme Special case formulas OD = 0,411 x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off

OD = **0,287** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(4,501-In(2,429/(MV(Sample) x0,83/ MV(STD)-0,003)-1)/1,016)

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(4.501-ln(2.429/(Sample*0.830/S-0.003)-1)/1.016)



Gültigkeitsbereich / Validity Range

0.415<=S1<=1.411



If OD Sample < Parameter A

if Ti < (0.003*(S1/0.830)) then Ti = (0.003+0.001)*(S1/0.830)



If OD Sample > Parameter D

if Ti>(2.432*(S1/0.830)) then Ti=(2.432-0.001)*(S1/0.830)



If OD Negative control < Parameter A

if NC1<(0.003*(S1/0.830)) then NCi=(0.003+0.001)*(S1/0.830)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg