SERION ELISA classic

ESR1372G

CHLAMYDIA TRACHOMATIS IgG

SLI.CM

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot SLI.CM IFU-Version 1372-6

Verw. bis / Exp. 2020-10

23.11.2018

Prüfdatum /

Date of control



15

10

Verwendete Reagenzien / Reagents used	Standard							Standard Kui	Standard Kurve / Standard curve			
Teststreifen / Antigen coated strips	SKI.BD	Ref Werte / Re	f. Value	es		Gültigke	itsbereich	/ Vali	dity Range	Parameter	Α	-0,007
Standardserum / Standard serum	SKI.EN	O	0,90			OD	0,45	-	1,53		В	0,957
Negativ Kontrolle / Negative control	SKI.DP										С	4,812
Konjugat / Conjugate	SKI.EK+++	Units	s 96,5	U/ml							D	2,044
Quantifizierungsgrenzen / Limits of quantifi	U/mI	4			250							
Grenzwertbereich / Borderline range	U/ml	10) .		15							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																					
0,45	-	0,50	0,51	- 0,55	0,56	- 0,61	0,62	- 0,67	0,68	- 0,72	0,73	0,78	0,79	- 0,83	0,84	- 0,89	0,90		U/ml		Interpretation
	<	0,08		< 0,09		< 0,10		< 0,11		< 0,12	<	0,13		< 0,14		< 0,15	<	0,16		< 10,	neg
0,08	-	0,12	0,09	- 0,14	0,10	- 0,15	0,11	- 0,16	0,12	- 0,18	0,13 -	0,19	0,14	- 0,21	0,15	- 0,22	0,16 -	0,23	10,0	- 15,	gw / borderline
	>	0,12		> 0,14		> 0,15		> 0,16		> 0,18	>	> 0,19		> 0,21		> 0,22	>	0,23		> 15,	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum												
U/ml		0,90	0,91 - 0,98	0,99 - 1,06	1,07 - 1,14	1,15 - 1,22	1,23 - 1,29	1,30 - 1,37	1,38 - 1,45 1,46 - 1,53	Interpretation		
<	10,0	< 0,16	< 0,17	< 0,18	< 0,20	< 0,21	< 0,22	< 0,24	< 0,25 < 0,27	neg		
,	15,0 15,0	0,16 - 0,23 > 0,23	0,17 - 0,24 > 0,24	0,18 - 0,26 > 0,26	0,20 - 0,28 > 0,28	0,21 - 0,30 > 0,30	0,22 - 0,32 > 0,32	0,24 - 0,34 > 0,34	0,25 - 0,36 0,27 - 0,38 > 0,36 > 0,38	gw / borderline		

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,261** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,182** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(4,812-ln(2,051/(MV(Sample) x0,9/ MV(STD)+0,007)-1)/0,957)

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(4.812-ln(2.051/(Sample*0.900/S+0.007)-1)/0.957)



Gültigkeitsbereich / Validity Range

0.450<=S1<=1.530



If OD Sample < Parameter A

if Ti < (-0.007*(S1/0.900)) then Ti = (-0.007+0.001)*(S1/0.900)



If OD Sample > Parameter D

if Ti>(2.044*(S1/0.900)) then Ti=(2.044-0.001)*(S1/0.900)



If OD Negative control < Parameter A

if NC1<(-0.007*(S1/0.900)) then NCi=(-0.007+0.001)*(S1/0.900)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg