## **Product Data Sheet - Recombinant Antigen**

Name of product	Echovirus Antigen, recombinant	
Order No.	BA135VS01	
Lot No.	Lot specific	
Source	E. coli derived recombinant antigen	
Product description	Proprietary Echovirus antigen	
Purification	By chromatography	
Presentation	The antigen is presented in buffer containing 39 mM Tris/HCl pH 8.0, 150 mM NaCl, 10 % glycerol.	
Storage	-65 °C or below. Avoid repeated freeze-thaw cycles.	
Date of manufacture	Lot specific	
Expiry	10 years from date of manufacture (when stored under recommended conditions)	
Manufacturer	Institut Virion\Serion GmbH, Germany	

 $For \ research \ or \ manufacturing \ use \ only. \ Not \ for \ use \ in \ diagnostic \ procedures.$ 

2018-01-15

## **Analysis:**

Test	Method	Specification
Protein concentration	Bradford assay (Bio-Rad) with BSA as reference standard	2 – 5 mg/mL
Identity / Purity	SDS-PAGE (reducing)	Major bands comparable to Reference Standard
Immunoreactivity	SERION ELISA classic Coxsackie-/Echovirus IgA, IgG & IgM	Individually pooled antigens shown to react with QC serum panel (multiple negative, borderline and positive sera) within defined reactivity range.