SERION ELISA classic

ESR121G

!New!

BORRELIA BURGDORFERI IgG

EK0078

Qualitätskontrollzertifikat / Quality Control Certificate

2021-07

Kitcharge / Lot EK0078 IFU-Ve

Verw. bis / Exp.

IFU-Version 121-20

OD =

06.08.2019 Prüfdatum /

Date of control



Verwendete Reagenzien / Reagents used	Lot	Standard						Standard Kur	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	ECK0115	Ref Werte / Ref	f. Values		Gültigke	itsbereich	/ Vali	dity Range	Parameter	Α	-0,021
Standardserum / Standard serum	ECK0124	OD	0,84		OD	0,42	-	1,43		В	0,848
Negativ Kontrolle / Negative control	ECK0122									С	3,422
Konjugat / Conjugate	SCK.CP+++	Units	20,7 U	/ml						D	2,038
Quantifizierungsgrenzen / Limits of quantific	U/ml	1	-	100							
Grenzwertbereich / Borderline range	U/ml	3	-	5							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																				
0,42	-	0,46	0,47	- 0,52	0,53	- 0,57	0,58	- 0,62	0,63	- 0,67	0,68	- 0,73	0,74	- 0,78	0,79	- 0,83	0,84	U/ml		Interpretation
	<	0,12		< 0,14		< 0,15		< 0,16		< 0,18		< 0,19		< 0,21		< 0,22	< 0,23	<	3,0	neg
0,12	-	0,18	0,14	- 0,20	0,15	- 0,22	0,16	- 0,24	0,18	- 0,26	0,19	- 0,28	0,21	- 0,31	0,22	- 0,33	0,23 - 0,34	3,0 -	5,0	gw / borderline
	>	0,18		> 0,20		> 0,22		> 0,24		> 0,26		> 0,28		> 0,31		> 0,33	> 0,34	>	5,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum											
U/ml	0,84	0,85 - 0,91	0,92 - 0,99	1,00 - 1,06	1,07 - 1,13	1,14 - 1,21	1,22 - 1,28	1,29 - 1,35 1,36 - 1,43	Interpretation		
< 3,0	< 0,23	< 0,24	< 0,26	< 0,28	< 0,30	< 0,32	< 0,34	< 0,36 < 0,38	neg		
3,0 - 5,0 > 5.0	0,23 - 0,34 > 0,34	0,24 - 0,36 > 0,36	0,26 - 0,39 > 0,39	0,28 - 0,42 > 0,42	0,30 - 0,45 > 0,45	0,32 - 0,48 > 0,48	0,34 - 0,51 > 0,51	0,36 - 0,54 0,38 - 0,57 > 0,54 > 0,57	gw / borderline		

Formeln für spezielle Auswertesysteme Special case formulas 0,409 x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off

OD = **0,275** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = \exp(3,422 - \ln(2,059/(MV(Sample) \ x0,84/\ MV(STD) + 0,021) - 1)/0,848)$

5 Institut Virion\Serion GmbH

3

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(3.422-ln(2.059/(Sample*0.840/S+0.021)-1)/0.848)



Gültigkeitsbereich / Validity Range

0.420<=S1<=1.428



If OD Sample < Parameter A

if Ti<(-0.021*(S1/0.840)) then Ti=(-0.021+0.001)*(S1/0.840)



If OD Sample > Parameter D

if Ti>(2.038*(S1/0.840)) then Ti=(2.038-0.001)*(S1/0.840)



If OD Negative control < Parameter A

if NC1<(-0.021*(S1/0.840)) then NCi=(-0.021+0.001)*(S1/0.840)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg