**SERION ELISA** classic

ESR139M

!New!

# **CAMPYLOBACTER JEJUNI IgM**

SAK.BL

## **Qualitätskontrollzertifikat / Quality Control Certificate**

Kitcharge / Lot SAK.BL IFU-V

IFU-Version 139-11

16.01.2019

Verw. bis / Exp. 2020-10

Prüfdatum /

Date of control



Verwendete Reagenzien / Reagents used	Standard							Standard Kur	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	SKI.EU	Ref Werte / Ref.	Values		Gültigke	itsbereich	/ Vali	dity Range	Parameter	Α	0,016
Standardserum / Standard serum	SMI.AM	OD	0,70		OD	0,35	-	1,19		В	0,935
Negativ Kontrolle / Negative control	SLI.EX									С	6,574
Konjugat / Conjugate	SII.AA+++	Units	130 U/m	ıl						D	4,073
Quantifizierungsgrenzen / Limits of quantific	U/ml	20	-	2000							
Grenzwertbereich / Borderline range	U/ml	40	-	60							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																			
0,35	-	0,38	0,39	- 0,43	0,44	- 0,47	0,48	- 0,52	0,53	- 0,56	0,57	- 0,60	0,61	- 0,65	0,66	- 0,69	0,70	U/ml	Interpretation
	<	0,14		< 0,16		< 0,18		< 0,19		< 0,21		< 0,23		< 0,24		< 0,26	< 0,27	< 40,0	neg
0,14	-	0,20	0,16	- 0,22	0,18	- 0,25	0,19	- 0,27	0,21	- 0,29	0,23	- 0,32	0,24	- 0,34	0,26	- 0,37	0,27 - 0,38	40,0 - 60,0	gw / borderline
	>	0,20		> 0,22		> 0,25		> 0,27		> 0,29		> 0,32		> 0,34		> 0,37	> 0,38	> 60,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum										
U/ml	0,70	0,71 - 0,76	0,77 - 0,82	0,83 - 0,88	0,89 - 0,95	0,96 - 1,01	1,02 - 1,07	1,08 - 1,13 1,14 - 1,19	Interpretation	
< 40,0	< 0,27	< 0,28	< 0,31	< 0,33	< 0,35	< 0,38	< 0,40	< 0,43 < 0,45	neg	
40,0 - 60,0 > 60,0	0,27 - 0,38 > 0,38	0,28 - 0,40 > 0,40	0,31 - 0,43 > 0,43	0,33 - 0,47 > 0,47	0,35 - 0,50 > 0,50	0,38 - 0,53 > 0,53	0,40 - 0,57 > 0,57	0,43 - 0,60 0,45 - 0,63 > 0.60 > 0.63	gw / borderline	

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,542** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,389** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(6,574-In(4,057/(MV(Sample) x0,7/ MV(STD)-0,016)-1)/0,935)

60 Institut Virion\Serion GmbH

40

Friedrich-Bergius-Ring 19 D-97076 Würzburg

# Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

#### 4PS- Formel / 4PS-formula

exp(6.574-ln(4.057/(Sample\*0.700/S-0.016)-1)/0.935)



#### Gültigkeitsbereich / Validity Range

0.350<=S1<=1.190



#### If OD Sample < Parameter A

if Ti < (0.016\*(S1/0.700)) then Ti = (0.016+0.001)\*(S1/0.700)



## If OD Sample > Parameter D

if Ti>(4.073\*(S1/0.700)) then Ti=(4.073-0.001)\*(S1/0.700)



### If OD Negative control < Parameter A

if NC1<(0.016\*(S1/0.700)) then NCi=(0.016+0.001)\*(S1/0.700)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg