SERION ELISA classic ESR125G LEPTOSPIRA IgG EK0075

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot EK0075 IFU-Version 125-16

Verw. bis / Exp. 2021-06 !New! Prüfdatum /



15

10

Date of control

31.07.2019

Verwendete Reagenzien / Reagents used	Standard							Standard Kur	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	Ref Werte / Ref. Values			Gültigkeitsbereich / Validity Range				Parameter	Α	0,013	
Standardserum / Standard serum	ECK0098	OD	0,78		OD	0,39	-	1,33		В	1,115
Negativ Kontrolle / Negative control	ECK0097									С	2,680
Konjugat / Conjugate	SDK.BI++	Units	44,6 U	/ml						D	1,001
Quantifizierungsgrenzen / Limits of quantific	U/ml	2	-	100							
Grenzwertbereich / Borderline range	U/ml	10	-	15			<u> </u>				

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																				
0,39	-	0,43	0,44	- 0,48	0,49	- 0,53	0,54	- 0,58	0,59	- 0,62	0,63	- 0,67	0,68	- 0,72	0,73	- 0,77	0,78	U/ml		Interpretation
-	<	0,21		< 0,23		< 0,26		< 0,28		< 0,31		< 0,33		< 0,36		< 0,38	< 0,40	<	10,0	neg
0,21	-	0,27	0,23	- 0,30	0,26	- 0,33	0,28	- 0,36	0,31	- 0,40	0,33	- 0,43	0,36	- 0,46	0,38	- 0,49	0,40 - 0,51	10,0 -	15,0	gw / borderline
	>	0,27		> 0,30		> 0,33		> 0,36		> 0,40		> 0,43		> 0,46		> 0,49	> 0,51	>	15,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum											
U/ml	0,78	0,79 - 0,85	0,86 - 0,92	0,93 - 0,98	0,99 - 1,05	1,06 - 1,12	1,13 - 1,19	1,20 - 1,26 1,27 - 1,33	Interpretation		
< 10,0	< 0,40	< 0,42	< 0,46	< 0,49	< 0,53	< 0,56	< 0,60	< 0,63 < 0,67	neg		
10,0 - 15,0 > 15,0	, ,	0,42 - 0,54 > 0,54	0,46 - 0,58 > 0,58	0,49 - 0,62 > 0,62	0,53 - 0,67 > 0,67	0,56 - 0,71 > 0,71	0,60 - 0,76 > 0,76	0,63 - 0,80	gw / borderline		

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,660** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,519** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(2,68-In(0,988/(MV(Sample) x0,78/ MV(STD)-0,013)-1)/1,115)

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(2.680-ln(0.988/(Sample*0.780/S-0.013)-1)/1.115)



Gültigkeitsbereich / Validity Range

0.390<=S1<=1.326



If OD Sample < Parameter A

if Ti<(0.013*(S1/0.780)) then Ti=(0.013+0.001)*(S1/0.780)



If OD Sample > Parameter D

if Ti>(1.001*(S1/0.780)) then Ti=(1.001-0.001)*(S1/0.780)



If OD Negative control < Parameter A

if NC1<(0.013*(S1/0.780)) then NCi=(0.013+0.001)*(S1/0.780)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg