SERION ELISA classic ESR116G BRUCELLA IgG EK0125

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot EK0125 IFU-Version 116-14

Verw. bis / Exp. 2021-10-31 !New! Prüfdatum /



30

20

Date of control

30.10.2019

Verwendete Reagenzien / Reagents used	Lot	Standard			Standard Kur	ve / Stand	lard curve				
Teststreifen / Antigen coated strips	ECK0263	Ref Werte / Ref	. Values		Gültigke	tsbereich	/ Valid	lity Range	Parameter	Α	-0,030
Standardserum / Standard serum	ECK0259	OD	0,98		OD	0,49	-	1,67		В	0,887
Negativ Kontrolle / Negative control	ECK0258									С	3,858
Konjugat / Conjugate	KJK009++	Units	64,3 U	/ml						D	1,750
Quantifizierungsgrenzen / Limits of quantific	U/ml	5	-	500							
Grenzwertbereich / Borderline range	U/ml	20	-	30					•		

OD B	D Bereich / OD Range 405 nm, Standardserum / Standard serum																										
0,49) -	0,54	0,55	- (0,60	0,61	- (0,66	0,67	- 0,7	3 0,	74 -	- 0,79	0,80	- 0,85	0,86	- 0,91	0,92	-	0,97	0,98			U/ml			Interpretation
		0.28		- (0.32			0.35		- 00	<u> </u>		< 0.42	I	< 0.45		< 0.49		_	0.52		_	0.54		_	20.0	noa
	_	-, -			, -			.,		< 0,3			-,		-, -		-, -			-,-		_	0,54		_	20,0	J
0,28	-	0,36	0,32	- (0,40	0,35	- (0,44	0,39	- 0,4	9 0,4	12 -	- 0,53	0,45	- 0,57	0,49	- 0,61	0,52	-	0,66	0,54	-	0,68	20,0	-	30,0	gw / borderline
	>	0,36		> (0,40		> (0,44		> 0,4	9	;	> 0,53		> 0,57		> 0,61		>	0,66		>	0,68		>	30,0	pos

OD Be	D Bereich / OD Range 405 nm, Standardserum / Standard serum																										
U/ml			(),98		0,99	- 1,07	1,08	-	1,15	1,16	-	1,24	1,25	- 1,32	1,33	- 1,41	1	1,42	- 1,49	1,50	-	1,58	1,59	-	1,67	Interpretation
	<	20,0		< 0,5	4		< 0,57		<	0,61		<	0,66		< 0,71		< 0,76			< 0,80		<	0,85		<	0,90	neg
20,0	-	30,0	0,54	- 0,6		0,57	- 0,71	0,61	-	0,77	0,66	-	•	0,71	- ,	0,76	- 0,95		0,80	- 1,01	0,85	-	1,07	0,90	-		gw / borderline
	>	30,0		> 0,6	8		> 0,71		>	0,77		>	0,83		> 0,89		> 0,95			> 1,01		>	1,07		>	1,13	pos

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,696** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,546** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

Concentration= exp(3,858-In(1,78/(MV(Sample) x0,98/ MV(STD)+0,03)-1)/0,887)

Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(3.858-ln(1.780/(Sample*0.980/S+0.030)-1)/0.887)



Gültigkeitsbereich / Validity Range

0.490<=S1<=1.666



If OD Sample < Parameter A

if Ti < (-0.030*(S1/0.980)) then Ti = (-0.030+0.001)*(S1/0.980)



If OD Sample > Parameter D

if Ti>(1.750*(S1/0.980)) then Ti=(1.750-0.001)*(S1/0.980)



If OD Negative control < Parameter A

if NC1<(-0.030*(S1/0.980)) then NCi=(-0.030+0.001)*(S1/0.980)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg