**SERION ELISA classic** 

**ESR1051M** 

**HERPES SIMPLEX VIRUS 1 IgM** 

## SCK.DD

#### **Qualitätskontrollzertifikat / Quality Control Certificate**

Kitcharge / Lot SCK.DD IFU-Version 105-23

Verw. bis / Exp. 02-2021 !New!

19.03.2019

Prüfdatum /

Date of control



30

20

Verwendete Reagenzien / Reagents used	Lot	Standard		Standard Kur	Standard Kurve / Standard curve						
Teststreifen / Antigen coated strips	Ref Werte / Re	f. Values		Gültigke	itsbereich	/ Vali	dity Range	Parameter	Α	0,038	
Standardserum / Standard serum	SBK.GA	OE	0,88		OD	0,44	-	1,50		В	0,980
Negativ Kontrolle / Negative control	SBK.FZ									С	4,049
Konjugat / Conjugate	SCK.AA++	Units	3 17,4 U	/ml						D	3,595
Quantifizierungsgrenzen / Limits of quantific	U/ml	10	-	500							
Grenzwertbereich / Borderline range	U/ml	20	-	30	<u> </u>						

OD B	DD Bereich / OD Range 405 nm, Standardserum / Standard serum																			
0,44		0,49	0,50	- 0,54	0,55	- 0,60	0,61	- 0,65	0,66	- 0,71	0,72	- 0,76	0,77	- 0,82	0,83	- 0,87	0,88	U/ml		Interpretation
		0,51		< 0.57	I	< 0.63		< 0.69		< 0.75	1	< 0,81	1	< 0.87		< 0.93	< 0.9	7	- 20 O	neg
0.54	_		0.57		0.00		0.00		0.75		0.04		0.07	,	0.00	-,	,			
0,51	-	•	0,57	- 0,75	0,63	- 0,83	0,69	•	0,75	- 0,98	0,81	,	0,87	- 1,14	0,93	- 1,22	0,97 - 1,2	,		gw / borderline
	>	0,67		> 0,75		> 0,83		> 0,91		> 0,98		> 1,06		> 1,14		> 1,22	> 1,2	7	> 30,0	pos

OD Be	D Bereich / OD Range 405 nm, Standardserum / Standard serum																					
U/ml			0,88		0,89	- 0,96	0,97	- 1,0	3 1,	04	- 1,11	1,12	- 1,19	1,20	- 1,27	1,28	- 1,34	1,35	- 1,42	1,43	- 1,50	Interpretation
	<	20,0	< 0,	97		< 1,02		< 1,1	)	<	< 1,19		< 1,27		< 1,36		< 1,44		< 1,53		< 1,61	neg
20,0	-	,	•	27	1,02	- 1,33	1,10	- 1,4	,	19	- 1,56	1,27	- 1,67	1,36	- 1,78	1,44	- 1,89	1,53	- 2,00	1,61		gw / borderline
	>	30,0	> 1,	27		> 1,33		> 1,4	4		> 1,56		> 1,67		> 1,78		> 1,89		> 2,00		> 2,1	pos

Formeln für spezielle Auswertesysteme Special case formulas OD = 1,443 x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = 1,105 x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = \exp(4,049 - \ln(3,557/(\text{MV(Sample}) \ x0,88/ \ \text{MV(STD)} - 0,038) - 1)/0,98)$ 

Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg

# Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

#### 4PS- Formel / 4PS-formula

exp(4.049-ln(3.557/(Sample\*0.880/S-0.038)-1)/0.980)



#### Gültigkeitsbereich / Validity Range

0.440<=S1<=1.496



#### If OD Sample < Parameter A

if Ti<(0.038\*(S1/0.880)) then Ti=(0.038+0.001)\*(S1/0.880)



#### If OD Sample > Parameter D

if Ti>(3.595\*(S1/0.880)) then Ti=(3.595-0.001)\*(S1/0.880)



### If OD Negative control < Parameter A

if NC1<(0.038\*(S1/0.880)) then NCi=(0.038+0.001)\*(S1/0.880)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg