SERION ELISA classic ESR125G LEPTOSPIRA IgG SMH.AX

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot SMH.AX IFU-Version 125-15

Verw. bis / Exp. 2019-11

06.12.2017

Prüfdatum /

Date of control



15

10

Verwendete Reagenzien / Reagents used	Standard		Standard Kui	Standard Kurve / Standard curve							
Teststreifen / Antigen coated strips	Ref Werte / Re	f. Value	es	Gültigk	eitsbereich	/ Vali	idity Range	Parameter	Α	0,019	
Standardserum / Standard serum	SLH.CO	OE	0,92		OD	0,46	-	1,56		В	1,133
Negativ Kontrolle / Negative control	SLH.CN									С	2,532
Konjugat / Conjugate	SLH.CD++	Units	41,1	U/ml						D	1,156
Quantifizierungsgrenzen / Limits of quantifi	U/mI	2	-	100							
Grenzwertbereich / Borderline range	U/ml	10) -	15							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																						
0,46	-	0,51	0,52	- 0,57	0,58	- 0,62	0,63	- 0,68	0,69	- 0,74	0,75	- 0,80	0,81	- 0,85	0,86	- 0,91	0,92		U/ml			Interpretation
	<	0,27		< 0,30		< 0,33		< 0,36		< 0,40		< 0,43		< 0,46		< 0,49	<	0,51		<	10,0	neg
0,27	-	0,34	0,30	- 0,38	0,33	- 0,42	0,36	- 0,46	0,40	- 0,50	0,43	- 0,54	0,46	- 0,58	0,49	- 0,62	0,51 -	0,64	10,0	-	15,0	gw / borderline
	>	0,34	:	> 0,38		> 0,42		> 0,46		> 0,50		> 0,54		> 0,58		> 0,62	>	0,64		>	15,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum												
U/ml	0,92	0,93 - 1,00	1,01 - 1,08	1,09 - 1,16	1,17 - 1,24	1,25 - 1,32	1,33 - 1,40	1,41 - 1,48 1,49 - 1,56	Interpretation			
< 10,0	< 0,51	< 0,54	< 0,58	< 0,62	< 0,67	< 0,71	< 0,76	< 0,80 < 0,85	neg			
10,0 - 15,0	, ,	0,54 - 0,67	, ,	0,62 - 0,78	- 1 1 -	0,71 - 0,90	0,76 - 0,95		gw / borderline			
> 15,0	, ,	> 0,67	> 0,73	> 0,78	> 0,84	> 0,90	> 0,95	, , , , , , , , , , , , , , , , , , , ,	pos			

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,700** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,559** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = exp(2,532-In(1,137/(MV(Sample)\ x0,92/\ MV(STD)-0,019)-1)/1,133)$

Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(2.532-ln(1.137/(Sample*0.920/S-0.019)-1)/1.133)



Gültigkeitsbereich / Validity Range

0.460<=S1<=1.564



If OD Sample < Parameter A

if Ti<(0.019*(S1/0.920)) then Ti=(0.019+0.001)*(S1/0.920)



If OD Sample > Parameter D

if Ti>(1.156*(S1/0.920)) then Ti=(1.156-0.001)*(S1/0.920)



If OD Negative control < Parameter A

if NC1<(0.019*(S1/0.920)) then NCi=(0.019+0.001)*(S1/0.920)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg