SERION ELISA classic ESR127A

MYCOPLASMA PNEUMONIAE IgA

SEI.CG

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot SEI.CG IFU-Version 127-14

Verw. bis / Exp. 2020-01

28.05.2018

Prüfdatum /

Date of control



14

10

Verwendete Reagenzien / Reagents used	Standard							Standard Kur	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	SKH.CW	Ref Werte / Re	f. Values		Gültigke	itsbereich	/ Valid	dity Range	Parameter	Α	0,034
Standardserum / Standard serum	SAI.CU	OD	0,88		OD	0,44	-	1,50		В	1,285
Negativ Kontrolle / Negative control	SAI.CT									С	4,149
Konjugat / Conjugate	SDI.AA++	Units	21,9 U	/ml						D	4,189
Quantifizierungsgrenzen / Limits of quantific	U/ml	2	-	150							
Grenzwertbereich / Borderline range	U/ml	10	-	14							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																					
0,44		0,49	0,50	- 0,54	0,55	- 0,60	0,61	- 0,65	0,66	- 0,71	0,72 -	0,76	0,77	- 0,82	0,83	- 0,87	0,88	U	J/ml		Interpretation
	<	0,20		< 0,23		< 0,25		< 0,28		< 0,30	<	0,33		< 0,35		< 0,38	< 0,	39		< 10,0	neg
0,20	-	0,29	0,23	- 0,33	0,25	- 0,36	0,28	- 0,40	0,30	- 0,43	0,33 -	0,47	0,35	- 0,50	0,38	- 0,54	0,39 - 0,	56	10,0	- 14,0	gw / borderline
	>	0,29		> 0,33		> 0,36		> 0,40		> 0,43	>	0,47		> 0,50		> 0,54	> 0,	56		> 14,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum											
U/ml	0,88	0,89 - 0,96	0,97 - 1,03	1,04 - 1,11	1,12 - 1,19	1,20 - 1,27	1,28 - 1,34	1,35 - 1,42 1,43 - 1,50	Interpretation		
< 10,0	< 0,39	< 0,41	< 0,44	< 0,48	< 0,51	< 0,55	< 0,58	< 0,61 < 0,65	neg		
10,0 - 14,0 > 14,0	, ,	0,41 - 0,59 > 0,59	0,44 - 0,64 > 0,64	0,48 - 0,69 > 0,69	0,51 - 0,73 > 0,73	0,55 - 0,78 > 0,78	0,58 - 0,83 > 0,83		gw / borderline pos		

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,632** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,441** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = exp(4,149-ln(4,155/(MV(Sample)\ x0,88/\ MV(STD)-0,034)-1)/1,285)$

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(4.149-ln(4.155/(Sample*0.880/S-0.034)-1)/1.285)



Gültigkeitsbereich / Validity Range

0.440<=S1<=1.496



If OD Sample < Parameter A

if Ti<(0.034*(S1/0.880)) then Ti=(0.034+0.001)*(S1/0.880)



If OD Sample > Parameter D

if Ti>(4.189*(S1/0.880)) then Ti=(4.189-0.001)*(S1/0.880)



If OD Negative control < Parameter A

if NC1<(0.034*(S1/0.880)) then NCi=(0.034+0.001)*(S1/0.880)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg