SERION ELISA classic

ESR142M

FRANCISELLA TULARENSIS IgM

EK0147

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot EK0147 IFU-Version 142-5

Verw. bis / Exp. 2021-10-31 !New!

02.12.2019

Prüfdatum /

Date of control



15

10

Verwendete Reagenzien / Reagents used	Standard							Standard Kur	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	Ref Werte / Ref. Values			Gültigkeitsbereich / Validity Range				Parameter	Α	0,026	
Standardserum / Standard serum	OD	0,89		OD	0,45	-	1,51		В	1,216	
Negativ Kontrolle / Negative control	ECK0275									С	4,574
Konjugat / Conjugate	KJK010++	Units	34,4 L	l/ml						D	3,937
Quantifizierungsgrenzen / Limits of quantific	U/ml	4	-	400							
Grenzwertbereich / Borderline range	U/ml	10	-	15							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																				
0,45	5 -	0,49	0,50	- 0,55	0,56	- 0,60	0,61	- 0,66	0,67	- 0,71	0,72	- 0,77	0,78	- 0,82	0,83	- 0,88	0,89	U/ml		Interpretation
	<	0,14		< 0,15		< 0,17		< 0,19		< 0,20		< 0,22		< 0,23		< 0,25	< 0,26	< '	10,0	neg
0,14	- 4	0,20	0,15	- 0,23	0,17	- 0,25	0,19	- 0,28	0,20	- 0,30	0,22	- 0,33	0,23	- 0,35	0,25	- 0,38	0,26 - 0,39	10,0 -	15,0	gw / borderline
	>	0,20		> 0,23		> 0,25		> 0,28		> 0,30		> 0,33		> 0,35		> 0,38	> 0,39	> 1	15,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum											
U/ml	0,89	0,90 - 0,97	0,98 - 1,05	1,06 - 1,12	1,13 - 1,20	1,21 - 1,28	1,29 - 1,36	1,37 - 1,44 1,45 - 1,51	Interpretation		
< 10,0	< 0,26	< 0,27	< 0,30	< 0,32	< 0,34	< 0,36	< 0,39	< 0,41 < 0,43	neg		
10,0 - 15,0 > 15,0	, ,	0,27 - 0,41 > 0,41	0,30 - 0,44 > 0,44	0,32 - 0,48 > 0.48	0,34 - 0,51 > 0,51	0,36 - 0,55 > 0,55	0,39 - 0,58 > 0,58	0,41 - 0,61 0,43 - 0,65 > 0,61 > 0,65	gw / borderline		

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,441** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off

OD = **0,290** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = \exp(4,574 - \ln(3,911/(MV(Sample)\ x0,89/\ MV(STD) - 0,026) - 1)/1,216)$

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(4.574-ln(3.911/(Sample*0.890/S-0.026)-1)/1.216)



Gültigkeitsbereich / Validity Range

0.445<=S1<=1.513



If OD Sample < Parameter A

if Ti < (0.026*(S1/0.890)) then Ti = (0.026+0.001)*(S1/0.890)



If OD Sample > Parameter D

if Ti>(3.937*(S1/0.890)) then Ti=(3.937-0.001)*(S1/0.890)



If OD Negative control < Parameter A

if NC1<(0.026*(S1/0.890)) then NCi=(0.026+0.001)*(S1/0.890)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg