SERION ELISA classic

ESR103M

MUMPS VIRUS / PAROTITIS / OREILLONS IgM

SBK.BT

Qualitätskontrollzertifikat / Quality Control Certificate

Kitcharge / Lot SBK.BT IFU-Version 103-13

Verw. bis / Exp. 2021-01 !New!

13.02.2019

Prüfdatum /

Date of control



15

10

Verwendete Reagenzien / Reagents used	Standard							Standard Kur	Standard Kurve / Standard curve		
Teststreifen / Antigen coated strips	SAK.CV	Ref Werte / Ref	. Values		Gültigke	itsbereich	/ Valid	dity Range	Parameter	Α	0,080
Standardserum / Standard serum	SAK.EW	OD	1,02		OD	0,51	-	1,73		В	1,198
Negativ Kontrolle / Negative control	SAK.EV									С	4,650
Konjugat / Conjugate	SKI.EL++	Units	41,1 U	l/ml						D	3,894
Quantifizierungsgrenzen / Limits of quantific	U/ml	5	-	600							
Grenzwertbereich / Borderline range	U/ml	10	-	15							

OD Bereich / OD Range 405 nm, Standardserum / Standard serum																				
0,51	-	0,56	0,57	- 0,63	0,64	- 0,69	0,70	- 0,76	0,77	- 0,82	0,83	- 0,88	0,89	- 0,95	0,96	- 1,01	1,02	U/ml		Interpretation
	<	0,16		< 0,18		< 0,20		< 0,21		< 0,23		< 0,25		< 0,27		< 0,29	< 0,30	<	10,0	neg
0,16	-	0,22	0,18	- 0,25	0,20	- 0,27	0,21	- 0,30	0,23	- 0,33	0,25	- 0,35	0,27	- 0,38	0,29	- 0,40	0,30 - 0,42	10,0 -	15,0	gw / borderline
	>	0,22		> 0,25		> 0,27		> 0,30		> 0,33		> 0,35		> 0,38		> 0,40	> 0,42	>	15,0	pos

OD Bereich / OD Range 405 nm, Standardserum / Standard serum											
U/ml	1,02	1,03 - 1,11	1,12 - 1,20	1,21 - 1,29	1,30 - 1,38	1,39 - 1,47	1,48 - 1,56	1,57 - 1,64 1,65 - 1,73	Interpretation		
< 10,0	< 0,30	< 0,31	< 0,34	< 0,37	< 0,39	< 0,42	< 0,45	< 0,47 < 0,50	neg		
10,0 - 15,0 > 15,0	0,30 - 0,42 > 0.42	0,31 - 0,44 > 0,44	0,34 - 0,48 > 0,48	0,37 - 0,51 > 0,51	0,39 - 0,55 > 0,55	0,42 - 0,59 > 0,59	0,45 - 0,62 > 0,62	0,47 - 0,66	gw / borderline		

Formeln für spezielle Auswertesysteme Special case formulas OD = **0,411** x MV(STD) entspricht oberem cut-off/corresponds to upper cut-off
OD = **0,290** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

OD = $0.290 \times MV(STD)$ entspricht unterem cut-off/corresponds to lower cut-o Concentration= $exp(4.65-ln(3.814/(MV(Sample) \times 1.02/ MV(STD)-0.08)-1)/1,198)$ Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

4PS- Formel / 4PS-formula

exp(4.650-ln(3.814/(Sample*1.020/S-0.080)-1)/1.198)



Gültigkeitsbereich / Validity Range

0.510<=S1<=1.734



If OD Sample < Parameter A

if Ti < (0.080*(S1/1.020)) then Ti = (0.080+0.001)*(S1/1.020)



If OD Sample > Parameter D

if Ti>(3.894*(S1/1.020)) then Ti=(3.894-0.001)*(S1/1.020)



If OD Negative control < Parameter A

if NC1<(0.080*(S1/1.020)) then NCi=(0.080+0.001)*(S1/1.020)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg