# SERION ELISA classic ESR126G

## **PARAINFLUENZA VIRUS IgG**

# **SKI.EV**

### **Qualitätskontrollzertifikat / Quality Control Certificate**

Kitcharge / Lot SKI.EV IFU-Version 126-14

Verw. bis / Exp. 2020-10

08.11.2018

Prüfdatum /

Date of control



15

10

| Verwendete Reagenzien / Reagents used         | Standard                |                |       |                                     |    |      |   | Standard Kur | Standard Kurve / Standard curve |        |       |
|---|-------------------------|----------------|-------|-------------------------------------|----|------|---|--------------|---------------------------------|--------|-------|
| Teststreifen / Antigen coated strips          | Ref Werte / Ref. Values |                |       | Gültigkeitsbereich / Validity Range |    |      |   | Parameter    | Α                               | -0,012 |       |
| Standardserum / Standard serum                | SKI.DE                  | OD <b>0,78</b> |       |                                     | OD | 0,39 | - | 1,33         |                                 | В      | 0,874 |
| Negativ Kontrolle / Negative control          | SKI.DD                  |                |       |                                     |    |      |   |              |                                 | С      | 4,864 |
| Konjugat / Conjugate                          | SDI.EB+++               | Units          | 102 U | ml                                  |    |      |   |              |                                 | D      | 1,753 |
| Quantifizierungsgrenzen / Limits of quantific | U/ml                    | 5              | -     | 1000                                |    |      |   |              |                                 |        |       |
| Grenzwertbereich / Borderline range           | U/ml                    | 10             | -     | 15                                  |    |      |   |              |                                 |        |       |

| OD Bereich / OD Range 405 nm, Standardserum / Standard serum |   |      |      |                   |    |      |        |      |        |      |        |      |        |      |        |      |        |             |             |                 |
|--|---|------|------|-------------------|----|------|--------|------|--------|------|--------|------|--------|------|--------|------|--------|-------------|-------------|-----------------|
| 0,39   | - | 0,43 | 0,44 | - 0,4             | 18 | 0,49 | - 0,53 | 0,54 | - 0,58 | 0,59 | - 0,62 | 0,63 | - 0,67 | 0,68 | - 0,72 | 0,73 | - 0,77 | 0,78        | U/ml        | Interpretation  |
|  | < | 0,08 |      | < 0,0             | )9 |      | < 0,10 |      | < 0,11 |      | < 0,12 |      | < 0,13 |      | < 0,14 |      | < 0,15 | < 0,16      | < 10,0      | neg             |
| 0,08   | - | 0,12 | 0,09 | - 0, <sup>-</sup> | 13 | 0,10 | - 0,14 | 0,11 | - 0,16 | 0,12 | - 0,17 | 0,13 | - 0,18 | 0,14 | - 0,20 | 0,15 | - 0,21 | 0,16 - 0,22 | 10,0 - 15,0 | gw / borderline |
|  | > | 0,12 |      | > 0,              | 13 |      | > 0,14 |      | > 0,16 |      | > 0,17 |      | > 0,18 |      | > 0,20 |      | > 0,21 | > 0,22      | > 15,0      | pos             |

| OD Bereich / OD Range 405 nm, Standardserum / Standard serum |        |                       |                       |                       |                       |                       |                       |   |                 |  |  |
|--|--------|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|---|-----------------|--|--|
| U/ml   | 0,78   | 0,79 - 0,85           | 0,86 - 0,92           | 0,93 - 0,98           | 0,99 - 1,05           | 1,06 - 1,12           | 1,13 - 1,19           | 1,20 - 1,26 1,27 - 1,33                     | Interpretation  |  |  |
| < 10,0   | < 0,16 | < 0,17                | < 0,18                | < 0,20                | < 0,21                | < 0,22                | < 0,24                | < 0,25 < 0,27                               | neg             |  |  |
| 10,0 - 15,0<br>> 15,0  | , ,    | 0,17 - 0,23<br>> 0,23 | 0,18 - 0,25<br>> 0,25 | 0,20 - 0,27<br>> 0,27 | 0,21 - 0,29<br>> 0,29 | 0,22 - 0,31<br>> 0,31 | 0,24 - 0,33<br>> 0,33 | 0,25 - 0,35   0,27 - 0,37   > 0,35   > 0,37 | gw / borderline |  |  |

Formeln für spezielle Auswertesysteme Special case formulas  $\mathsf{OD} = \qquad \textbf{0,283} \qquad \mathsf{x} \; \mathsf{MV}(\mathsf{STD}) \; \mathsf{entspricht} \; \mathsf{oberem} \; \mathsf{cut\text{-}off/corresponds} \; \mathsf{to} \; \mathsf{upper} \; \mathsf{cut\text{-}off}$ 

OD = **0,203** x MV(STD) entspricht unterem cut-off/corresponds to lower cut-off

 $Concentration = exp(4,864-In(1,765/(MV(Sample)\ x0,78/\ MV(STD)+0,012)-1)/0,874)$ 

Institut Virion\Serion GmbH

Friedrich-Bergius-Ring 19 D-97076 Würzburg

# Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for Revelation™ DSX / DS-Matrix ™

#### 4PS- Formel / 4PS-formula

exp(4.864-ln(1.765/(Sample\*0.780/S+0.012)-1)/0.874)



#### Gültigkeitsbereich / Validity Range

0.390<=S1<=1.326



#### If OD Sample < Parameter A

if Ti < (-0.012\*(S1/0.780)) then Ti = (-0.012+0.001)\*(S1/0.780)



# If OD Sample > Parameter D

if Ti>(1.753\*(S1/0.780)) then Ti=(1.753-0.001)\*(S1/0.780)



#### If OD Negative control < Parameter A

if NC1<(-0.012\*(S1/0.780)) then NCi=(-0.012+0.001)\*(S1/0.780)



Institut Virion\Serion GmbH Friedrich-Bergius-Ring 19 D-97076 Würzburg