## virion\serion

## Qualitätskontrollzertifikat

## **Quality Control Certificate**

## SERION ELISA *classic* ADENOVIRUS IgM

| BestNr. / Order No.         | ESR128M            |
|-----------------------------|--------------------|
| Chargen-Nr. / Lot No.       | SBK.CZ             |
| Verfallsdatum / Expiry Date | 2020-02            |
| IFU-Version / IFU-Version   | <b>NEW!</b> 128-15 |

| Kontrollseren /<br>Control sera        | Chargen-Nr. /<br>Lot No | Gemessener Wert /<br>Measured value<br>OD 405 nm | Gültigkeitsbereich /<br>Validity range |
|--|-------------------------|--|--|
| Positivkontrolle /<br>Positive control | SAK.DP                  | 0,97   | > 0,49                                 |
| Cut off-Serum /<br>Cut off serum       | SAK.DQ                  | 0,25   | 0,13 - 0,50                            |

| Negativkontrolle / Negative Control | ADENOVIRUS IgM |  |
|-------------------------------------|----------------|--|
| Chargen-Nr. / Lot No.               | SAK.DO         |  |
| Teststreifen / Strips               | ADENOVIRUS IgM |  |
| Chargen-Nr. / Lot No.               | SBI.ER         |  |
| Konjugat / Conjugate                | IgM ++         |  |
| Chargen-Nr. / Lot No.               | SKI.EL         |  |

04.03.2019

Prüfdatum / Date of control

Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100

E-Mail: info@virion-serion.de Internet: www.virion-serion.de