#	Function	SET_SERIAL	SET_PARALLEL	OUT_ENABLE1	OUT_ENABLE2	SENSE1_ON	SENSE2_ON	Serial/Parallel (LED_SP)	Out2 (LED_O2)	Sense2 (LED_S2)	Rprog2 (LED_PROG2)	CC2 (LED_CC2)	CV2 (LED_CV2)	Out1 (LED_O1)	Sense1 (LED_S1)	Rprog1 (LED_PROG1)	CC1 (LED_CC1)	CV1 (LED_CV1)	Standy-by	LED_PWR 15	K_S2 14	LED_O2 13	LED_S2 12	LED_PROG2 11	LED_CC2 10	K_PAR 9	K_SER 8	LED_CV1 7	LED_CC1 6	LED_S1 5	LED_PROG1 4	LED_SP 3	LED_O1 2	K_S1 1	LED_CV2 0	TLC5925 code
1	Serial (no remote sensing)	1	0	1	1	0	0	●R												0	0	0	0	0	X	0	1	X	X	0	0	1	0	0	X	0108h
2	Parallel (no remote sensing)	0	1	1	1	0	0	●R												0	0	0	0	0	X	1	0	X	X	0	0	1	0	0	X	0208h
3	Parallel (with remote sensing)	0	1	1	1	1	0	●R								●Y				0	1	0	1	0	X	1	0	X	X	1	0	1	0	1	X	522Ah
4	Ch#2 only (no remote sensing)	0	0	0	1	0	0		●G											0	0	1	0	X	X	0	0	0	0	0	0	0	0	0	0	2000h
5	Ch#2 only (with remote sensing)	0	0	0	1	0	1		●G	●Y										0	1	1	0	X	X	0	0	0	0	0	0	0	0	0	0	6000h
6	Ch#1 only (no remote sensing)	0	0	1	0	0	0							●G						0	0	0	0	0	0	0	0	X	X	0	X	0	1	0	X	0004h
7	Ch#1 only (with remote sensing)	0	0	1	0	1	0							●G	●Y					0	0	0	0	0	0	0	0	X	X	1	X	0	1	1	X	0026h
8	Both ch (no remote sensing)	0	0	1	1	0	0		●G					●G						0	0	1	0	0	X	0	0	X	X	0	X	0	1	0	X	2004h
9	Both ch (with remote sensing)	0	0	1	1	1	1		●G	●Y				●G	●Y					0	1	1	0	0	X	0	0	X	X	1	X	0	1	1	X	6026h
10	PSU in stand-by mode	0	0	0	0	X	X												●B	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	8000h
11	Ch#1 remote programming	0	0	1	0	X	X									●R				0	X	X	X	X	X	X	X	X	X	X	1	0	X	X	X	0010h
12	Ch#2 remote programming	0	0	0	1	X	X				●R									0	X	X	X	1	X	X	X	X	X	X	X	0	X	X	X	0800h
13	Ch#1 CV mode	X	X	1	X	X	X							●G				●Y		0	X	X	X	X	X	X	X	1	0	X	X	0	1	X	X	0084h
14	Ch#1 CC mode	X	X	1	X	X	X							●G			●R			0	X	X	X	X	X	X	X	0	1	X	X	0	1	X	X	0044h
15	Ch#2 CV mode	X	X	X	1	X	X		●G				●Y							0	X	1	X	X	0	X	X	X	X	X	X	0	X	X	1	2001h
16	Ch#2 CC mode	X	X	X	1	X	X		●G			●R								0	X	1	X	X	1	X	X	X	X	X	X	0	X	X	0	2400h

Notes:

- The Remote sense (VOLTage:SENSe) is forbidden when channels are grouped in serial (SYSTem:GROup:SERial).
- Before channel grouping in serial or parallel all outputs has to be switched off (OUTPut OFF). New output enable can follow only after 20ms guard time that is required for relay contacts to stabilize (debouncing).