



721 West 1800 North
Logan, UT 84321

Certificate of Calibration Ultraviolet Sensor SU-200 and MU-200 Series

Serial Number : SU-200-SS_1252
Calibration Date : Jun-2023
Recommended Recalibration Date : Jun-2025
Calibration Uncertainty : $\pm 10 \%$
Measurement Repeatability : $< 1 \%$
Non-stability (Long-term Drift) : $< 2 \%$ per year
Calibration Factor : $10.00 \text{ W m}^{-2} \text{ per mV}$

Calibration Procedure

The calibration is based on a side-by-side comparison under high intensity discharge metal halide lamps using (3) Apogee transfer standard UV sensors. The Apogee transfer standards are calibrated to a spectroradiometer under clear sky conditions at Apogee Instruments in Logan, UT. The spectroradiometer is calibrated with an LI-1800-02 Optical Radiation Calibrator using a 200 Watt quartz halogen lamp. The LI-1800-02 and all replacement quartz halogen lamps are traceable to the U.S. National Institute of Standards and Technology (NIST).

Traceability

Reference Instrument	Serial Number	Classification
Li-Cor 1800-02L	ORL-1003L	Irradiance Lamp
Apogee PS-300	17110932	Spectroradiometer
Apogee SU-200	TS1	Photodiode Transfer Standard
Apogee SU-200	TS2	Photodiode Transfer Standard
Apogee SU-200	TS3	Photodiode Transfer Standard
Apogee SU-200	TS4	Photodiode Transfer Standard

Technical Manager :

Jacob Bingham

Date : 19-Jun-2023

Please keep this document for your records