
PATIENT:

Delivered by: Carolin
3750 NW CARY PKWY, Suite 120
Cary NC 27513
A (919)439-8580
a Pulmonary & Sleep Clinic

IDS
Instant Diagnostic Systems ™



RX#

Patient Instructions

Overnight Oximetry

Your physician has ordered this test to monitor your blood's oxygen saturation level while sleeping.

Do not turn on the Oximeter until you qo to bed for the night or your results may not be valid!

1. Preparation

Your doctor has ordered this test o be performed: _____

WEARING CPAP

It is suggested that you remove fingernail polish and coverings from the finger you will be using.

2. Sensor should be attached to oximeter.



3. When ready for bed, place sensor on finger, (wire on top of index finger)



___


(num
appear
di



4. Press and release the power button to turn on oximeter.

berswill
ear on
splay) I

5. Go to sleep, (wear the sensor all night)





6.

Wake up, press and hold power button for 2 seconds to turn off, oximeter. Remove sensor from finger.

Call scheduling # at

bottom of page for

pick-up

TROUBLESHOOTING

IF THE SENSOR FALLS OFF DURING THE NIGHT:

1) Replace the sensor on your finger
2) Check that the oximeter is on and displaying numbers.
3) If the oximeter is OFF, press the power button to turn it on.

IF THE PULSE INDICATOR ABOVE THE POWER BUTTON BLINKS YELLOW OR RED FOR SEVERAL MINUTES:

1) Reposition the sensor on finger.
2) Continue the test.

IF THE LOW BATTERY INDICATOR IS YELLOW:

1) DO NOT REPLACE BATTERIES.
2) Continue the test.
3) Advise your local supplier when they pick up the equipment.

Patient Testing Questions (Not for equipment pick-up) 1 -866-486-4891

In case of a medical emergency, call 911. ***To schedule pickup of test equipment call:*** **(919) 439-8580**

CAROLINA PULMONARY & **Sleep** clinic

05//2018