



Ultima[®] X Series Gas Monitors Instruction Manual Addendum

CORRECTIONS:

Page iii, General Warnings and Cautions

9. CSA performance Certification to standard C22.2 No. 152 is valid only when the instrument is calibrated on methane per the instruction manual.