



GWS Metrology Services

Certificate Number:
CBRE-44109-CET

CBRE - GWS LLC
Metrology Services
9410 Bunsen Parkway
Suite 100B
Louisville, KY 40220
502-495-5700



Date of Cal:
09-Jun-2023

Calibration Certificate

Customer :

GE APPLIANCES - A HAIER COMPANY
GE APPLIANCE PARK
LOUISVILLE, KY 40225

PO Number
46417112

Work Order
MAY 2023 AP TECH

Asset Number : AP2/CH 6B-MW100
Manufacturer : YOKOGAWA
Description : DATA ACQUISITION UNIT
Department : AP2
Location : WATER HEATER LAB

Serial Number : NSN
Model Number : MW100
Cal. Location : ON-SITE
Sublocation : CHAMBER 6B

Environmental Data

Temp : 22 °C
Humidity : 44 %

Calibration Information

***Calibration Due Date : 06/09/2024
Condition As Received : Meets Listed Specifications
Condition As Returned : Passed

Traceability Information

Asset Number	Description	Cal. Due Date	Reference Number
MET-0245	THERMOMETER, DIGITAL	11/4/2023	CBRE-4254-JTA
MET-0540	PLATINUM RESISTANCE THERMOMETER,	11/4/2023	CBRE-4255-JTA
MET-0642	CALIBRATOR, PROCESS	8/24/2023	CBRE-41780-CET
MET-0845	HUMIDITY INDICATOR	12/21/2023	CBRE-4692-JTA
MET-0846	HUMIDITY PROBE	12/21/2023	CBRE-4691-JTA

Traceability to NIST or other national metrology institutes for secondary measurement standards is established through laboratories approved by the CBRE-GWS, LLC Metrology Services quality assurance program. Test reports and calibration certificates maintained by CBRE - GWS are available upon request to the recipient of this calibration report.

Calibrated By: CHARLES THOMPSON - Metrologist

This is to certify that the above listed instrument/gage was inspected by CBRE GWS Metrology Services using a procedure(s) developed from the manufacturer specifications, accepted industry practices and/or customer requirements. The CBRE GWS Metrology Services Quality System conforms to ISO/IEC-17025:2017. It is hereby further certified that the inspection described herein was performed using standards whose values are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) or other National Metrology Institute (NMI), or have been derived from accepted values of natural constants, or have been derived by the ratio type of self calibration techniques. Uncertainties are estimated at a 95% confidence level. (k=2). The results indicated in this certificate relate only to the item(s) listed above. CBRE GWS Metrology Services responsibility shall in no event nor for any reason whatsoever exceed the purchase price of this calibration.

*** Calibration due dates are only issued if requested by the customer and are based upon customer dictated recall intervals.

Certificate Comments :

No Additional Comments

Expanded measurement uncertainty at an approximate 95% confidence level (k=2)