

CBRE - GWS LLC Metrology Services 9410 Bunsen Parkway Suite 100B Louisville, KY 40220 502-495-5700





Date of Cal: 12-Jun-2023

Certificate Number: CBRE-44116-CET

# **Calibration Certificate**

#### **Customer:**

**PO Number** 46417112

GE APPLIANCES - A HAIER COMPANY GE APPLIANCE PARK

Work Order MAY 2023 AP TECH

Asset Number : C2VK21023V

LOUISVILLE, KY 40225
I need to demonstrate +/-0.5% of reading in kWh. Please include gray bur ball brail of the care include g

Manufacturer : YOKOGAWA Model Number : WT310E

Description : POWER METER, DIGITAL Manufacturer spec is: Cal. Location : ON-SITE

Department : AP2 We test at two ranges.

Location: WATER HEATER GAS LAB1. ±(0.1% of reading + 0.05 % of range) - Range = 750Watts (150V and 5A)

2.  $\pm$ (0.1% of reading + 0.05 % of range) - Range = 6000Watts (300V and 20A)

**Environmental Data** 

Calibration Information

Temp: 23 °C

Humidity: 46 %

\*\*\*Calibration Due Date: 06/12/2024

Condition As Received: Meets Listed Specifications

Condition As Returned: Passed

### **Procedures used for this Calibration:**

Procedure #	Procedure Description	Rev #	Rev Date
CP-0553	YOKOGAWA WT310/330 DIGITAL POWER METER	00	10/21/2014

### **Traceability Information**

Asset Number	Description	Cal. Due Date	Reference Number
MET-0001	CALIBRATOR, MULTIFUNCTION	2/17/2024	CBRE-2733-TSM
MET-0731	SENSOR, TEMP/HUMIDITY	9/6/2023	CBRE-7120-TLB

Traceability to NIST or other national metrology institutes for secondary measurement standards is established through laboratories aproved by the CBRE-GWS, LLC Metrology Services quality assurance program. Test reports and calibration certificates maintained by CBRE - GWS are available upon request to the reciepient of this calibration report.

Charles Thompson

Calibrated By: CHARLES THOMPSON - Metrologist

This is to certify that the above listed instrument/gage was inspected by CBRE GWS Metrology Services using a procedure(s) developed from the manufacturer specifications, accepted industry practices and/or customer requirements. The CBRE GWS Metrology Services Quality System conforms to ISO/IEC-17025:2017. It is hereby further certified that the inspection described herein was performed using standards whose values are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) or other National Metrology Institute (NMI), or have been derived from accepted values of natural constants, or have been derived by the ratio type of self calibration techniques. Uncertainties are estimated at a 95% confidence level. (k=2). The results indicated in this certificate relate only to the item(s) listed above. CBRE GWS Metrology Services responsibility shall in no event nor for any reason whatsoever exceed the purchase price of this calibration.

\*\*\* Calibration due dates are only issued if requested by the customer and are based upon customer dictated recall intervals.

C2VK21023V Page 1 of 4 CBRE-44116-CET Form CCA Rev.16 Rev.Date: 3AUG2021

# Calibration Certificate (Cont)

Certificate Number: CBRE-44116-CET

CALIBRATION TEST POINTS			/	* Not accredited for this parameter VERDICT- (P)=Passed, (A)=Adjusted, (L)=Limited, (F)=Failed, (R)=Report of Value only				
Description	Neminal	Tal /	Tol .		1	1	1	<u> </u>
Description	Nominal	Tol -	Tol +	As Found	As Left	Units	Verdict	Comment
Freq 150V/0A	60.000	59.964	60.036	59.999	59.999	Hz	Р	
Freq 0V/1A	60.000	59.964	60.036	59.999	59.999	Hz	Р	
15V @ 60Hz	15.000	14.970	15.030	14.995	14.995	V	P	
30V @ 60Hz	30.000	29.94	30.060	29.990	29.990	V	P	
60V @ 60Hz	60.000	59.880	60.120	59.981	59.981	v	P	
150V @ 60Hz	150.00	149.70	150.30	149.96	149.96	V	P	
150V MEAN @ 60Hz	150.00	149.70	150.30	149.96	149.96	v	P	
150V @ 1kHz	100.00	99/60	100.40	99.97	99.97	V	P	
150V @ 10kHz	100.00	98.85	101.15	100.01	100.01	V	P	
150V @ 20kHz	100.00	98.35	101.65	100.12	100.12	V	P	
150V @ 50kHz	100.00	97.15	102.85	100.30	100.30	V	P	
150V @ 100kHz	100.00	95.15	104.85	98.84	98.84	V	P	
150V MEAN @ 100kHz	100.00	95.15	104.85	98.81	98.81	V	P	
300V @ 60Hz	300.00	299.40	300.60	299.93	299.93	V	P	
600V DC	60.00	58.74	61.26	59.99	59.99	V	P	
600V DC	-60.00	-61.26	-58.74	-59.98	-59.98	V	P	
600V DC	100.00	98.70	101.30	99.98	99.98	V	P	
600V DC	-100.00	-101.30	-98.70	-99.98	-99.98	V	P	
600V DC	300.00	298.50	301.50	299.97	299.97	V	P	
600V DC	-300.00	-301.50	-298.50	-299.91	-299.91	V	P	
600V DC	600.00	598.20	601.80	599.96	599.96	V	P	
600V DC	-600.00	-601.80	-598.20	-599.89	-599.89	V	P	
600V @ 60Hz	60.00	59.34	60.66	59.99	59.99	V	P	
600V @ 60Hz	100.00	99.30	100.70	99.99	99.99	V	P	
600V @ 60Hz	300.00	299.10	300.90	299.95	299.95	V	P	
600V @ 60Hz	600.00	598.80	601.20	599.96	599.96	V	P	
						·		
5mA @ 60Hz	5.0000	4.9900	5.0100	4.9982	4.9982	mA	Р	
10mA @ 60Hz	10.000	9.980	10.020	9.996	9.996	mA	Р	
10mA @ 1kHz	10.000	9.970	10.030	9.996	9.996	mA	Р	
10mA @ 10kHz	10.000	9.900	10.100	10.001	10.001	mA	Р	
10mA @ 20kHz	10.000	9.860	10.140	10.017	10.017	mA	Р	
20mA @ 60Hz	20.000	19.960	20.040	19.992	19.992	mA	Р	
50mA @ 60Hz	50.000	49.900	50.100	49.982	49.982	mA	Р	
100mA @ 60Hz	100.00	99.80	100.20	99.96	99.96	mA	Р	
0.5A @ 60Hz	500.00	499.00	501.00	499.95	499.95	mA	Р	
1A DC	0.1000	0.0980	0.1020	0.1001	0.1001	Α	Р	
1A DC	-0.1000	-0.1020	-0.0980	-0.0999	-0.0999	Α	Р	
1A DC	0.2000	0.1980	0.2020	0.2000	0.2000	Α	Р	
1A DC	-0.2000	-0.2020	-0.1980	-0.1999	-0.1999	Α	Р	
1A DC	0.6000	0.5970	0.6030	0.6000	0.6000	Α	Р	
1A DC	-0.6000	-0.6030	-0.5970	-0.5999	-0.5999	Α	Р	
1A DC	1.0000	0.9970	1.0030	0.9999	0.9999	Α	Р	
1A DC	-1.0000	-1.0030	-0.9970	-0.9999	-0.9999	Α	Р	
1A @ 10Hz	1.0000	0.9970	1.0030	0.9997	0.9997	Α	Р	
1A @ 60Hz	0.1000	0.0990	0.1010	0.1000	0.1000	Α	Р	
1A @ 60Hz	0.2000	0.1990	0.2010	0.2000	0.2000	Α	Р	
1A @ 60Hz	0.6000	0.5980	0.6020	0.5999	0.5999	Α	Р	
1A @ 60Hz	1.0000	0.9980	1.0020	0.9999	0.9999	Α	Р	
1A @ 1kHz	1.0000	0.9970	1.0030	1.0001	1.0001	Α	P	
1A @ 10kHz	1.0000	0.9900	1.0100	1.0056	1.0056	Α	P	
2A @ 60Hz	2.0000	1.9960	2.0040	2.0001	2.0001	Α	Р	
5A DC	5.0000	4.9850	5.0150	4.9999	4.9999	Α	Р	
5A DC	-5.0000	-5.0150	-4.9850	-5.0007	-5.0007	Α	P	
5A @ 60Hz	5.0000	4.9900	5.0100	5.0001	5.0001	Α	Р	
5A @ 1kHz	5.0000	4.9850	5.0150	5.0015	5.0015	Α	P	
10A @ 60Hz	10.000	9.980	10.020	10.002	10.002	A	P	
20A @ 60Hz	20.000	19.960	20.040	20.005	20.005	Α	Р	I



## Calibration Certificate (Cont)

Certificate Number : CBRE-44116-CET

\* Not accredited for this parameter

**CALIBRATION TEST POINTS** 

VERDICT- (P)=Passed, (A)=Adjusted, (L)=Limited, (F)=Failed, (R)=Report of Value only

				· · · · · · · · · · · · · · · · · · ·	LRDICT- (F)=Fasse	u, (A)=Aujustet	u, (L)=Lillilleu,	)=Failed, (R)=Report of Value of Comment		
Description	Nominal	Tol -	Tol +	As Found	As Left	Units	Verdict	Comment		
15V/1A @ 60Hz - PF 1	15.000	14.970	15.030	14.999	14.999	W	Р			
30V/1A @ 60Hz - PF 1	30.000	29.940	30.060	29.992	29.992	W	Р			
60V/1A @ 60Hz - PF 1	60.000	59.880	60.120	59.985	59.985	W	Р			
150V/1A @ 60Hz - PF 1	150.00	149.70	150.30	149.98	149.98	W	Р			
150V-100V/1A @ 1kHz - PF 1	100.00	99.50	100.50	99.98	99.98	W	Р			
150V-100V/1A @ 10kHz - PF 1	100.00	98.85	101.15	100.55	100.55	W	Р			
300V/1A @ 60Hz - PF 1	300.00	299.40	300.60	299.95	299.95	W	Р			
600V-60V/1A @ DC	60.00	58.74	61.26	59.99	59.99	W	Р			
600V60V/1A @ DC	-60.00	-61.14	-58.86	-59.97	-59.97	W	Р			
600V-100V/1A @ DC	100.00	98.70	101.30	99.96	99.96	W	Р			
600V100V/1A @ DC	-100.00	-101.10	-98.90	-99.96	-99.96	W	Р			
000V/1A @ DC	600.00	598.20	601.80	599.85	599.85	W	Р			
600V600/1A @ DC	-600.00	-600.60	-599.40	-599.79	-599.79	W	Р			
600V-60/1A @ 60Hz - PF 1	60.00	59.34	60.66	59.99	59.99	W	Р			
600V-100V/1A @ 60Hz - PF 1	100.00	99.30	100.70	99.99	99.99	W	Р			
600V/1A @ 60Hz - PF 1	600.00	598.80	601.20	599.97	599.97	W	Р			
150V-100V/5mA @ 60Hz - PF 1	500.00	498.75	501.25	499.90	499.90	mW	Р			
150V-100V/10mA @ 60Hz - PF 1	1.0000	0.9980	1.0030	0.9998	0.9998	W	Р			
150V-100V/10mA @ 1kHz - PF 1	1.0000	0.9950	1.0050	0.9997	0.9997	W	Р			
150V-100V/10mA @ 10kHz - PF 1	1.0000	0.9890	1.0120	0.9997	0.9997	W	Р			
50V-100V/10mA @ 20kHz - PF 1	1.0000	0.9790	1.0220	0.9997	0.9997	W	Р			
150V-100V/20mA @ 60Hz - PF 1	2.0000	1.9950	2.0050	2.0033	2.0033	W	Р			
150V-100V/50mA @ 60Hz - PF 1	5.0000	4.9900	5.0100	4.9992	4.9992	W	Р			
50V-100V/100mA @ 60Hz - PF 1	10.000	9.975	10.025	9.998	9.998	W	Р			
50V-100V/2A @ 60Hz - PF 1	200.00	199.50	200.50	199.98	199.98	W	Р			
150V-100V/5A @ 60Hz - PF 1	500.00	498.75	501.25	499.99	499.99	W	Р			
150V-100V/10A @ 60Hz - PF 1	1.0000	0.9975	1.0025	0.9999	0.9999	kW	Р			
150V-100V/20A @ 60Hz - PF 1	2.0000	1.9950	2.0050	2.0000	2.0000	kW	Р			

The verdicts above are based upon a direct comparison of the measured value at the time of calibration, to a published or customer supplied tolerance for the specification listed. CBRE-GWS does not include the measurement uncertainty in making these determinations unless specifically requested. It is the responsibility of the user of this equipment to determine if the accept / reject tolerances meet the requirements of the intended measurement process.

### **Certificate Comments:**

No Additional Comments

### Expanded measurement uncertainty at an approximate 95% confidence level (k=2)

DC Voltage 0 to 200 mV

 $\pm 0.1 \; \mu V + 4.5 \mu V / V + 0.6 R$ 

200 mV to 2 V

 $\pm 0.4 \; \mu V + 3 \mu V / V + 0.6 R$ 

(2 to 20) V

±4 µV 3µV/V+0.6R

(20 to 200) V

±40 µV+4.5µV/V+0.6R

(200 to 1 000) V

 $\pm 0.5$ mV+4.5 $\mu$ V/V+0.6R

AC Voltage

1 mV to 200 mV

(10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz

(2 to 10) kHz

±4 µV+130µV/V +0.6R ±4 µV+110µV/V + 0.6R ±2µV+105µV/V+ 0.6R ±4µV+105µV/V+0.6R ±8µV+305µV/V+ 0.6R

(10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (300 to 500) kHz

±20µV+705µV/V+0.6R ±0.3%iv+0.2 mV+ 0.6R ±1 %iv + 2 mV + 0.6R

200 mV to 2 V

(10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz ±20μV+105μV/V+0.6R ±20μV+ 85μV/V+0.6R ±20μV+ 65 μV/V +0.6R ±22μV+ 85 μV/V +0.6R



#### **GWS Metrology Services**

(10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (300 to 500) kHz (300 to 500) kHz (300 to 500) kHz (40µV+205µV/V+0.6R ±0.2mV+505µV/V+0.6R ±0.3 % iv + 2 mV +0.6R ±1 % iv + 2 mV +0.6R

2 V to 20 V

(10 to 40) Hz ±0.2mV+105μV/V+0.6R (40 to 100) Hz ±0.2mV+85 μV/V+0.6R 100 Hz to 2 kHz ±0.2mV+65 μV/V+0.6R (2 to 10) kHz ±0.2mV+ 85 μV/V+0.6R ±0.2mV+ 85 μV/V+0.6R ±0.2mV+ 95 μV/V+0.6R

20 V to 200 V

(10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 20) kHz (2 to 20) kHz (3 to 20) kHz (4 to 20) kHz (5 to 20) kHz (5 to 20) kHz (6 to 20) kHz (7 to 20) kHz (8 to 20) kHz (9 to 20) kHz (1 to 20) kHz

200 V to 1 000 V

(10 to 40) Hz  $\pm 20 \text{ mV} + 110 \mu \text{V/V} + 0.6 \text{R}$ (40 to 10) kHz  $\pm 20 \text{ mV} + 95 \mu \text{V/V} + 0.6 \text{R}$ 

Resistance - Source

0 to 110hm ±0.009%iv+0.006Ohm+0.6 (11 to 33)Ohm ±0.009 4%iv+0.01Ohm+0.6R (33 to 330)Ohm ±0.007%iv + 0.01Ohm+0.6R 3300hm to 3.3 kOhm ±0.007%iv + 0.060hm+0.6R ±0.007%iv + 0.60hm+0.6R (3.3 to 33) kOhm ±0.008 6%iv+ 6Ohm+0.6R (33 to 110) kOhm (110 to 330) kOhm ±0.009 3%iv + 60hm+0.6R 330 kOhm to 3.3 MOhm ±0.012%iv+ 550hm+0.6R (3.3 to 11) MOhm ±0.047%iv+ 550Ohm+0.6R (11 to 33) MOhm ±0.078%iv+ 550Ohm+0.6R (33 to 110) MOhm ±0.39%iv+ 5.55kOhm+0.6R (110 to 330) MOhm ±0.39%iv+ 16.5kOhm+0.6R

DC Current

AC Current 29 μA to 200 μA

10 Hz to 10 kHz  $\pm 20 \text{ nA} + 280 \mu\text{A/A} + 0.6\text{R}$ 

200  $\mu A$  to 2 mA

10 Hz to 10 kHz ±200 nA +280 μA/A +0.6R

2 mA to 20 mA  $\,$ 

10 Hz to 10 kHz ±2 μA+ 280 μA/A+0.6R

20 mA to 200 mA

10 Hz to 10 kHz  $\pm$ 20  $\mu$ A+ 250  $\mu$ A/A +0.6R

200 mA to 2 A

 $\begin{array}{lll} \mbox{10 Hz to 2 kHz} & \pm 0.2 \mbox{ mA} + 600 \mbox{ $\mu$A/A+ 0.6R} \\ \mbox{2 kHz to 5 kHz} & \pm 0.2 \mbox{ mA} + 700 \mbox{ $\mu$A/A+ 0.6R} \end{array}$ 

Calibration Certificate (Cont)

**Certificate Number: CBRE-44116-CET** 

C2VK21023V Page 4 of 4 CBRE-44116-CET Form CCA Rev.16 Rev.Date: 3AUG2021