Out-of-specification investigation summary, Unit 1, Q3 2026

Scope

Out-of-specification handling for finished product tests at Unit 1 from July to September 2026. 5 failed test results were investigated.

Results

OOS reference	Test	Phase 1 finding	Outcome
OOS-26-031	Assay	Analyst error, wrong dilution	Retest valid, batch released
OOS-26-034	Dissolution	No lab error found	Deviation raised, batch held
OOS-26-036	Residual solvent	No lab error found	Deviation raised, cleaning validation reviewed
OOS-26-039	Assay	Instrument calibration overdue	Retest valid, batch released
OOS-26-041	Related substances	No lab error found	Deviation raised, batch rejected

CAPA

Three deviations led to corrective and preventive action. The residual solvent result showed that the residue limit study of the shared granulator did not cover the new solvent. Cleaning validation is repeated for that equipment before the next campaign. CAPA tracking shows all three actions due by 30 November 2026.

Synthetic sample written for the RPG NeuralMesh demonstration. It is not a record of the RPG Group.