Title:

Certificate of Analysis (CoA)

Date: 9/25/2026
Date Tested: 9/23/2026
Customer: TIDElwave Research Solutions
Testing material: Vitamin B12
Lot Number: 20260425-B12-3BF2
Sample ID: BTL260811-012-026
Content/Potency: 10 mL
Storage: R.T.
Visual Description: Large clear vial: red sample, black label, gold crimp, clear plastic cap.
Labeled as: Vitamin B12
Manufacturer: N/A
Testing Purpose: FTIR and HPLC analysis for the identification, purity, potency and composition of a vitamin product. It does not provide information on particulate matter, microbial contamination or presence of endotoxins.



Test	Method	Specification	Result
General Appearance	USP <630>	red liquid	red liquid
Mass	USP <41>	As recorded	10 mL
FTIR Identification and Composition Analysis	USP <197A>	Sample spectrum should confirm the content of vitamin via characteristic bands	FTIR sample spectrum confirms the presence of Vitamin B12 in BAC water.
HPLC Purity of Peptide Assay	USP <621>	Specifications: ≥ 98 %	99.7 %
HPLC Potency Assay	USP <621>	Specifications: 90 – 110% of 10 mg	9.1 mg (90.6 %)

The results of the CoA relate only to the item(s) tested and applied to the sample as received.

Andrea Castro, BS
Scientist-II
BTLabs


Escarlet Cabrieles, BS
Scientist-II
BTLabs

5601 Corporate Way | Suite 206 | West Palm Beach, FL 33407
Phone: (561) 625-0133
E-mail: info@btlabtesting.com | Website: https://btlabtesting.com

Title:

Certificate of Analysis (CoA)

FTIR ID and Composition Analysis: Vitamin B12 Lot 20260425-B12-3BF2



HPLC Purity and Potency Assay @ 220 nm: Vitamin B12 Lot 20260425-B12-3BF2



Vitamin B12 Lot 20260425-B12-3BF2 @ 220 nm		
Peak #:	Retention Time (min)	Area (mAU*s)
1	1.888	4757.8

5601 Corporate Way | Suite 206 | West Palm Beach, FL 33407
Phone: (561) 625-0133
E-mail: info@btlabtesting.com | Website: https://btlabtesting.com