

ASSET MAINTENANCE & ENGINEERING ACTION REPORT

MATRIX GROUP GLOBAL LLC • CORPORATE FACILITIES & EQUIPMENT INFRASTRUCTURE REGISTRY

Work Order ID	MNT-2026-304	Execution Date	May 30, 2026
Target Asset Name	Gamma Lab Centrifuge #12		
Assigned Technician	Bruce Banner, Lab Safety Officer	Physical Location	Biomedical Research Wing, Lab G

CURRENT MAINTENANCE STATUS: COMPLETED

1. REPORTED ISSUE / DIAGNOSTIC BASELINE FAULTS

Rotor locking mechanism reported intermittent sensor calibration faults during active sample acceleration validation tests.

2. CORRECTIVE REMEDIATION ACTIONS EXECUTED

Replaced the optical lid engagement sensor, re-calibrated the magnetic braking parameters, and performed a zero-drift balancing test run.

This document serves as an authorized technical record. All post-maintenance asset metrics have been verified against corporate equipment safety rules and re-logged into the primary operational framework dashboard.

Field Technician Signature

Facilities Operations Sign-off