

Dry Cannabis Flower for Medical use·*Cannabis sativa* L.·Ph. Eur. Monograph 3028·*Cannabis flos*

Wedding Crasher

THC-dominant

THC 23 : CBD 1

Grade II

Lab Sample ID

Лаб. ИД на примерок

PP-QCL-26-095

Production Batch №

Производна серија №

P060089

Requested Test

Побарано испитување

Loss on Drying (Param. 8)

Specification Ref.

Референца на спецификација

QCSP-IMB-001 v.02 · §8

01 Analytical Method | Аналитички метод

Method Метода	Method validation status: Cp · VERIFIED SOP · QCSOP-006 v.02
Loss on Drying — Vacuum-Oven over Molecular Sieve R(Ph. Eur. 2.2.32 Method E)	
Cut and homogenise sample; weigh accurately ≈ 1.0 g into a tared, low-form weighing bottle (n=2 replicates). Dry in vacuum oven over molecular sieve R at 40 °C ± 1 °C under 1.5–2.5 kPa for 24 h until constant mass. Cool to room temperature in a desiccator over NaOH-CaCO ₃ trap; weigh on calibrated 4-decimal balance (Sartorius MSA225P-1CE-DA · Cal-Cert 2026-04 · u(m) = 0.0002 g). Constant mass verified by <0.5 mg difference between two successive 1-h weighings at the end of the cycle. Test executed PRIOR to sampling for batch release , immediately before primary packaging.	

02 Result(s) · Replicate Gravimetric Data | Резултат(u)

Rep	Replicate Description	Initial m ₀ (g)	Final m ₁ (g)	Δm = m ₀ – m ₁ (g)	LoD (% w/w)	Status
R1	Replicate 1Bottle ID LB-26-088-1 · tare	1.0010	0.9224	0.0786	7.85 %	OK
	11.3245 g · room T 21.4 °C / 47 % RH at weigh-out					
R2	Replicate 2Bottle ID LB-26-088-2 · tare	1.0008	0.9226	0.0782	7.81 %	OK
	11.4012 g · room T 21.5 °C / 47 % RH at weigh-out					
Mean LoD (n=2)					7.83 % w/w	PASS
RSD · Replicate Spread (R1–R2)					0.36 % · Δ 0.04 %	≤ 0.2 % OK

CM Constant-mass verification: R1 Δ(23 h → 24 h) = 0.3 mg · R2 Δ(23 h → 24 h) = 0.2 mg — both < **0.5 mg** threshold. **Constant mass achieved at 24 h.** Molecular sieve regeneration check: 18.3 % mass loss at 250 °C confirms full activity.

03 Method Validation · Measurement Uncertainty | Валидација на метод · Мерна неодреденост

Method Status Статус на метод
Compendial · Verified for cannabis flower matrix per Ph. Eur. 5.x · Verification dossier QCV-008 v.01 · 6-lot inter-day reproducibility < 0.10 % w/w absolute · 2-analyst inter-bias < 0.05 % w/w · Method E chosen for heat-labile cannabinoid preservation.
Measurement Uncertainty Мерна неодреденост
u(m ₀) = u(m ₁) = 0.0002 g (balance certificate) · u(Δm) = √2·0.0002 = 0.00028 g · u(LoD) _{rel} ≈ 0.028 % · Combined expanded uncertainty (k=2, ~95 %) on LoD ≈ ±0.06 % w/w absolute . Reported 7.83 % w/w ± 0.06 %.

Conforms

Result: The tested sample **conforms** to the product specification **QCSP-IMB-001 v.02 §8 (Loss on Drying)** per Ph. Eur. 11.5 Monograph 3028 + general chapter 2.2.32 — mean LoD **7.83 % w/w** (≤ 12.0 %). Replicate-agreement criterion met (Δ 0.04 % < 0.2 %). Test executed prior to primary packaging. QC disposition only; final batch release is certified separately by the QP in linked CoQ CoQ-PP-2026-0012.

Notes | Забелешки

- These results relate **only to the tested sample(s)** as received in the QC Laboratory.
- No part of this report may be reproduced** except in full, without prior written permission of the QC Laboratory.
- Measurement uncertainty is evaluated in accordance with the QC Laboratory's quality system; the **detailed uncertainty calculation is available on request**.
- Testing was performed **in-house** by Purely Plant QC personnel under an **ISO/IEC 17025-aligned** quality system within the **MK GMP** facility, using the compendial Ph. Eur. 2.2.32 Method E procedure verified for the cannabis-flower matrix (QCV-008). **No analyses outsourced**. OOS handling per QCSOP-014.

Analysis Performed by

Анализата ја изврши

Analyst · QC Laboratory

Аналитичар · Лабораторија КК

Stojanka Pavlova

Lab ID PP-QCL-04

Date · Датум11.03.2026

Reviewed by

Прегледал

Senior Analyst · QC Laboratory

Виш аналитичар · Лабораторија КК

Marija Trajkovska

Lab ID PP-QCL-02

Date · Датум11.03.2026

Approved by

Одобрил

Head of QC Laboratory

Раководител на Лабораторија КК

Blagoj Nikolov

QC Manager · GMP-trained

Date · Датум11.03.2026

Purely Plant DOOEL · QC Analytical Laboratory · Industriska ulica 9, br. 9,
s. Kojlija 1043 · Petrovec-Skopje, North Macedonia

1 | 1

QCSOP-012 v. 03